

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kinderprep **SDG No.:** Q2145
Contract: KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168201BS							
Copper	ug/L	5000	5030		101	85 - 115	P
Lead	ug/L	2500	2390		96	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Kinderprep **SDG No.:** Q2145
Contract: KIND02 **Lab Code:** CHEM **Case No.:** Q2145 **SAS No.:** Q2145

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168416BS							
Copper	ug/L	5000	5310		106	85 - 115	P
Lead	ug/L	2500	2460		98	85 - 115	P