

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Sciacca General Contractors, LLC Date Collected:

Project: 98 Morse Ave Nutley Date Received:

Client Sample ID: 3MSD SDG No.: Q2147

Lab Sample ID:Q2147-05MSDMatrix:SolidAnalytical Method:NJEPH% Solid:86.3

Sample Wt/Vol: 30.06 Units: g Final Vol: 2000 uL

Soil Aliquot Vol: uL Test: EPH F2

Prep Method:

Prep Date : Date Analyzed : Prep Batch ID

06/03/25 10:00 06/03/25 17:26 PB168239

Datafile

| CAS Number | Parameter | Conc. | Qualifier | Dilution | MDL | LOQ / CRQL | Units(Dry Weight) | |
|-------------------------|--------------------|-------|-----------|----------|------|------------|-------------------|--------|
| TARGETS Aliphatic C9-C2 | 8 Aliphatic C9-C28 | 99.5 | Е | 1 | 1.05 | 4.63 | mg/kg FE05 | 4156.D |
| Total EPH | Total EPH | 99.5 | | | 1.05 | 4.63 | mg/kg | |

^{*} As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution



284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: Sciacca General Contractors, LLC Date Collected:

Project: 98 Morse Ave Nutley Date Received:

Client Sample ID: 3MSD SDG No.: Q2147

Lab Sample ID:Q2147-05MSDMatrix:SolidAnalytical Method:NJEPH% Solid:86.3

Sample Wt/Vol: 30.06 Units: g Final Vol: 2000 uL

Soil Aliquot Vol: uL Test: EPH F2

Prep Method:

Prep Date : Date Analyzed : Prep Batch ID

06/03/25 10:00 06/03/25 17:26 PB168239

Datafile

| CAS Number | Parameter | Conc. | Qualifier | Dilution | MDL | LOQ / CRQL | Units(Dry Weight) | |
|-------------------------|--------------------|-------|-----------|----------|------|------------|-------------------|--------|
| TARGETS Aliphatic C9-C2 | 8 Aliphatic C9-C28 | 99.5 | Е | 1 | 1.05 | 4.63 | mg/kg FE05 | 4156.D |
| Total EPH | Total EPH | 99.5 | | | 1.05 | 4.63 | mg/kg | |

^{*} As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C28 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C28 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution



uL



Report of Analysis

Client: Sciacca General Contractors, LLC Date Collected:

Project: 98 Morse Ave Nutley Date Received:

Client Sample ID: 3MSD SDG No.: Q2147

Lab Sample ID: Q2147-05MSD Matrix: Solid

Analytical Method: NJEPH % Solid: 86.3

Sample Wt/Vol: 30.06 Units: g Final Vol: 2000

Soil Aliquot Vol: uL Test: EPH_F2

Prep Method:

 File ID :
 Dilution:
 Prep Date :
 Date Analyzed :
 Prep Batch ID

 FE054156.D
 1
 06/03/25
 06/03/25
 PB168239

| CAS Number | Parameter | | Conc. Q | ualifier | MDL | LOQ / CRQL | Units |
|----------------|-----------|---------------------------|---------|----------|----------|------------|---------|
| TARGETS | | | | | | | |
| Aliphatic C9-C | C28 | Aliphatic C9-C28 | 99.5 | E | 1.05 | 4.63 | mg/kg |
| Aliphatic C28- | -C40 | Aliphatic C28-C40 | 56.7 | E | 1.36 | 2.31 | mg/kg |
| SURROGATES | 8 | | | | | | |
| 3383-33-2 | | 1-chlorooctadecane (SURR) | 36.1 | | 40 - 140 | 72% | SPK: 50 |
| 84-15-1 | | ortho-Terphenyl (SURR) | 30.8 | | 40 - 140 | 62% | SPK: 50 |



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID: Q2147-05MSD Acq On: 03 Jun 2025 17:26

Client Sample ID: Q2147-05MSD Operator: YP\AJ

Data file: FE054156.D Misc:

Instrument: FID_E ALS Vial: 13

Dilution Factor: 1 Sample Multiplier: 1.00

| Compound | R.T. | | Response | Conc | highest_standard | Units |
|---------------------------|--------|--------|-----------|---------|------------------|-------|
| Aliphatic C9-C12 | 3.085 | 6.730 | 26770159 | 197.474 | 300 | ug/ml |
| Aliphatic C12-C16 | 6.731 | 10.178 | 34755542 | 257.56 | 200 | ug/ml |
| Aliphatic C16-C21 | 10.179 | 13.552 | 49219153 | 373.124 | 300 | ug/ml |
| Aliphatic C21-C28 | 13.553 | 17.220 | 57399957 | 462.243 | 400 | ug/ml |
| Aliphatic C28-C40 | 17.221 | 22.091 | 84830276 | 734.901 | 600 | ug/ml |
| Aliphatic EPH | 3.085 | 22.091 | 252975087 | 2030 | | ug/ml |
| ortho-Terphenyl (SURR) | 11.839 | 11.839 | 5003569 | 30.79 | | ug/ml |
| 1-chlorooctadecane (SURR) | 13.283 | 13.283 | 4282711 | 36.13 | | ug/ml |
| Aliphatic C9-C28 | 3.085 | 17.220 | 168144811 | 1290 | 1200 | ug/ml |