

Analytical Sequence

Client: Sciacca General Contractors, LLC

SDG No.: Q2147

Project: 98 Morse Ave Nutley

Instrument ID: FID_G

GC Column: RXI-1MS **ID:** 0.18 (mm)

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

MEAN SUROGATE RT FROM INITIAL CALIBRATION 14.9848					
EPA SAMPLE NO.	LAB SAMPLE ID	DATE AND TIME ANALYZED	DATAFILE	RT	#
PIBLK01	LBLK01	10 Jun 2025 10:32	FG016001.D	14.975	
50 PPM TRPH STD	50 PPM TRPH STD	10 Jun 2025 11:01	FG016002.D	14.978	
PB168365BL	PB168365BL	10 Jun 2025 12:10	FG016004.D	14.975	
PB168365BS	PB168365BS	10 Jun 2025 12:39	FG016005.D	14.974	
WASTE	Q2147-01	10 Jun 2025 13:08	FG016006.D	14.973	
WASTEMS	Q2147-01MS	10 Jun 2025 13:38	FG016007.D	14.972	
WASTEMSD	Q2147-01MSD	10 Jun 2025 14:07	FG016008.D	14.974	
PIBLK02	LBLK02	10 Jun 2025 14:37	FG016009.D	14.973	
50 PPM TRPH STD	50 PPM TRPH STD	10 Jun 2025 15:06	FG016010.D	14.976	