

Surrogate Summary

SDG No.: Q2147

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2147-02	VOC	1,2-Dichloroethane-d4	50	38.5	77	63	155
		Dibromofluoromethane	50	41.0	82	70	134
		Toluene-d8	50	44.2	88	74	123
		4-Bromofluorobenzene	50	23.9	48	17	146
VY0529SBL01	VY0529SBL01	1,2-Dichloroethane-d4	50	53.3	107	63	155
		Dibromofluoromethane	50	50.8	102	70	134
		Toluene-d8	50	50.1	100	74	123
		4-Bromofluorobenzene	50	40.9	82	17	146
VY0529SBS01	VY0529SBS01	1,2-Dichloroethane-d4	50	53.5	107	63	155
		Dibromofluoromethane	50	49.6	99	70	134
		Toluene-d8	50	49.6	99	74	123
		4-Bromofluorobenzene	50	48.2	96	17	146
VY0529SBSD01	VY0529SBSD01	1,2-Dichloroethane-d4	50	55.1	110	63	155
		Dibromofluoromethane	50	52.5	105	70	134
		Toluene-d8	50	51.6	103	74	123
		4-Bromofluorobenzene	50	49.9	100	17	146