

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2147

Client: Sciaccia General Contractors, LLC

Analytical Method: SW8260D

Datafile : VY022456.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0529SBSD01	Dichlorodifluoromethane	20	23.4	ug/Kg	117	3		64	136	20
	Chloromethane	20	26.4	ug/Kg	132	0		52	151	20
	Vinyl chloride	20	24.5	ug/Kg	123	3		56	148	20
	Bromomethane	20	28.8	ug/Kg	144	3	*	58	141	20
	Chloroethane	20	27.0	ug/Kg	135	1	*	69	130	20
	Trichlorofluoromethane	20	25.3	ug/Kg	127	6		69	134	20
	1,1,2-Trichlorotrifluoroethane	20	21.3	ug/Kg	106	1		81	123	20
	1,1-Dichloroethene	20	22.1	ug/Kg	111	5		79	121	20
	Acetone	100	120	ug/Kg	120	9		40	171	20
	Carbon disulfide	20	21.6	ug/Kg	108	4		59	130	20
	Methyl tert-butyl Ether	20	23.0	ug/Kg	115	3		77	129	20
	Methyl Acetate	20	25.2	ug/Kg	126	9		69	149	20
	Methylene Chloride	20	21.4	ug/Kg	107	3		72	131	20
	trans-1,2-Dichloroethene	20	21.9	ug/Kg	110	6		80	123	20
	1,1-Dichloroethane	20	22.8	ug/Kg	114	5		82	123	20
	Cyclohexane	20	20.7	ug/Kg	104	0		76	122	20
	2-Butanone	100	110	ug/Kg	110	0		69	131	20
	Carbon Tetrachloride	20	20.5	ug/Kg	103	2		76	129	20
	cis-1,2-Dichloroethene	20	21.9	ug/Kg	110	4		82	123	20
	Bromochloromethane	20	22.0	ug/Kg	110	2		80	127	20
	Chloroform	20	22.8	ug/Kg	114	4		82	125	20
	1,1,1-Trichloroethane	20	21.6	ug/Kg	108	1		80	126	20
	Methylcyclohexane	20	20.0	ug/Kg	100	2		77	123	20
	Benzene	20	21.4	ug/Kg	107	3		84	121	20
	1,2-Dichloroethane	20	21.9	ug/Kg	110	2		81	126	20
	Trichloroethene	20	20.9	ug/Kg	104	1		83	122	20
	1,2-Dichloropropane	20	21.9	ug/Kg	110	4		83	122	20
	Bromodichloromethane	20	21.5	ug/Kg	108	4		82	123	20
	4-Methyl-2-Pentanone	100	110	ug/Kg	110	0		70	135	20
	Toluene	20	20.6	ug/Kg	103	2		83	122	20
	t-1,3-Dichloropropene	20	21.2	ug/Kg	106	0		78	124	20
	cis-1,3-Dichloropropene	20	21.2	ug/Kg	106	2		81	122	20
	1,1,2-Trichloroethane	20	21.2	ug/Kg	106	0		82	125	20
	2-Hexanone	100	110	ug/Kg	110	0		66	138	20
	Dibromochloromethane	20	21.1	ug/Kg	106	0		79	125	20
	1,2-Dibromoethane	20	21.6	ug/Kg	108	2		80	125	20
	Tetrachloroethene	20	20.9	ug/Kg	104	0		83	125	20
	Chlorobenzene	20	20.9	ug/Kg	104	3		84	122	20
	Ethyl Benzene	20	20.2	ug/Kg	101	2		82	124	20
	m/p-Xylenes	40	40.0	ug/Kg	100	2		83	124	20
	o-Xylene	20	20.3	ug/Kg	102	2		83	123	20
	Styrene	20	20.4	ug/Kg	102	1		82	124	20
	Bromoform	20	20.9	ug/Kg	104	3		75	127	20
	Isopropylbenzene	20	20.4	ug/Kg	102	3		82	124	20
	1,1,2,2-Tetrachloroethane	20	21.8	ug/Kg	109	3		77	127	20
	1,3-Dichlorobenzene	20	20.5	ug/Kg	103	6		83	122	20
	1,4-Dichlorobenzene	20	21.2	ug/Kg	106	7		84	121	20
	1,2-Dichlorobenzene	20	21.1	ug/Kg	106	4		83	124	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2147

Client: Sciacca General Contractors, LLC

Analytical Method: SW8260D

Datafile : VY022456.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0529SBSD01	1,2-Dibromo-3-Chloropropane	20	22.3	ug/Kg	112	3		66	134	20
	1,2,4-Trichlorobenzene	20	21.7	ug/Kg	109	10		78	127	20
	1,2,3-Trichlorobenzene	20	22.0	ug/Kg	110	7		70	137	20