

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: SCIA01
 Lab Code: CHEM Case No.: Q2147 SAS No.: Q2147 SDG NO.: Q2147
 Lab File ID: VY022453.D Date Analyzed: 05/29/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:23
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	217073	7.72	367403	8.62	311885	11.42
UPPER LIMIT	434146	8.219	734806	9.122	623770	11.92
LOWER LIMIT	108537	7.219	183702	8.122	155943	10.92
EPA SAMPLE NO.						
VOC	22984 *	7.71	34413 *	8.62	20399 *	11.42
VY0529SBL01	320837	7.72	584950	8.62	455837	11.43
VY0529SBS01	217849	7.72	376540	8.62	319754	11.42
VY0529SBSD01	205776	7.71	359832	8.62	305847	11.42

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.