

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2148

Project: Monthly 2025

Run No.: LB135951

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135951BS							
BOD5	mg/L	198	212		107	1	84.6-115.4	05/29/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2148

Project: Monthly 2025

Run No.: LB135984

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135984BS							
TSS	mg/L	550	533		97	1	90-110	06/02/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2148

Project: Monthly 2025

Run No.: LB136055

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136055BS							
TPH	mg/L	20.0	16.9		84	1	78-114	06/09/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2148

Project: Monthly 2025

Run No.: LB136055

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136055BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	06/09/2025