

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV32	Mercury	4.16	4.0	104	95 - 105	CV	06/11/2025	09:46	LB136099

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV23	Mercury	5.08	5.0	102	90 - 110	CV	06/11/2025	09:50	LB136099
CCV24	Mercury	5.01	5.0	100	90 - 110	CV	06/11/2025	10:23	LB136099
CCV25	Mercury	4.92	5.0	98	90 - 110	CV	06/11/2025	10:35	LB136099

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	504	510	99	95 - 105	P	06/05/2025	15:55	LB136035
	Copper	527	510	103	95 - 105	P	06/05/2025	15:55	LB136035
	Lead	985	1000	98	95 - 105	P	06/05/2025	15:55	LB136035
	Nickel	535	530	101	95 - 105	P	06/05/2025	15:55	LB136035
	Zinc	995	1000	100	95 - 105	P	06/05/2025	15:55	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	6.12	6.0	102	80 - 120	P	06/05/2025	16:09	LB136035
	Copper	21.5	20.0	107	80 - 120	P	06/05/2025	16:09	LB136035
	Lead	14.0	12.0	117	80 - 120	P	06/05/2025	16:09	LB136035
	Nickel	40.1	40.0	100	80 - 120	P	06/05/2025	16:09	LB136035
	Zinc	45.3	40.0	113	80 - 120	P	06/05/2025	16:09	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2470	2500	99	90 - 110	P	06/05/2025	16:50	LB136035
	Copper	1260	1250	101	90 - 110	P	06/05/2025	16:50	LB136035
	Lead	4930	5000	99	90 - 110	P	06/05/2025	16:50	LB136035
	Nickel	2470	2500	99	90 - 110	P	06/05/2025	16:50	LB136035
	Zinc	2540	2500	102	90 - 110	P	06/05/2025	16:50	LB136035
CCV02	Cadmium	2450	2500	98	90 - 110	P	06/05/2025	17:46	LB136035
	Copper	1260	1250	101	90 - 110	P	06/05/2025	17:46	LB136035
	Lead	4870	5000	97	90 - 110	P	06/05/2025	17:46	LB136035
	Nickel	2440	2500	98	90 - 110	P	06/05/2025	17:46	LB136035
	Zinc	2490	2500	100	90 - 110	P	06/05/2025	17:46	LB136035
CCV03	Cadmium	2430	2500	97	90 - 110	P	06/05/2025	19:43	LB136035
	Copper	1230	1250	98	90 - 110	P	06/05/2025	19:43	LB136035
	Lead	4870	5000	97	90 - 110	P	06/05/2025	19:43	LB136035
	Nickel	2440	2500	98	90 - 110	P	06/05/2025	19:43	LB136035
	Zinc	2610	2500	104	90 - 110	P	06/05/2025	19:43	LB136035
CCV04	Cadmium	2500	2500	100	90 - 110	P	06/05/2025	20:28	LB136035
	Copper	1270	1250	102	90 - 110	P	06/05/2025	20:28	LB136035
	Lead	4990	5000	100	90 - 110	P	06/05/2025	20:28	LB136035
	Nickel	2500	2500	100	90 - 110	P	06/05/2025	20:28	LB136035
	Zinc	2630	2500	105	90 - 110	P	06/05/2025	20:28	LB136035
CCV05	Cadmium	2500	2500	100	90 - 110	P	06/05/2025	23:51	LB136035
	Copper	1250	1250	100	90 - 110	P	06/05/2025	23:51	LB136035
	Lead	4970	5000	100	90 - 110	P	06/05/2025	23:51	LB136035
	Nickel	2490	2500	100	90 - 110	P	06/05/2025	23:51	LB136035
	Zinc	2670	2500	107	90 - 110	P	06/05/2025	23:51	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Cadmium	502	510	98	95 - 105	P	06/06/2025	13:16	LB136052
	Copper	527	510	103	95 - 105	P	06/06/2025	13:16	LB136052
	Lead	980	1000	98	95 - 105	P	06/06/2025	13:16	LB136052
	Nickel	533	530	100	95 - 105	P	06/06/2025	13:16	LB136052
	Zinc	991	1000	99	95 - 105	P	06/06/2025	13:16	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Cadmium	6.02	6.0	100	80 - 120	P	06/06/2025	13:29	LB136052
	Copper	21.0	20.0	105	80 - 120	P	06/06/2025	13:29	LB136052
	Lead	13.9	12.0	116	80 - 120	P	06/06/2025	13:29	LB136052
	Nickel	39.7	40.0	99	80 - 120	P	06/06/2025	13:29	LB136052
	Zinc	43.4	40.0	108	80 - 120	P	06/06/2025	13:29	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Cadmium	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Copper	1250	1250	100	90 - 110	P	06/06/2025	14:05	LB136052
	Lead	4910	5000	98	90 - 110	P	06/06/2025	14:05	LB136052
	Nickel	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Zinc	2480	2500	99	90 - 110	P	06/06/2025	14:05	LB136052
CCV02	Cadmium	2470	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Copper	1280	1250	102	90 - 110	P	06/06/2025	14:56	LB136052
	Lead	4960	5000	99	90 - 110	P	06/06/2025	14:56	LB136052
	Nickel	2480	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Zinc	2490	2500	100	90 - 110	P	06/06/2025	14:56	LB136052
CCV03	Cadmium	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Copper	1280	1250	103	90 - 110	P	06/06/2025	15:50	LB136052
	Lead	5000	5000	100	90 - 110	P	06/06/2025	15:50	LB136052
	Nickel	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Zinc	2470	2500	99	90 - 110	P	06/06/2025	15:50	LB136052
CCV04	Cadmium	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Copper	1240	1250	99	90 - 110	P	06/06/2025	16:46	LB136052
	Lead	4890	5000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Nickel	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Zinc	2460	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
CCV05	Cadmium	2390	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Copper	1250	1250	100	90 - 110	P	06/06/2025	17:40	LB136052
	Lead	4810	5000	96	90 - 110	P	06/06/2025	17:40	LB136052
	Nickel	2410	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Zinc	2450	2500	98	90 - 110	P	06/06/2025	17:40	LB136052
CCV06	Cadmium	2300	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Copper	1170	1250	94	90 - 110	P	06/07/2025	03:03	LB136052
	Lead	4580	5000	92	90 - 110	P	06/07/2025	03:03	LB136052
	Nickel	2300	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Zinc	2190	2500	88	90 - 110	P	06/07/2025	03:03	LB136052
CCV07	Cadmium	2330	2500	93	90 - 110	P	06/07/2025	03:40	LB136052
	Copper	1180	1250	94	90 - 110	P	06/07/2025	03:40	LB136052
	Lead	4630	5000	92	90 - 110	P	06/07/2025	03:40	LB136052
	Nickel	2330	2500	93	90 - 110	P	06/07/2025	03:40	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2148
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2148 **SAS No.:** Q2148
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Zinc	2300	2500	92	90 - 110	P	06/07/2025	03:40	LB136052
