



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q2149

Project: Filter Cake

RunNo.: LB135949

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	05/29/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	05/29/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	05/29/2025

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q2149

Project: Filter Cake

RunNo.: LB135966

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.092	0.099	93	85-115	05/30/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	05/30/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	05/30/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.26	0.25	104	90-110	05/30/2025

Initial and Continuing Calibration Verification

Client: Aramark Uniforms

SDG No.: Q2149

Project: Filter Cake

RunNo.: LB135966

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------