

Initial and Continuing Calibration Blank Summary

Client: Aramark Uniforms

SDG No.: Q2149

Project: Filter Cake

RunNo.: LB135966

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.0016	0.0025	J	0.00096	0.005	05/30/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	0.0014	0.0025	J	0.00096	0.005	05/30/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	05/30/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0015	0.0025	J	0.00096	0.005	05/30/2025

Initial and Continuing Calibration Blank Summary

Client: Aramark Uniforms

SDG No.: Q2149

Project: Filter Cake

RunNo.: LB135966

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------