

Duplicate Sample Summary

Client: Aramark Uniforms Project: Filter Cake Client ID: TP03-MHMDUP	SDG No.: Q2149 Sample ID: Q2128-01 Percent Solids for Spike Sample: 84.6
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		05/28/2025

Duplicate Sample Summary

Client:	Aramark Uniforms	SDG No.:	Q2149
Project:	Filter Cake	Sample ID:	Q2143-01
Client ID:	ELI-46-36-25-58-53DUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Cyanide	mg/Kg	+/-20	0.015	J	0.015	J	1	0		05/30/2025
Reactive Sulfide	mg/Kg	+/-20	1.59	J	1.59	J	1	0		05/30/2025

Duplicate Sample Summary

Client: Aramark Uniforms Project: Filter Cake Client ID: WC-1DUP	SDG No.: Q2149 Sample ID: Q2151-04 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Corrosivity	pH	+/-20	6.03		6.04		1	0.17		05/29/2025