

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV23	Mercury	3.69	4.0	92	90 - 110	CV	06/02/2025	11:39	LB135970

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV83	Mercury	5.43	5.0	108	90 - 110	CV	06/02/2025	11:46	LB135970
CCV84	Mercury	5.28	5.0	106	90 - 110	CV	06/02/2025	12:20	LB135970
CCV85	Mercury	5.22	5.0	104	90 - 110	CV	06/02/2025	12:55	LB135970

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms SDG No.: Q2149
 Contract: ARAM01 Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1070	1000	107	90 - 110	P	06/02/2025	12:37	LB135981
	Barium	503	520	97	90 - 110	P	06/02/2025	12:37	LB135981
	Cadmium	535	510	105	90 - 110	P	06/02/2025	12:37	LB135981
	Chromium	546	520	105	90 - 110	P	06/02/2025	12:37	LB135981
	Lead	1040	1000	104	90 - 110	P	06/02/2025	12:37	LB135981
	Selenium	1100	1000	110	90 - 110	P	06/02/2025	12:37	LB135981
	Silver	263	250	105	90 - 110	P	06/02/2025	12:37	LB135981



- 2a -

Client:	<u>Aramark Uniforms</u>	SDG No.:	<u>Q2149</u>				
Contract:	<u>ARAM01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2149</u>	SAS No.:	<u>Q2149</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
LLICV01	Arsenic	23.6	20.0	118	80 - 120	P	06/02/2025	12:59	LB135981
	Barium	94.9	100	95	80 - 120	P	06/02/2025	12:59	LB135981
	Cadmium	5.95	6.0	99	80 - 120	P	06/02/2025	12:59	LB135981
	Chromium	10.2	10.0	102	80 - 120	P	06/02/2025	12:59	LB135981
	Lead	13.3	12.0	111	80 - 120	P	06/02/2025	12:59	LB135981
	Selenium	18.2	20.0	91	80 - 120	P	06/02/2025	12:59	LB135981
	Silver	9.21	10.0	92	80 - 120	P	06/02/2025	12:59	LB135981

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2149

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4980	5000	100	90 - 110	P	06/02/2025	14:14	LB135981
	Barium	9890	10000	99	90 - 110	P	06/02/2025	14:14	LB135981
	Cadmium	2450	2500	98	90 - 110	P	06/02/2025	14:14	LB135981
	Chromium	1000	1000	100	90 - 110	P	06/02/2025	14:14	LB135981
	Lead	4920	5000	98	90 - 110	P	06/02/2025	14:14	LB135981
	Selenium	4820	5000	96	90 - 110	P	06/02/2025	14:14	LB135981
	Silver	1260	1250	101	90 - 110	P	06/02/2025	14:14	LB135981
CCV02	Arsenic	5000	5000	100	90 - 110	P	06/02/2025	15:11	LB135981
	Barium	9760	10000	98	90 - 110	P	06/02/2025	15:11	LB135981
	Cadmium	2450	2500	98	90 - 110	P	06/02/2025	15:11	LB135981
	Chromium	1010	1000	101	90 - 110	P	06/02/2025	15:11	LB135981
	Lead	4910	5000	98	90 - 110	P	06/02/2025	15:11	LB135981
	Selenium	4970	5000	99	90 - 110	P	06/02/2025	15:11	LB135981
	Silver	1250	1250	100	90 - 110	P	06/02/2025	15:11	LB135981
CCV03	Arsenic	4820	5000	96	90 - 110	P	06/02/2025	15:58	LB135981
	Barium	9720	10000	97	90 - 110	P	06/02/2025	15:58	LB135981
	Cadmium	2380	2500	95	90 - 110	P	06/02/2025	15:58	LB135981
	Chromium	976	1000	98	90 - 110	P	06/02/2025	15:58	LB135981
	Lead	4770	5000	95	90 - 110	P	06/02/2025	15:58	LB135981
	Selenium	4850	5000	97	90 - 110	P	06/02/2025	15:58	LB135981
	Silver	1220	1250	98	90 - 110	P	06/02/2025	15:58	LB135981
CCV04	Arsenic	4990	5000	100	90 - 110	P	06/02/2025	16:49	LB135981
	Barium	9540	10000	95	90 - 110	P	06/02/2025	16:49	LB135981
	Cadmium	2460	2500	99	90 - 110	P	06/02/2025	16:49	LB135981
	Chromium	1000	1000	100	90 - 110	P	06/02/2025	16:49	LB135981
	Lead	4910	5000	98	90 - 110	P	06/02/2025	16:49	LB135981
	Selenium	5020	5000	100	90 - 110	P	06/02/2025	16:49	LB135981
	Silver	1210	1250	97	90 - 110	P	06/02/2025	16:49	LB135981
CCV05	Arsenic	4820	5000	96	90 - 110	P	06/02/2025	17:38	LB135981
	Barium	9460	10000	95	90 - 110	P	06/02/2025	17:38	LB135981
	Cadmium	2380	2500	95	90 - 110	P	06/02/2025	17:38	LB135981
	Chromium	991	1000	99	90 - 110	P	06/02/2025	17:38	LB135981
	Lead	4760	5000	95	90 - 110	P	06/02/2025	17:38	LB135981
	Selenium	4850	5000	97	90 - 110	P	06/02/2025	17:38	LB135981

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2149

Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1220	1250	98	90 - 110	P	06/02/2025	17:38	LB135981
CCV06	Arsenic	4860	5000	97	90 - 110	P	06/02/2025	18:24	LB135981
	Barium	9660	10000	97	90 - 110	P	06/02/2025	18:24	LB135981
	Cadmium	2410	2500	96	90 - 110	P	06/02/2025	18:24	LB135981
	Chromium	989	1000	99	90 - 110	P	06/02/2025	18:24	LB135981
	Lead	4800	5000	96	90 - 110	P	06/02/2025	18:24	LB135981
	Selenium	4890	5000	98	90 - 110	P	06/02/2025	18:24	LB135981
	Silver	1230	1250	98	90 - 110	P	06/02/2025	18:24	LB135981
CCV07	Arsenic	4820	5000	96	90 - 110	P	06/02/2025	19:11	LB135981
	Barium	9350	10000	94	90 - 110	P	06/02/2025	19:11	LB135981
	Cadmium	2430	2500	97	90 - 110	P	06/02/2025	19:11	LB135981
	Chromium	1000	1000	100	90 - 110	P	06/02/2025	19:11	LB135981
	Lead	4820	5000	96	90 - 110	P	06/02/2025	19:11	LB135981
	Selenium	4820	5000	96	90 - 110	P	06/02/2025	19:11	LB135981
	Silver	1230	1250	99	90 - 110	P	06/02/2025	19:11	LB135981
CCV08	Arsenic	4890	5000	98	90 - 110	P	06/02/2025	19:57	LB135981
	Barium	9520	10000	95	90 - 110	P	06/02/2025	19:57	LB135981
	Cadmium	2430	2500	97	90 - 110	P	06/02/2025	19:57	LB135981
	Chromium	1000	1000	100	90 - 110	P	06/02/2025	19:57	LB135981
	Lead	4830	5000	97	90 - 110	P	06/02/2025	19:57	LB135981
	Selenium	4920	5000	98	90 - 110	P	06/02/2025	19:57	LB135981
	Silver	1240	1250	99	90 - 110	P	06/02/2025	19:57	LB135981
CCV09	Arsenic	4840	5000	97	90 - 110	P	06/02/2025	20:45	LB135981
	Barium	9650	10000	96	90 - 110	P	06/02/2025	20:45	LB135981
	Cadmium	2410	2500	97	90 - 110	P	06/02/2025	20:45	LB135981
	Chromium	977	1000	98	90 - 110	P	06/02/2025	20:45	LB135981
	Lead	4810	5000	96	90 - 110	P	06/02/2025	20:45	LB135981
	Selenium	4870	5000	98	90 - 110	P	06/02/2025	20:45	LB135981
	Silver	1200	1250	96	90 - 110	P	06/02/2025	20:45	LB135981
CCV10	Arsenic	5.12	5000	0	90 - 110	P	06/02/2025	21:06	LB135981
	Barium	14.6	10000	0	90 - 110	P	06/02/2025	21:06	LB135981
	Cadmium	0.50	2500	0	90 - 110	P	06/02/2025	21:06	LB135981
	Chromium	2.12	1000	0	90 - 110	P	06/02/2025	21:06	LB135981
	Lead	2.30	5000	0	90 - 110	P	06/02/2025	21:06	LB135981

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Aramark Uniforms **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Selenium	9.64	5000	0	90 - 110	P	06/02/2025	21:06	LB135981
	Silver	1.62	1250	0	90 - 110	P	06/02/2025	21:06	LB135981