



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Aramark Uniforms **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.16	0.2	81	70 - 130	CV	06/02/2025	11:50	LB135970
CRI01	Arsenic	20.4	20.0	102	65 - 135	P	06/02/2025	13:08	LB135981
	Barium	92.5	100	92	65 - 135	P	06/02/2025	13:08	LB135981
	Cadmium	5.90	6.0	98	65 - 135	P	06/02/2025	13:08	LB135981
	Chromium	9.95	10.0	100	65 - 135	P	06/02/2025	13:08	LB135981
	Lead	13.1	12.0	109	65 - 135	P	06/02/2025	13:08	LB135981
	Selenium	19.1	20.0	96	65 - 135	P	06/02/2025	13:08	LB135981
	Silver	9.30	10.0	93	65 - 135	P	06/02/2025	13:08	LB135981