



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Aramark Uniforms				SDG No.:	Q2149				
Contract:	ARAM01	Lab Code:	CHEM		Case No.:	Q2149	SAS No.:	Q2149		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB23	Mercury	0.20	+/-0.20	U	0.20	CV	06/02/2025	11:44	LB135970	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Aramark Uniforms

SDG No.: Q2149

Contract: ARAM01

Lab Code: CHEM

Case No.: Q2149

SAS No.: Q2149

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB83	Mercury	0.20	+/-0.20	U	0.20	CV	06/02/2025	11:48	LB135970
CCB84	Mercury	0.20	+/-0.20	U	0.20	CV	06/02/2025	12:22	LB135970
CCB85	Mercury	0.20	+/-0.20	U	0.20	CV	06/02/2025	12:59	LB135970

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Aramark Uniforms</u>		SDG No.: <u>Q2149</u>							
Contract: <u>ARAM01</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q2149</u>		SAS No.: <u>Q2149</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	13:03	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	13:03	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	13:03	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	13:03	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	13:03	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	13:03	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	13:03	LB135981

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Aramark Uniforms</u>		SDG No.: <u>Q2149</u>							
Contract: <u>ARAM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2149</u>	SAS No.: <u>Q2149</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	14:28	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	14:28	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	14:28	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	14:28	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	14:28	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	14:28	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	14:28	LB135981
CCB02	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	15:16	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	15:16	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	15:16	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	15:16	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	15:16	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	15:16	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	15:16	LB135981
CCB03	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	16:02	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	16:02	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	16:02	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	16:02	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	16:02	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	16:02	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	16:02	LB135981
CCB04	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	16:55	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	16:55	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	16:55	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	16:55	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	16:55	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	16:55	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	16:55	LB135981
CCB05	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	17:42	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	17:42	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	17:42	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	17:42	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	17:42	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	17:42	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	17:42	LB135981
CCB06	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	18:28	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	18:28	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	18:28	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	18:28	LB135981

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Aramark Uniforms</u>		SDG No.: <u>Q2149</u>							
Contract: <u>ARAM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2149</u>	SAS No.: <u>Q2149</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.52	+/-12.0	J	12.0	P	06/02/2025	18:28	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	18:28	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	18:28	LB135981
CCB07	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	19:15	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	19:15	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	19:15	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	19:15	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	19:15	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	19:15	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	19:15	LB135981
CCB08	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	20:01	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	20:01	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	20:01	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	20:01	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	20:01	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	20:01	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	20:01	LB135981
CCB09	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	20:49	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	20:49	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	20:49	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	20:49	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	20:49	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	20:49	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	20:49	LB135981
CCB10	Arsenic	20.0	+/-20.0	U*	20.0	P	06/02/2025	21:11	LB135981
	Barium	100	+/-100	U*	100	P	06/02/2025	21:11	LB135981
	Cadmium	6.00	+/-6.00	U*	6.00	P	06/02/2025	21:11	LB135981
	Chromium	10.0	+/-10.0	U*	10.0	P	06/02/2025	21:11	LB135981
	Lead	12.0	+/-12.0	U*	12.0	P	06/02/2025	21:11	LB135981
	Selenium	20.0	+/-20.0	U*	20.0	P	06/02/2025	21:11	LB135981
	Silver	10.0	+/-10.0	U*	10.0	P	06/02/2025	21:11	LB135981

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Aramark Uniforms</u>				SDG No.:	<u>Q2149</u>			
Contract:	<u>ARAM01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q2149</u>	SAS No.:	<u>Q2149</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number