

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149
Matrix: Water **Sample ID:** Q2151-04 **Client ID:** WC-1DUP
Percent Solids for Sample: NA **Duplicate ID** Q2151-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1560		1670		7		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	11.7	J	13.8	J	16		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Aramark Uniforms **Level:** LOW **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149
Matrix: Water **Sample ID:** Q2151-04MS **Client ID:** WC-1MSD
Percent Solids for Sample: NA **Duplicate ID** Q2151-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3560		3720		4		P
Barium	ug/L	20	2620		2520		4		P
Cadmium	ug/L	20	882		906		3		P
Chromium	ug/L	20	1780		1850		4		P
Lead	ug/L	20	4220		4340		3		P
Mercury	ug/L	20	39.3		38.6		2		CV
Selenium	ug/L	20	8860		9200		4		P
Silver	ug/L	20	319		333		4		P