

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Aramark Uniforms **SDG No.:** Q2149
Contract: ARAM01 **Lab Code:** CHEM **Case No.:** Q2149 **SAS No.:** Q2149

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168214BS							
Arsenic	ug/L	4000	4080		102	80 - 120	P
Barium	ug/L	1000	956		96	80 - 120	P
Cadmium	ug/L	1000	1000		100	80 - 120	P
Chromium	ug/L	2000	2090		104	80 - 120	P
Lead	ug/L	5000	4910		98	80 - 120	P
Selenium	ug/L	10000	10400		104	80 - 120	P
Silver	ug/L	380	379		100	80 - 120	P



- 7 -

Client:	<u>Aramark Uniforms</u>	SDG No.:	<u>Q2149</u>				
Contract:	ARAM01	Lab Code:	CHEM	Case No.:	Q2149	SAS No.:	Q2149

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168230BS Mercury	ug/L	4.0	4.30		108	80 - 120	CV