

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2136-05MS	Client Sample ID:	OR-646-COMP-52MS					DataFile:	BF142575.D		
Pyridine	500	0	380	ug/L	76				10	109	
1,4-Dichlorobenzene	500	0	410	ug/L	82				55	125	
2-Methylphenol	500	0	430	ug/L	86				60	131	
3+4-Methylphenols	500	0	400	ug/L	80				54	136	
Hexachloroethane	500	0	400	ug/L	80				19	146	
Nitrobenzene	500	0	440	ug/L	88				62	112	
Hexachlorobutadiene	500	0	420	ug/L	84				52	125	
2,4,6-Trichlorophenol	500	0	460	ug/L	92				78	112	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				71	111	
2,4-Dinitrotoluene	500	0	480	ug/L	96				74	137	
Hexachlorobenzene	500	0	470	ug/L	94				72	115	
Pentachlorophenol	1000	0	830	ug/L	83				52	162	