

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2136-05MSD	Client Sample ID:	OR-646-COMP-52MSD					DataFile:	BF142576.D		
Pyridine	500	0	350	ug/L	70	8			10	109	20
1,4-Dichlorobenzene	500	0	380	ug/L	76	8			55	125	20
2-Methylphenol	500	0	390	ug/L	78	10			60	131	20
3+4-Methylphenols	500	0	380	ug/L	76	5			54	136	20
Hexachloroethane	500	0	380	ug/L	76	5			19	146	20
Nitrobenzene	500	0	410	ug/L	82	7			62	112	20
Hexachlorobutadiene	500	0	380	ug/L	76	10			52	125	20
2,4,6-Trichlorophenol	500	0	420	ug/L	84	9			78	112	20
2,4,5-Trichlorophenol	500	0	420	ug/L	84	9			71	111	20
2,4-Dinitrotoluene	500	0	430	ug/L	86	11			74	137	20
Hexachlorobenzene	500	0	430	ug/L	86	9			72	115	20
Pentachlorophenol	1000	0	770	ug/L	77	8			52	162	20