

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: 8270E

DataFile: BF142592.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168219BS	Pyridine	50	40.1	ug/L	80				29	97	
	1,4-Dichlorobenzene	50	42.1	ug/L	84				76	103	
	2-Methylphenol	50	43.7	ug/L	87				69	109	
	3+4-Methylphenols	50	42.0	ug/L	84				67	106	
	Hexachloroethane	50	42.1	ug/L	84				76	118	
	Nitrobenzene	50	42.1	ug/L	84				58	106	
	Hexachlorobutadiene	50	41.5	ug/L	83				69	101	
	2,4,6-Trichlorophenol	50	42.7	ug/L	85				61	110	
	2,4,5-Trichlorophenol	50	43.7	ug/L	87				70	106	
	2,4-Dinitrotoluene	50	47.7	ug/L	95				60	115	
	Hexachlorobenzene	50	42.5	ug/L	85				73	106	
	Pentachlorophenol	100	85.2	ug/L	85				47	114	