

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: ARAM01

Lab Code: CHEM

SAS No.: Q2149

SDG NO.: Q2149

Lab File ID: BF142589.D

DFTPP Injection Date: 06/02/2025

Instrument ID: BNA_F

DFTPP Injection Time: 11:01

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	27
68	Less than 2.0% of mass 69	0.4 (1.8) 1
69	Mass 69 relative abundance	24.1
70	Less than 2.0% of mass 69	0.1 (0.6) 1
127	10.0 - 80.0% of mass 198	33.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	4.9
275	10.0 - 60.0% of mass 198	21.6
365	Greater than 1% of mass 198	2.9
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142590.D	06/02/2025	11:30
PB168219BL	PB168219BL	BF142591.D	06/02/2025	11:59
PB168219BS	PB168219BS	BF142592.D	06/02/2025	12:28
PB168190TB	PB168190TB	BF142593.D	06/02/2025	12:57
FILTER-CAKE	Q2149-01	BF142599.D	06/02/2025	16:03
FILTER-CAKEDL	Q2149-01DL	BF142600.D	06/02/2025	17:11