

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ARAM01

Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG No.: Q2149

Instrument ID: MSVOA_X Calibration Date/Time: 05/30/2025 10:14

Lab File ID: VX046409.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.722		4.49	20
1,1-Dichloroethene	0.593	0.644		8.6	20
2-Butanone	0.543	0.600		10.5	20
Carbon Tetrachloride	0.544	0.593		9.01	20
Chloroform	1.271	1.387		9.13	20
Benzene	1.417	1.556		9.81	20
1,2-Dichloroethane	0.612	0.666		8.82	20
Trichloroethene	0.341	0.373		9.38	20
Tetrachloroethene	0.354	0.397		12.15	20
Chlorobenzene	1.094	1.183	0.3	8.14	20
1,2-Dichloroethane-d4	0.932	0.920		-1.29	20
Dibromofluoromethane	0.360	0.375		4.17	20
Toluene-d8	1.246	1.233		-1.04	20
4-Bromofluorobenzene	0.478	0.485		1.46	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.