

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q2149	OrderDate:	5/28/2025 1:31:11 PM
Client:	Aramark Uniforms	Project:	Waste Characterization
Contact:	Jarrod Mills	Location:	L31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2149-01	FILTER-CAKE	TCLP	TCLP VOA Group1	8260D	05/28/25		05/30/25	05/28/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units

Client ID:

0

Total Voc :

Total Concentration:



QC SUMMARY

Surrogate Summary

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2149-01	FILTER-CAKE	1,2-Dichloroethane-d4	50	54.9	110	74	125
		Dibromofluoromethane	50	51.9	104	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	51.3	103	77	121

Surrogate Summary

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0530WBL01	VX0530WBL01	1,2-Dichloroethane-d4	50	54.0	108	74	125
		Dibromofluoromethane	50	50.2	100	75	124
		Toluene-d8	50	50.1	100	86	113
		4-Bromofluorobenzene	50	50.8	102	77	121
VX0530WBS01	VX0530WBS01	1,2-Dichloroethane-d4	50	49.9	100	74	125
		Dibromofluoromethane	50	50.4	101	75	124
		Toluene-d8	50	47.8	96	86	113
		4-Bromofluorobenzene	50	49.4	99	77	121
VX0530WBSD0	VX0530WBSD01	1,2-Dichloroethane-d4	50	50.5	101	74	125
		Dibromofluoromethane	50	50.3	101	75	124
		Toluene-d8	50	47.5	95	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260-Low

Datafile : VX046412.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0530WBS01	Vinyl chloride	20	16.5	ug/L	83			65	117	
	1,1-Dichloroethene	20	17.9	ug/L	90			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	18.7	ug/L	94			77	113	
	Chloroform	20	20.5	ug/L	103			79	113	
	Benzene	20	19.3	ug/L	97			82	109	
	1,2-Dichloroethane	20	19.9	ug/L	100			80	115	
	Trichloroethene	20	18.8	ug/L	94			77	113	
	Tetrachloroethene	20	18.4	ug/L	92			67	123	
	Chlorobenzene	20	19.1	ug/L	96			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260-Low

Datafile : VX046413.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0530WBSD01	Vinyl chloride	20	16.2	ug/L	81	2		65	117	20
	1,1-Dichloroethene	20	17.1	ug/L	86	5		74	110	20
	2-Butanone	100	110	ug/L	110	0		65	122	20
	Carbon Tetrachloride	20	18.6	ug/L	93	1		77	113	20
	Chloroform	20	20.9	ug/L	104	1		79	113	20
	Benzene	20	19.2	ug/L	96	1		82	109	20
	1,2-Dichloroethane	20	20.0	ug/L	100	0		80	115	20
	Trichloroethene	20	18.7	ug/L	94	0		77	113	20
	Tetrachloroethene	20	19.2	ug/L	96	4		67	123	20
	Chlorobenzene	20	19.5	ug/L	98	2		82	109	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0530WBL01

Lab Name: CHEMTECH

Contract: ARAM01

Lab Code: CHEM Case No.: Q2149

SAS No.: Q2149 SDG NO.: Q2149

Lab File ID: VX046411.D

Lab Sample ID: VX0530WBL01

Date Analyzed: 05/30/2025

Time Analyzed: 11:06

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0530WBS01	VX0530WBS01	VX046412.D	05/30/2025
VX0530WBSD01	VX0530WBSD01	VX046413.D	05/30/2025
FILTER-CAKE	Q2149-01	VX046418.D	05/30/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ARAM01
Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG NO.: Q2149
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: ARAM01
Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG NO.: Q2149
Lab File ID: VX046408.D BFB Injection Date: 05/30/2025
Instrument ID: MSVOA_X BFB Injection Time: 08:20
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.6
75	30.0 - 60.0% of mass 95	57.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.2
173	Less than 2.0% of mass 174	0.9 (1.3) 1
174	50.0 - 100.0% of mass 95	68.3
175	5.0 - 9.0% of mass 174	5.2 (7.7) 1
176	95.0 - 101.0% of mass 174	65 (95.2) 1
177	5.0 - 9.0% of mass 176	4.2 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046409.D	05/30/2025	10:14
VX0530WBL01	VX0530WBL01	VX046411.D	05/30/2025	11:06
VX0530WBS01	VX0530WBS01	VX046412.D	05/30/2025	11:29
VX0530WBSD01	VX0530WBSD01	VX046413.D	05/30/2025	11:55
FILTER-CAKE	Q2149-01	VX046418.D	05/30/2025	13:52

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARAM01
 Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG NO.: Q2149
 Lab File ID: VX046409.D Date Analyzed: 05/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91458	5.54	155659	6.75	130840	10.05
UPPER LIMIT	182916	6.038	311318	7.251	261680	10.549
LOWER LIMIT	45729	5.038	77829.5	6.251	65420	9.549
EPA SAMPLE NO.						
FILTER-CAKE	62556	5.54	124459	6.76	115572	10.05
VX0530WBL01	64533	5.55	129905	6.76	123011	10.05
VX0530WBS01	86619	5.55	153415	6.76	134836	10.05
VX0530WBSD01	81578	5.54	145807	6.76	127712	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARAM01
 Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG NO.: Q2149
 Lab File ID: VX046409.D Date Analyzed: 05/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	64536	12.018				
UPPER LIMIT	129072	12.518				
LOWER LIMIT	32268	11.518				
EPA SAMPLE NO.						
FILTER-CAKE	52207	12.02				
VX0530WBL01	51208	12.02				
VX0530WBS01	64120	12.02				
VX0530WBSD01	60852	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	Aramark Uniforms			Date Collected:	05/28/25	
Project:	Waste Characterization			Date Received:	05/28/25	
Client Sample ID:	FILTER-CAKE			SDG No.:	Q2149	
Lab Sample ID:	Q2149-01			Matrix:	TCLP	
Analytical Method:	8260D			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	TCLP VOA Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :	SW5035					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046418.D	50		05/30/25 13:52	VX053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	13.0	U	13.0	250	ug/L
75-35-4	1,1-Dichloroethene	11.5	U	11.5	250	ug/L
78-93-3	2-Butanone	49.0	U	49.0	1300	ug/L
56-23-5	Carbon Tetrachloride	12.5	U	12.5	250	ug/L
67-66-3	Chloroform	12.5	U	12.5	250	ug/L
71-43-2	Benzene	7.50	U	7.50	250	ug/L
107-06-2	1,2-Dichloroethane	11.0	U	11.0	250	ug/L
79-01-6	Trichloroethene	4.70	U	4.70	250	ug/L
127-18-4	Tetrachloroethene	11.5	U	11.5	250	ug/L
108-90-7	Chlorobenzene	6.00	U	6.00	250	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.9		74 - 125	110%	SPK: 50
1868-53-7	Dibromofluoromethane	51.9		75 - 124	104%	SPK: 50
2037-26-5	Toluene-d8	51.1		86 - 113	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		77 - 121	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	62600	5.544			
540-36-3	1,4-Difluorobenzene	124000	6.757			
3114-55-4	Chlorobenzene-d5	116000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	52200	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046418.D
 Acq On : 30 May 2025 13:52
 Operator : JC/MD
 Sample : Q2149-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FILTER-CAKE

Quant Time: May 30 23:27:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	62556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124459	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64010	54.886	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.780%	
35) Dibromofluoromethane	5.379	113	46528	51.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	8.647	98	158494	51.094	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.180%	
62) 4-Bromofluorobenzene	11.079	95	61081	51.334	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.660%	
Target Compounds						Qvalue
16) Acetone	2.386	43	2562	5.479	ug/l	# 85


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046418.D
Acq On : 30 May 2025 13:52
Operator : JC/MD
Sample : Q2149-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FILTER-CAKE

Quant Time: May 30 23:27:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX046418.D
Acq: 30 May 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
FILTER-CAKE


Tgt Ion:168 Resp: 62556
Ion Ratio Lower Upper
168 100
99 72.2 54.9 82.3



#16
Acetone
Concen: 5.479 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046418.D
Acq: 30 May 2025 13:52

Tgt Ion: 43 Resp: 2562
Ion Ratio Lower Upper
43 100
58 18.6 21.2 31.8#





#33

1,2-Dichloroethane-d4

Concen: 54.886 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046418.D

Acq: 30 May 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :





FILTER-CAKE

Tgt Ion: 65 Resp: 64010

Ion Ratio Lower Upper

65 100

67 48.5 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046418.D

Acq: 30 May 2025 13:52




Tgt Ion: 114 Resp: 124459


Ion Ratio Lower Upper

114 100

63 25.4 0.0 49.2

88 17.2 0.0 33.6





#35

Dibromofluoromethane

Concen: 51.915 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046418.D


Acq: 30 May 2025 13:52

Instrument :

MSVOA_X




ClientSampleId :

FILTER-CAKE



Tgt Ion:113 Resp: 46528

Ion	Ratio	Lower	Upper
113	100		
111	103.5	83.1	124.7
192	16.9	13.3	19.9



#50

Toluene-d8

Concen: 51.094 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. -0.000 min


Lab File: VX046418.D

Acq: 30 May 2025 13:52

Tgt Ion: 98 Resp: 158494

Ion	Ratio	Lower	Upper
98	100		
100	64.1	53.5	80.3









#62
4-Bromofluorobenzene
Concen: 51.334 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046418.D
Acq: 30 May 2025 13:52




Instrument :
MSVOA_X
ClientSampleId :
FILTER-CAKE


Tgt Ion: 95 Resp: 61081
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 63.7 0.0 131.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046418.D
Acq: 30 May 2025 13:52

Tgt Ion: 117 Resp: 115572
Ion Ratio Lower Upper
117 100
82 63.4 50.6 76.0
119 32.6 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046418.D


Acq: 30 May 2025 13:52

Instrument :

MSVOA_X

ClientSampleId :

FILTER-CAKE





Tgt Ion:152 Resp: 52207

Ion Ratio Lower Upper

152 100

115 65.9 46.9 140.7

150 155.6 0.0 351.0





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARAM01
 Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG No.: Q2149
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD	
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2	
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9	
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8	
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6	
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3	
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3	
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5	
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3	
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8	
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7	
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6	
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7	
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2	
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M
 Title : SW846 8260
 Last Update : Tue May 06 07:12:22 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX046047.D 5 =VX046046.D 20 =VX046041.D 50 =VX046042.D 100 =VX046043.D 150 =VX046044.D

Compound		1	5	20	50	100	150	Avg	%RSD

1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.658	0.639	0.697	0.864	0.859	0.875	0.765	14.62
3) P	Chloromethane	0.694	0.679	0.727	0.775	0.787	0.791	0.742	6.58
4) C	Vinyl Chloride	0.673	0.619	0.660	0.710	0.727	0.755	0.691	7.17#
5) T	Bromomethane		0.305	0.296	0.326	0.340	0.334	0.320	5.84
6) T	Chloroethane	0.467	0.368	0.354	0.378	0.329	0.317	0.369	14.44
7) T	Trichlorofluor...	1.064	0.990	1.035	1.068	0.983	0.985	1.021	3.92
8) T	Diethyl Ether	0.403	0.311	0.340	0.337	0.338	0.355	0.347	8.83
9) T	1,1,2-Trichlor...	0.633	0.610	0.628	0.641	0.629	0.648	0.632	2.10
10) T	Methyl Iodide		0.608	0.767	0.806	0.793	0.763	0.747	10.68
11) T	Tert butyl alc...		0.114	0.122	0.129	0.144	0.146	0.131	10.45
12) CM	1,1-Dichloroet...	0.594	0.567	0.565	0.601	0.607	0.625	0.593	3.94#
13) T	Acrolein		0.117	0.158	0.152	0.154	0.163	0.149	12.33
14) T	Allyl chloride	1.052	1.058	1.127	1.179	1.187	1.196	1.133	5.75
15) T	Acrylonitrile	0.363	0.345	0.378	0.388	0.381	0.390	0.374	4.52
16) T	Acetone	0.380	0.408	0.361	0.362	0.361	0.370	0.374	4.90
17) T	Carbon Disulfide	1.423	1.141	1.295	1.455	1.522	1.597	1.406	11.68
18) T	Methyl Acetate	1.006	0.816	0.814	0.848	0.845	0.875	0.867	8.27
19) T	Methyl tert-bu...	1.949	1.908	2.044	2.160	2.172	2.239	2.079	6.39
20) T	Methylene Chlo...	0.853	0.689	0.689	0.684	0.691	0.691	0.716	9.39
21) T	trans-1,2-Dich...	0.604	0.557	0.573	0.610	0.612	0.622	0.596	4.27
22) T	Diisopropyl ether	2.095	1.924	2.219	2.278	2.295	2.321	2.189	6.97
23) T	Vinyl Acetate	1.660	1.698	1.928	2.048	2.082	2.134	1.925	10.52
24) P	1,1-Dichloroet...	1.116	1.154	1.233	1.263	1.263	1.286	1.219	5.60
25) T	2-Butanone	0.495	0.539	0.540	0.555	0.558	0.569	0.543	4.77
26) T	2,2-Dichloropr...	0.965	0.850	0.910	0.957	1.003	1.039	0.954	7.03
27) T	cis-1,2-Dichlo...	0.719	0.642	0.716	0.737	0.738	0.755	0.718	5.55
28) T	Bromochloromet...	0.576	0.553	0.628	0.578	0.595	0.590	0.587	4.28
29) T	Tetrahydrofuran	0.318	0.318	0.340	0.350	0.351	0.362	0.340	5.36
30) C	Chloroform	1.265	1.199	1.287	1.296	1.277	1.300	1.271	2.95#
31) T	Cyclohexane		1.059	1.090	1.128	1.128	1.150	1.111	3.26
32) T	1,1,1-Trichlor...	1.015	1.013	1.106	1.131	1.155	1.188	1.101	6.60
33) S	1,2-Dichloroet...		0.935	0.953	0.910	0.930	0.932	0.932	1.65
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.354	0.359	0.355	0.364	0.368	0.360	1.70
36) T	1,1-Dichloropr...	0.493	0.462	0.463	0.495	0.483	0.505	0.484	3.68
37) T	Ethyl Acetate	0.586	0.582	0.569	0.611	0.609	0.631	0.598	3.83
38) T	Carbon Tetrach...	0.541	0.505	0.528	0.558	0.552	0.577	0.544	4.61
39) T	Methylcyclohexane	0.627	0.587	0.596	0.641	0.627	0.658	0.623	4.34
40) TM	Benzene	1.348	1.337	1.426	1.474	1.441	1.477	1.417	4.30
41) T	Methacrylonitrile	0.233	0.288	0.318	0.346	0.343	0.348	0.313	14.50
42) TM	1,2-Dichloroet...	0.579	0.594	0.632	0.627	0.611	0.625	0.612	3.45
43) T	Isopropyl Acetate	0.764	0.826	0.905	0.963	0.982	1.030	0.912	11.05
44) TM	Trichloroethene	0.324	0.315	0.344	0.355	0.345	0.362	0.341	5.28
45) C	1,2-Dichloropr...	0.317	0.324	0.356	0.371	0.368	0.378	0.352	7.37#
46) T	Dibromomethane	0.263	0.262	0.285	0.287	0.280	0.289	0.278	4.35
47) T	Bromodichlorom...	0.485	0.498	0.557	0.577	0.573	0.594	0.547	8.22
48) T	Methyl methacr...	0.370	0.426	0.465	0.502	0.500	0.531	0.466	12.75
49) T	1,4-Dioxane	0.007	0.009	0.009	0.009	0.009	0.010	0.009	8.45
50) S	Toluene-d8		1.221	1.246	1.223	1.266	1.275	1.246	1.95
51) T	4-Methyl-2-Pen...	0.561	0.555	0.620	0.634	0.630	0.631	0.605	6.04
52) CM	Toluene	0.803	0.838	0.884	0.898	0.885	0.904	0.869	4.54#
53) T	t-1,3-Dichloro...	0.371	0.406	0.468	0.528	0.555	0.591	0.487	17.86
54) T	cis-1,3-Dichlo...	0.423	0.469	0.531	0.578	0.602	0.623	0.538	14.61
55) T	1,1,2-Trichlor...	0.308	0.337	0.349	0.354	0.351	0.356	0.343	5.30
56) T	Ethyl methacry...	0.377	0.508	0.540	0.595	0.617	0.639	0.546	17.58

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M

57)	T	1,3-Dichloropr...	0.610	0.601	0.618	0.623	0.613	0.627	0.615	1.53
58)	T	2-Chloroethyl ...	0.230	0.247	0.270	0.307	0.303	0.313	0.278	12.41
59)	T	2-Hexanone	0.385	0.414	0.466	0.473	0.477	0.473	0.448	8.69
60)	T	Dibromochlorom...	0.306	0.326	0.378	0.400	0.415	0.431	0.376	13.28
61)	T	1,2-Dibromoethane	0.322	0.333	0.359	0.373	0.368	0.381	0.356	6.54
62)	S	4-Bromofluorob...		0.464	0.455	0.470	0.500	0.500	0.478	4.41
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.347	0.323	0.390	0.375	0.345	0.344	0.354	6.84
65)	PM	Chlorobenzene	1.131	1.046	1.093	1.098	1.085	1.114	1.094	2.65
66)	T	1,1,1,2-Tetrac...	0.369	0.341	0.365	0.390	0.382	0.395	0.374	5.22
67)	C	Ethyl Benzene	1.803	1.816	1.919	2.022	1.979	2.036	1.929	5.24#
68)	T	m/p-Xylenes	0.648	0.678	0.706	0.740	0.721	0.740	0.706	5.21
69)	T	o-Xylene	0.642	0.639	0.688	0.727	0.706	0.726	0.688	5.75
70)	T	Styrene	0.951	1.012	1.135	1.219	1.214	1.230	1.127	10.56
71)	P	Bromoform	0.234	0.236	0.270	0.304	0.312	0.327	0.281	14.17
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.789	3.562	3.843	4.130	3.876	4.156	3.893	5.72
74)	T	N-amyl acetate	1.715	1.652	1.846	2.067	2.068	2.192	1.924	11.32
75)	P	1,1,2,2-Tetrac...	1.552	1.350	1.315	1.338	1.284	1.345	1.364	6.97
76)	T	1,2,3-Trichlor...	1.405	1.167	1.151	1.187	1.131	1.181	1.204	8.36
77)	T	Bromobenzene	0.926	0.862	0.896	0.928	0.883	0.928	0.904	3.08
78)	T	n-propylbenzene	4.272	4.186	4.394	4.854	4.583	4.868	4.526	6.45
79)	T	2-Chlorotoluene	3.184	2.748	2.832	2.994	2.805	2.953	2.919	5.45
80)	T	1,3,5-Trimethy...	3.036	3.053	3.275	3.487	3.255	3.405	3.252	5.60
81)	T	trans-1,4-Dich...		0.269	0.335	0.385	0.410	0.449	0.370	18.91
82)	T	4-Chlorotoluene	3.226	2.939	3.196	3.430	3.255	3.379	3.238	5.32
83)	T	tert-Butylbenzene	3.341	3.098	3.115	3.435	3.255	3.411	3.276	4.44
84)	T	1,2,4-Trimethy...	3.150	3.034	3.274	3.522	3.335	3.444	3.293	5.52
85)	T	sec-Butylbenzene	3.708	3.767	3.937	4.282	4.095	4.343	4.022	6.55
86)	T	p-Isopropyltol...	3.025	3.084	3.206	3.555	3.450	3.599	3.320	7.44
87)	T	1,3-Dichlorobe...	1.619	1.558	1.633	1.701	1.656	1.729	1.649	3.71
88)	T	1,4-Dichlorobe...	1.817	1.606	1.629	1.693	1.639	1.722	1.684	4.64
89)	T	n-Butylbenzene	2.443	2.650	2.748	3.147	3.139	3.346	2.912	12.00
90)	T	Hexachloroethane	0.511	0.523	0.551	0.622	0.622	0.680	0.585	11.44
91)	T	1,2-Dichlorobe...	1.710	1.577	1.613	1.696	1.634	1.702	1.655	3.34
92)	T	1,2-Dibromo-3-...	0.259	0.248	0.299	0.322	0.329	0.356	0.302	13.89
93)	T	1,2,4-Trichlor...	0.862	0.842	0.861	0.981	1.035	1.123	0.951	12.03
94)	T	Hexachlorobuta...	0.393	0.414	0.394	0.427	0.418	0.445	0.415	4.79
95)	T	Naphthalene	3.499	2.929	3.204	3.613	3.690	3.984	3.487	10.69
96)	T	1,2,3-Trichlor...	0.941	0.846	0.921	1.019	1.051	1.107	0.981	9.74

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	25.640%#		
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	25.600%#		
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	26.400%#		
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	26.480%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	27845m	11.039	ug/l	
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.001 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:168 Resp: 8367
Ion Ratio Lower Upper
168 100
99 72.7 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 12.638 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 85 Resp: 23341
Ion Ratio Lower Upper
85 100
87 33.5 15.7 47.1









#3
Chloromethane
Concen: 13.371 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 50 Resp: 24310
Ion Ratio Lower Upper
50 100
52 33.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 13.868 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 62 Resp: 22096
Ion Ratio Lower Upper
62 100
64 34.9 25.2 37.8





#5
Bromomethane
Concen: 12.279 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 94 Resp: 9920
Ion Ratio Lower Upper
94 100
96 98.1 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 14.414 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 64 Resp: 11832
Ion Ratio Lower Upper
64 100
66 32.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 14.271 ug/l
RT: 1.873 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:101 Resp: 34630
Ion Ratio Lower Upper
101 100
103 63.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 14.278 ug/l
RT: 2.129 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 74 Resp: 11392
Ion Ratio Lower Upper
74 100
45 103.7 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 14.499 ug/l

RT: 2.324 min Scan# 202

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	101	Resp:	2102
Ion Ratio	Lower	Upper	
101	100		
85	46.6	38.6	58.0
151	70.4	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 15.070 ug/l




RT: 2.446 min Scan# 223


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	142	Resp:	25661
Ion Ratio	Lower	Upper	
142	100		
127	53.0	41.7	62.5
141	14.7	11.5	17.3





#11

Tert butyl alcohol

Concen: 68.662 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 59 Resp: 20370

Ion Ratio Lower Upper

59 100





57 11.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 13.385 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 96 Resp: 18899


Ion Ratio Lower Upper

96 100

61 179.8 146.2 219.2

98 61.8 51.0 76.6





#13

Acrolein

Concen: 76.689 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 56 Resp: 2642

Ion Ratio Lower Upper

56 100





55 74.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 14.133 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 37708


Ion Ratio Lower Upper

41 100

39 75.0 60.6 90.8

76 30.6 24.9 37.3





#15

Acrylonitrile

Concen: 71.312 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	53	Resp:	63290
Ion	Ratio	Lower	Upper
53	100		
52	83.5	65.3	97.9
51	38.4	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 70.884 ug/l




RT: 2.379 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	43	Resp:	60363
Ion	Ratio	Lower	Upper
43	100		
58	26.0	21.2	31.8





#17

Carbon Disulfide

Concen: 13.251 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 76 Resp: 4334

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 13.397 ug/l

RT: 2.702 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 43 Resp: 27234

Ion Ratio Lower Upper

43 100

74 20.7 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 13.890 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion: 73 Resp: 6840
Ion Ratio Lower Upper
73 100
57 22.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 13.362 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 84 Resp: 23050
Ion Ratio Lower Upper
84 100
49 146.0 113.9 170.9
51 43.8 33.5 50.3
86 62.2 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 13.294 ug/l
RT: 3.086 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 96 Resp: 1916
Ion Ratio Lower Upper
96 100
61 155.6 119.5 179.3
98 63.2 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 14.896 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 45 Resp: 74257
Ion Ratio Lower Upper
45 100
43 78.2 66.6 100.0
87 24.0 19.8 29.6
59 11.7 8.6 12.8





#23
Vinyl Acetate
Concen: 73.230 ug/l
RT: 3.720 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020




Tgt Ion: 43 Resp: 322598
Ion Ratio Lower Upper
43 100
86 9.3 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#24
1,1-Dichloroethane
Concen: 14.082 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35



Tgt Ion: 63 Resp: 41257
Ion Ratio Lower Upper
63 100
98 7.9 3.6 10.8
100 5.0 2.1 6.3





#25

2-Butanone

Concen: 73.489 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 9033

Ion Ratio Lower Upper

43 100





72 23.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 13.729 ug/l

RT: 4.470 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 77 Resp: 30472

Ion Ratio Lower Upper

77 100

97 20.9 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 13.791 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 96 Resp: 23970

Ion Ratio Lower Upper

96 100

61 158.3 0.0 322.8





98 63.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 13.497 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 49 Resp: 21028


Ion Ratio Lower Upper

49 100

129 1.9 0.0 4.0

130 70.9 56.2 84.2





#29

Tetrahydrofuran

Concen: 71.175 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	42	Resp:	56819
Ion Ratio	Lower	Upper	
42	100		
72	42.3	33.7	50.5
71	39.2	31.4	47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 14.144 ug/l




RT: 5.086 min Scan# 656


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	83	Resp:	43065
Ion Ratio	Lower	Upper	
83	100		
85	65.7	49.3	73.9









#31
Cyclohexane
Concen: 14.769 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 56 Resp: 36470
Ion Ratio Lower Upper
56 100
69 28.8 24.4 36.6
84 78.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 14.092 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 97 Resp: 37003
Ion Ratio Lower Upper
97 100
99 62.8 51.8 77.6
61 49.5 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 12.820 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 65 Resp: 31909

Ion Ratio Lower Upper

65 100





67 49.9 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:114 Resp: 147096


Ion Ratio Lower Upper

114 100

63 25.2 0.0 49.2


88 18.0 0.0 33.6





#35
Dibromofluoromethane
Concen: 12.804 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 113 Resp: 2111
Ion Ratio Lower Upper
113 100
111 103.0 83.1 124.7
192 17.0 13.3 19.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#36
1,1-Dichloropropene
Concen: 14.231 ug/l
RT: 5.684 min Scan# 754
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 75 Resp: 27242
Ion Ratio Lower Upper
75 100
110 31.6 16.3 48.9
77 31.9 24.3 36.5





#37

Ethyl Acetate

Concen: 13.931 ug/l

RT: 4.720 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion:	43	Resp:	3345
Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.3	15.5
70	10.9	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 14.137 ug/l




RT: 5.671 min Scan# 752


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:	117	Resp:	31091
Ion	Ratio	Lower	Upper
117	100		
119	95.6	75.2	112.8
121	32.1	24.2	36.4





#39

Methylcyclohexane

Concen: 14.623 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 83	Resp: 3507
Ion Ratio	Lower Upper
83 100	
55 80.7	64.7 97.1
98 44.8	37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 14.140 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35


Tgt Ion: 78	Resp: 83898
Ion Ratio	Lower Upper
78 100	
77 22.3	19.0 28.4





#41
Methacrylonitrile
Concen: 14.038 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 41 Resp: 18732
Ion Ratio Lower Upper
41 100
39 62.6 47.2 70.8
67 68.3 50.7 76.1
52 32.5 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 15.165 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 62 Resp: 37203
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 14.512 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020





Tgt Ion: 43	Resp: 5327
Ion Ratio	Lower Upper
43	100
61	17.3 14.3 21.5
87	11.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 14.429 ug/l




RT: 7.116 min Scan# 989


Delta R.T. -0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Tgt Ion:130	Resp: 20239
Ion Ratio	Lower Upper
130	100
95	102.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 14.204 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 63 Resp: 20940

Ion Ratio Lower Upper

63 100





65 29.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 14.439 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 93 Resp: 16781


Ion Ratio Lower Upper

93 100

95 81.5 65.6 98.4

174 83.1 68.2 102.2





#47

Bromodichloromethane

Concen: 14.545 ug/l

RT: 7.817 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 83 Resp: 32762

Ion Ratio Lower Upper

83 100

85 68.4 50.5 75.7





127 8.3 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 14.521 ug/l

RT: 7.689 min Scan# 1083

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 41 Resp: 27374


Ion Ratio Lower Upper

41 100

69 73.5 58.5 87.7

39 64.9 51.7 77.5





#49

1,4-Dioxane

Concen: 282.442 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020





Tgt Ion:	88	Resp:	1058
Ion	Ratio	Lower	Upper
88	100		
43	45.5	33.4	50.2
58	73.6	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 13.200 ug/l




RT: 8.646 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35


Tgt Ion:	98	Resp:	73333
Ion	Ratio	Lower	Upper
98	100		
100	64.8	53.5	80.3





#51
4-Methyl-2-Pentanone
Concen: 75.775 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion: 43 Resp: 18245
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 14.788 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 92 Resp: 52030
Ion Ratio Lower Upper
92 100
91 172.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 14.566 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 75 Resp: 27549

Ion Ratio Lower Upper

75 100





77 31.4 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 14.154 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 75 Resp: 31226


Ion Ratio Lower Upper

75 100

77 30.8 25.4 38.0

39 64.0 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 14.442 ug/l

RT: 9.152 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 97 Resp: 20530

Ion Ratio Lower Upper

97 100

83 87.0 69.0 103.6

85 58.7 44.3 66.5





99 66.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 14.425 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion: 69 Resp: 31787


Ion Ratio Lower Upper

69 100

41 78.2 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 14.417 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 76 Resp: 3634

Ion Ratio Lower Upper

76 100





78 31.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 80.312 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion: 63 Resp: 79531

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 74.997 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 43 Resp: 137054

Ion Ratio Lower Upper

43 100





58 48.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 14.387 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:129 Resp: 22267

Ion Ratio Lower Upper

129 100

127 78.0 39.3 117.8









#61
1,2-Dibromoethane
Concen: 14.458 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:107 Resp: 2111
Ion Ratio Lower Upper
107 100
109 95.3 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 13.236 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 95 Resp: 26760
Ion Ratio Lower Upper
95 100
174 67.4 0.0 135.8
176 67.0 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1470
Delta R.T. 0.006 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion:117 Resp: 127829
Ion Ratio Lower Upper
117 100
82 57.7 50.6 76.0
119 32.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 15.263 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:164 Resp: 19924
Ion Ratio Lower Upper
164 100
166 127.2 105.0 157.6
129 107.6 83.5 125.3
131 97.7 76.5 114.7









#65
Chlorobenzene
Concen: 14.202 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:112 Resp: 5586
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 14.231 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:131 Resp: 18671
Ion Ratio Lower Upper
131 100
133 94.5 47.3 141.9
119 66.7 31.6 95.0





#67

Ethyl Benzene

Concen: 14.631 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion: 91 Resp: 9811

Ion Ratio Lower Upper

91 100





106 29.3 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 29.806 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:106 Resp: 72202

Ion Ratio Lower Upper

106 100

91 216.1 171.2 256.8









#69
o-Xylene
Concen: 14.408 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:106 Resp: 35178
Ion Ratio Lower Upper
106 100
91 231.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 14.853 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion:104 Resp: 58034
Ion Ratio Lower Upper
104 100
78 58.7 45.7 68.5
103 55.6 43.7 65.5





#71
Bromoform
Concen: 14.030 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:173 Resp: 13825
Ion Ratio Lower Upper
173 100
175 46.4 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:152 Resp: 60503
Ion Ratio Lower Upper
152 100
115 76.6 46.9 140.7
150 163.6 0.0 351.0









#73
Isopropylbenzene
Concen: 14.096 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion:105 Resp: 9301
Ion Ratio Lower Upper
105 100
120 26.1 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 13.581 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion: 43 Resp: 44684
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 24.1 18.7 28.1
61 22.3 17.1 25.7









#75
1,1,2-Tetrachloroethane
Concen: 13.560 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 83 Resp: 31830
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 63.3 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 11.039 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 27845
Ion Ratio Lower Upper
75 100
77 43.0 20.5 61.5





#77
Bromobenzene
Concen: 14.212 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion	Ratio	Lower	Upper
156	100		
77	179.9	92.8	278.3
158	93.1	48.8	146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 14.376 ug/l
RT: 11.298 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion	Ratio	Lower	Upper
91	100		
120	22.0	10.8	32.4









#79
2-Chlorotoluene
Concen: 13.748 ug/l
RT: 11.359 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Tgt Ion: 91 Resp: 68534
Ion Ratio Lower Upper
91 100
126 31.4 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 14.468 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion:105 Resp: 79261
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 13.371 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020



Tgt Ion: 75 Resp: 8104
Ion Ratio Lower Upper
75 100
53 138.3 100.5 150.7
89 44.6 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 14.116 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 91 Resp: 77338
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 14.011 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:119 Resp: 75382
Ion Ratio Lower Upper
119 100
91 66.8 32.9 98.7
134 23.7 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 14.480 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:105 Resp: 79232
Ion Ratio Lower Upper
105 100
120 42.0 21.2 63.6









#85
sec-Butylbenzene
Concen: 14.263 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC020

Tgt Ion:105 Resp: 9527
Ion Ratio Lower Upper
105 100
134 19.3 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 14.387 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:119 Resp: 77599
Ion Ratio Lower Upper
119 100
134 25.9 12.5 37.5
91 28.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 13.898 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC020

Tgt Ion:146 Resp: 39520

Ion Ratio Lower Upper

146 100

111 42.7 22.1 66.3





148 62.9 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 14.146 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35




Tgt Ion:146 Resp: 39419


Ion Ratio Lower Upper

146 100

111 42.8 21.3 63.9

148 64.2 31.9 95.5





#89

n-Butylbenzene

Concen: 14.184 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046041.D

Acq: 05 May 2025 11:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC020

Tgt Ion: 91 Resp: 66500

Ion Ratio Lower Upper

91 100

92 52.9 26.9 80.7





134 23.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 13.287 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046041.D




Acq: 05 May 2025 11:35


Tgt Ion:117 Resp: 13332

Ion Ratio Lower Upper

117 100


201 62.9 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 14.092 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:146 Resp: 3903
Ion Ratio Lower Upper
146 100
111 46.6 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 13.535 ug/l
RT: 12.938 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion: 75 Resp: 7232
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 88.4 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 13.745 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:180 Resp: 20840
Ion Ratio Lower Upper
180 100
182 96.6 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 13.426 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35


Tgt Ion:225 Resp: 9524
Ion Ratio Lower Upper
225 100
223 65.6 30.8 92.4
227 61.4 30.9 92.7





#95
Naphthalene
Concen: 14.040 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020



Tgt Ion:128 Resp: 77542
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.1 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 14.023 ug/l
RT: 13.956 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046041.D
Acq: 05 May 2025 11:35

Tgt Ion:180 Resp: 22301
Ion Ratio Lower Upper
180 100
182 92.1 47.8 143.3
145 36.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80678m	28.468	ug/l	
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:168 Resp: 97070
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1









#3
Chloromethane
Concen: 35.658 ug/l
RT: 1.307 min Scan# 36
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 50 Resp: 75238
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 37.273 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 62 Resp: 68901
Ion Ratio Lower Upper
62 100
64 31.5 25.2 37.8





#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5





#7
Trichlorofluoromethane
Concen: 36.824 ug/l
RT: 1.874 min Scan# 129
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:101 Resp: 10369
Ion Ratio Lower Upper
101 100
103 63.7 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 35.350 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 74 Resp: 32724
Ion Ratio Lower Upper
74 100
45 109.9 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 36.997 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:101 Resp: 6225

Ion Ratio Lower Upper

101 100

85 48.3 38.6 58.0





151 69.0 55.2 82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 39.622 ug/l

RT: 2.447 min Scan# 223

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:142 Resp: 78277


Ion Ratio Lower Upper

142 100

127 52.1 41.7 62.5

141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 181.745 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 59 Resp: 6255

Ion Ratio Lower Upper

59 100





57 10.7 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 35.610 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 96 Resp: 58334


Ion Ratio Lower Upper

96 100

61 182.7 146.2 219.2

98 63.8 51.0 76.6





#13

Acrolein

Concen: 184.907 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 7392

Ion Ratio Lower Upper

56 100





55 70.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 36.971 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 41 Resp: 114445


Ion Ratio Lower Upper

41 100

39 75.7 60.6 90.8

76 31.1 24.9 37.3





#15

Acrylonitrile

Concen: 182.874 ug/l

RT: 3.062 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	53	Resp:	188304
Ion	Ratio	Lower	Upper
53	100		
52	81.6	65.3	97.9
51	37.3	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 177.910 ug/l




RT: 2.386 min Scan# 213


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	43	Resp:	175777
Ion	Ratio	Lower	Upper
43	100		
58	26.5	21.2	31.8





#17

Carbon Disulfide

Concen: 37.223 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 14128

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 34.915 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 43 Resp: 82347

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 36.695 ug/l

RT: 3.111 min Scan# 311

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 20966

Ion Ratio Lower Upper

73 100





57 22.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 33.182 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 84 Resp: 66412




Ion Ratio Lower Upper


84 100

49 142.4 113.9 170.9

51 41.9 33.5 50.3


86 67.3 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8





#23

Vinyl Acetate

Concen: 194.528 ug/l

RT: 3.721 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 994230

Ion Ratio Lower Upper

43 100

86 9.4 7.5 11.3

Instrument :

MSVOA_X

ClientSampleId :





VSTDICCC050

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#24

1,1-Dichloroethane

Concen: 36.067 ug/l

RT: 3.605 min Scan# 413

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 63 Resp: 122601


Ion Ratio Lower Upper

63 100

98 7.2 3.6 10.8

100 4.2 2.1 6.3





#25

2-Butanone

Concen: 188.783 ug/l

RT: 4.556 min Scan# 569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 269224

Ion Ratio Lower Upper

43 100





72 23.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 36.059 ug/l

RT: 4.465 min Scan# 554

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 77 Resp: 92857

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 35.466 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 96 Resp: 71518

Ion Ratio Lower Upper

96 100

61 161.4 0.0 322.8





98 64.5 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 31.031 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 49 Resp: 56091


Ion Ratio Lower Upper

49 100

129 2.0 0.0 4.0

130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 170124

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 39.3 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 35.626 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 83 Resp: 125850

Ion Ratio Lower Upper

83 100


85 61.6 49.3 73.9





#31
Cyclohexane
Concen: 38.205 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 10945
Ion Ratio Lower Upper
56 100
69 30.5 24.4 36.6
84 83.6 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 36.034 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 97 Resp: 109781
Ion Ratio Lower Upper
97 100
99 64.7 51.8 77.6
61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 8837

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.8 0.0 33.6





#35

Dibromofluoromethane

Concen: 31.633 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	113	Resp:	59550
Ion	Ratio	Lower	Upper
113	100		
111	103.9	83.1	124.7
192	16.6	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 38.073 ug/l




RT: 5.684 min Scan# 754


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	75	Resp:	83215
Ion	Ratio	Lower	Upper
75	100		
110	32.6	16.3	48.9
77	30.4	24.3	36.5





#37

Ethyl Acetate

Concen: 37.394 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 10253

Ion Ratio Lower Upper

43 100

61 12.9 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 37.320 ug/l

RT: 5.672 min Scan# 752

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:117 Resp: 93712


Ion Ratio Lower Upper

117 100

119 94.0 75.2 112.8

121 30.3 24.2 36.4





#39

Methylcyclohexane

Concen: 39.307 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 10765

Ion Ratio Lower Upper

83 100

55 80.9 64.7 97.1





98 46.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 36.530 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 78 Resp: 247476

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 612
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 58082
Ion Ratio Lower Upper
41 100
39 59.0 47.2 70.8
67 63.4 50.7 76.1
52 30.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 105332
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 38.598 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 16178

Ion Ratio Lower Upper

43

100

61

17.9

87

11.9

14.3

21.5





14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 37.230 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:130 Resp: 59623

Ion Ratio Lower Upper

130




100


95

102.1

0.0

204.2





#45

1,2-Dichloropropane

Concen: 37.064 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 6238

Ion Ratio Lower Upper

63 100





65 30.1 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 36.326 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 93 Resp: 48201


Ion Ratio Lower Upper

93 100

95 82.0 65.6 98.4

174 85.2 68.2 102.2





#47

Bromodichloromethane

Concen: 37.685 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	83	Resp:	96910
Ion	Ratio	Lower	Upper
83	100		
85	63.1	50.5	75.7
127	8.1	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 39.157 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	41	Resp:	84277
Ion	Ratio	Lower	Upper
41	100		
69	73.1	58.5	87.7
39	64.6	51.7	77.5









#49
1,4-Dioxane
Concen: 713.477 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 88 Resp: 30529
Ion Ratio Lower Upper
88 100
43 41.8 33.4 50.2
58 73.3 58.6 88.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#50
Toluene-d8
Concen: 32.395 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 98 Resp: 205479
Ion Ratio Lower Upper
98 100
100 66.9 53.5 80.3





#51

4-Methyl-2-Pentanone

Concen: 193.653 ug/l

RT: 8.574 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 53238

Ion Ratio Lower Upper

43 100





58 36.1 28.9 43.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#52

Toluene

Concen: 37.530 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 92 Resp: 150757

Ion Ratio Lower Upper

92 100

91 170.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 41.056 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 8865

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 38.521 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 97031


Ion Ratio Lower Upper

75 100

77 31.7 25.4 38.0

39 65.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 36.648 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 97 Resp: 5949

Ion Ratio Lower Upper

97 100

83 86.3 69.0 103.6

85 55.4 44.3 66.5





99 59.7 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 39.726 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 69 Resp: 99947


Ion Ratio Lower Upper

69 100

41 76.0 60.8 91.2

39 48.8 39.0 58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 104600

Ion Ratio Lower Upper

76 100





78 32.9 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100

106 26.9 21.5 32.3





#59

2-Hexanone

Concen: 190.168 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 396784

Ion Ratio Lower Upper

43 100





58 49.7 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 38.027 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:129 Resp: 67198

Ion Ratio Lower Upper

129 100

127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 6263
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 14625
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7









#65
Chlorobenzene
Concen: 35.686 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 160600
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 38.017 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:131 Resp: 57066
Ion Ratio Lower Upper
131 100
133 94.6 47.3 141.9
119 63.3 31.6 95.0









#67
Ethyl Benzene
Concen: 38.543 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 295712
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 78.142 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:106 Resp: 216578
Ion Ratio Lower Upper
106 100
91 214.0 171.2 256.8





#69

o-Xylene

Concen: 38.064 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:106 Resp: 106333

Ion Ratio Lower Upper

106 100





91 225.4 112.7 338.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#70

Styrene

Concen: 39.888 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:104 Resp: 178313


Ion Ratio Lower Upper

104 100

78 57.1 45.7 68.5


103 54.6 43.7 65.5





#71
Bromoform
Concen: 39.457 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:173 Resp: 44480
Ion Ratio Lower Upper
173 100
175 46.7 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:152 Resp: 67976
Ion Ratio Lower Upper
152 100
115 93.8 46.9 140.7
150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 34.495 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 90975
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 63.8 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 28.468 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 75 Resp: 80678
Ion Ratio Lower Upper
75 100
77 41.0 20.5 61.5





#77
Bromobenzene
Concen: 36.763 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:156 Resp: 6305
Ion Ratio Lower Upper
156 100
77 185.5 92.8 278.3
158 97.5 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 39.707 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 329981
Ion Ratio Lower Upper
91 100
120 21.6 10.8 32.4





#79

2-Chlorotoluene

Concen: 36.336 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 203509

Ion Ratio Lower Upper

91 100





126 31.1 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 38.516 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 237066

Ion Ratio Lower Upper

105 100

120 46.1 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 38.445 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 26179
Ion Ratio Lower Upper
75 100
53 125.6 100.5 150.7
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 37.880 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 91 Resp: 233172
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 38.622 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:119 Resp: 233468
Ion Ratio Lower Upper
119 100
91 65.8 32.9 98.7
134 22.7 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 38.946 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:105 Resp: 239433
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6









#85
sec-Butylbenzene
Concen: 38.786 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 291074
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 39.879 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:119 Resp: 241658
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 36.199 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 11565

Ion Ratio Lower Upper

146 100

111 44.2 22.1 66.3





148 64.3 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 36.763 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:146 Resp: 115097


Ion Ratio Lower Upper

146 100

111 42.6 21.3 63.9

148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 21389

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7





134 23.5 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion:117 Resp: 42314

Ion Ratio Lower Upper

117 100


201 63.1 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:146 Resp: 115280
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 39.117 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:180 Resp: 6665
Ion Ratio Lower Upper
180 100
182 95.9 47.9 143.8
145 34.1 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 36.395 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:225 Resp: 29007
Ion Ratio Lower Upper
225 100
223 61.6 30.8 92.4
227 61.8 30.9 92.7





#95

Naphthalene

Concen: 39.575 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:128 Resp: 245568

Ion Ratio Lower Upper

128 100

127 13.0 10.4 15.6





129 10.8 8.6 13.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#96

1,2,3-Trichlorobenzene

Concen: 38.765 ug/l

RT: 13.957 min Scan# 2111

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:180 Resp: 69262

Ion Ratio Lower Upper

180 100

182 95.5 47.8 143.3

145 36.2 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 125.060%#			
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 130.020%#			
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 134.060%#			
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 145.640%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	150182m	54.277	ug/l	
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDIC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025


Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:168 Resp: 87028
Ion Ratio Lower Upper
168 100
99 72.1 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 77.828 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 85 Resp: 149508
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.1





#3
Chloromethane
Concen: 72.392 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 50 Resp: 13693
Ion Ratio Lower Upper
50 100
52 32.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 76.398 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 62 Resp: 126609
Ion Ratio Lower Upper
62 100
64 30.6 25.2 37.8





#5
Bromomethane
Concen: 70.507 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 94 Resp: 59240
Ion Ratio Lower Upper
94 100
96 91.7 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 67.148 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 64 Resp: 57332
Ion Ratio Lower Upper
64 100
66 30.8 24.3 36.5





#7
Trichlorofluoromethane
Concen: 67.754 ug/l
RT: 1.867 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100



Tgt Ion:101 Resp: 17104
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 70.807 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 74 Resp: 58762
Ion Ratio Lower Upper
74 100
45 110.1 54.9 164.8





#9
1,1,2-Trichlorotrifluoroethane
Concen: 72.604 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:101 Resp: 109520
Ion Ratio Lower Upper
101 100
85 48.1 38.6 58.0
151 69.9 55.2 82.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#10
Methyl Iodide
Concen: 77.943 ug/l
RT: 2.440 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion:142 Resp: 138045
Ion Ratio Lower Upper
142 100
127 52.0 41.7 62.5
141 14.5 11.5 17.3





#11

Tert butyl alcohol

Concen: 405.394 ug/l

RT: 2.995 min Scan# 310

Delta R.T. 0.018 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 59 Resp: 125094

Ion Ratio Lower Upper

59 100





57 10.9 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 71.884 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 96 Resp: 105566


Ion Ratio Lower Upper

96 100

61 189.9 146.2 219.2

98 63.3 51.0 76.6





#13

Acrolein

Concen: 375.132 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 56 Resp: 134450

Ion Ratio Lower Upper

56 100





55 71.3 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 74.419 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 206519


Ion Ratio Lower Upper

41 100

39 73.8 60.6 90.8


76 31.4 24.9 37.3





#15
Acrylonitrile
Concen: 358.723 ug/l
RT: 3.068 min Scan# 3114
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 53 Resp: 33114
Ion Ratio Lower Upper
53 100
52 84.7 65.3 97.9
51 37.9 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16
Acetone
Concen: 354.936 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 314383
Ion Ratio Lower Upper
43 100
58 27.6 21.2 31.8





#17

Carbon Disulfide

Concen: 77.867 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 76 Resp: 264950

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 69.531 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 43 Resp: 147014

Ion Ratio Lower Upper

43 100

74 22.6 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 73.818 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 73 Resp: 37811
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 67.030 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 84 Resp: 120270
Ion Ratio Lower Upper
84 100
49 136.4 113.9 170.9
51 41.1 33.5 50.3
86 64.0 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 71.109 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 96 Resp: 106600
Ion Ratio Lower Upper
96 100
61 149.7 119.5 179.3
98 64.4 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 77.048 ug/l
RT: 3.763 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 45 Resp: 399494
Ion Ratio Lower Upper
45 100
43 71.2 66.6 100.0
87 25.7 19.8 29.6
59 11.0 8.6 12.8





#23
Vinyl Acetate
Concen: 395.363 ug/l
RT: 3.721 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 43 Resp: 1811550
Ion Ratio Lower Upper
43 100
86 9.5 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 72.114 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 63 Resp: 219759
Ion Ratio Lower Upper
63 100
98 6.8 3.6 10.8
100 4.5 2.1 6.3





#25

2-Butanone

Concen: 379.704 ug/l

RT: 4.556 min Scan# 569

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 485448

Ion Ratio Lower Upper

43 100





72 23.1 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 75.621 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 77 Resp: 174579

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 71.056 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 96 Resp: 128454

Ion Ratio Lower Upper

96 100

61 160.1 0.0 322.8





98 64.9 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 63.932 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 49 Resp: 103602


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.1 56.2 84.2





#29

Tetrahydrofuran

Concen: 368.234 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 42 Resp: 30575

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 70.180 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 83 Resp: 222253

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 76.428 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 56 Resp: 196302
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.3 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 73.635 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 97 Resp: 201112
Ion Ratio Lower Upper
97 100
99 63.8 51.8 77.6
61 48.7 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 62.526 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 65 Resp: 16187

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:114 Resp: 152958


Ion Ratio Lower Upper

114 100

63 24.3 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 65.006 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:113 Resp: 11145

Ion Ratio Lower Upper

113 100

111 101.7 83.1 124.7





192 16.6 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 74.285 ug/l

RT: 5.690 min Scan# 755

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 147873


Ion Ratio Lower Upper

75 100

110 33.5 16.3 48.9

77 31.1 24.3 36.5





#37

Ethyl Acetate

Concen: 74.565 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 186215

Ion Ratio Lower Upper

43 100

61 12.4 10.3 15.5





70 9.9 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 73.875 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:117 Resp: 168946


Ion Ratio Lower Upper

117 100

119 95.0 75.2 112.8

121 29.6 24.2 36.4





#39

Methylcyclohexane

Concen: 76.952 ug/l

RT: 7.372 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 19194

Ion Ratio Lower Upper

83 100

55 84.2 64.7 97.1





98 49.0 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 71.457 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 78 Resp: 440885

Ion Ratio Lower Upper

78 100


77 24.6 19.0 28.4





#41
Methacrylonitrile
Concen: 75.574 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
Client Sample Id :
VSTDICC100



Tgt Ion: 41 Resp: 10486
Ion Ratio Lower Upper
41 100
39 57.0 47.2 70.8
67 64.6 50.7 76.1
52 29.7 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 73.318 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 62 Resp: 187036
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 78.714 ug/l

RT: 6.342 min Scan# 80

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 30049

Ion Ratio Lower Upper

43 100

61 18.0 14.3 21.5





87 11.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 72.420 ug/l

RT: 7.123 min Scan# 990

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:130 Resp: 105628

Ion Ratio Lower Upper

130 100

95 103.8 0.0 204.2





#45

1,2-Dichloropropane

Concen: 73.420 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 63 Resp: 112554

Ion Ratio Lower Upper

63 100





65 30.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 70.938 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 93 Resp: 85728


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4

174 85.0 68.2 102.2





#47

Bromodichloromethane

Concen: 74.845 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 17530

Ion Ratio Lower Upper

83 100

85 63.6 50.5 75.7





127 7.8 6.5 9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 78.101 ug/l

RT: 7.689 min Scan# 1083

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 41 Resp: 153093


Ion Ratio Lower Upper

41 100

69 74.1 58.5 87.7

39 65.5 51.7 77.5





#49

1,4-Dioxane

Concen: 1470.687 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 88 Resp: 5731

Ion Ratio Lower Upper

88 100

43 40.1 33.4 50.2





58 73.3 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 67.031 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 98 Resp: 387230

Ion Ratio Lower Upper

98 100

100 65.2 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 384.740 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 43 Resp: 963319
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 74.010 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 92 Resp: 270763
Ion Ratio Lower Upper
92 100
91 171.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 86.372 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 75 Resp: 16986

Ion Ratio Lower Upper

75 100





77 31.5 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 80.264 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 75 Resp: 184134


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 59.6 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 72.543 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 107264

Ion Ratio Lower Upper

97 100

83 86.8 69.0 103.6

85 56.4 44.3 66.5





99 61.8 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 82.359 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 69 Resp: 188717


Ion Ratio Lower Upper

69 100

41 75.7 60.8 91.2

39 47.7 39.0 58.6





#57

1,3-Dichloropropane

Concen: 71.590 ug/l

RT: 9.305 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 76 Resp: 187649

Ion Ratio Lower Upper

76 100





78 32.1 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 449.791 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 63 Resp: 463169

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 383.945 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 43 Resp: 729602

Ion Ratio Lower Upper

43 100





58 49.9 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 78.918 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:129 Resp: 127012

Ion Ratio Lower Upper

129 100

127 77.0 39.3 117.8





#61

1,2-Dibromoethane

Concen: 74.200 ug/l

RT: 9.610 min Scan# 11266

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:107 Resp: 11266

Ion Ratio Lower Upper

107 100





109 93.9 73.3 109.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#62

4-Bromofluorobenzene

Concen: 72.818 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion: 95 Resp: 153085


Ion Ratio Lower Upper

95 100

174 67.1 0.0 135.8


176 65.3 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:117 Resp: 135214
Ion Ratio Lower Upper
117 100
82 61.7 50.6 76.0
119 30.6 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 67.493 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:164 Resp: 93195
Ion Ratio Lower Upper
164 100
166 123.4 105.0 157.6
129 101.4 83.5 125.3
131 98.0 76.5 114.7





#65
Chlorobenzene
Concen: 70.490 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:112 Resp: 293284
Ion Ratio Lower Upper
112 100
114 31.9 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 74.349 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:131 Resp: 103177
Ion Ratio Lower Upper
131 100
133 95.3 47.3 141.9
119 64.5 31.6 95.0





#67
Ethyl Benzene
Concen: 75.444 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 91 Resp: 53512
Ion Ratio Lower Upper
91 100
106 29.6 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 152.181 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:106 Resp: 389935
Ion Ratio Lower Upper
106 100
91 215.8 171.2 256.8









#69
o-Xylene
Concen: 73.890 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion:106 Resp: 190833
Ion Ratio Lower Upper
106 100
91 230.3 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 79.411 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:104 Resp: 328195
Ion Ratio Lower Upper
104 100
78 57.4 45.7 68.5
103 54.8 43.7 65.5





#71
Bromoform
Concen: 80.927 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:173 Resp: 8435
Ion Ratio Lower Upper
173 100
175 48.3 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:152 Resp: 66369
Ion Ratio Lower Upper
152 100
115 120.4 46.9 140.7
150 190.4 0.0 351.0









#73
Isopropylbenzene
Concen: 71.083 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Tgt Ion:105 Resp: 514528
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 76.073 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 43 Resp: 274553
Ion Ratio Lower Upper
43 100
70 38.2 30.9 46.3
55 22.5 18.7 28.1
61 22.2 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 66.207 ug/l

RT: 11.213 min Scan# 1661

Delta R.T. 0.006 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion: 83 Resp: 17048

Ion Ratio Lower Upper

83 100

131 9.7 5.0 14.9





85 64.9 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#76

1,2,3-Trichloropropane

Concen: 54.277 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion: 75 Resp: 150182

Ion Ratio Lower Upper

75 100


77 42.2 20.5 61.5





#77
Bromobenzene
Concen: 69.959 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:156 Resp: 11715
Ion Ratio Lower Upper
156 100
77 187.4 92.8 278.3
158 96.1 48.8 146.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#78
n-propylbenzene
Concen: 74.974 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion: 91 Resp: 608332
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 68.100 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Tgt Ion: 91 Resp: 372392
Ion Ratio Lower Upper
91 100
126 30.9 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 71.903 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 432096
Ion Ratio Lower Upper
105 100
120 46.3 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 81.820 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion: 75 Resp: 54398
Ion Ratio Lower Upper
75 100
53 115.8 100.5 150.7
89 47.0 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 71.890 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion: 91 Resp: 432053
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83
tert-Butylbenzene
Concen: 73.197 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:119 Resp: 43201
Ion Ratio Lower Upper
119 100
91 64.4 32.9 98.7
134 22.8 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 73.758 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21


Tgt Ion:105 Resp: 442728
Ion Ratio Lower Upper
105 100
120 42.4 21.2 63.6





#85
sec-Butylbenzene
Concen: 74.188 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100




Tgt Ion:105 Resp: 54359
Ion Ratio Lower Upper
105 100
134 19.4 9.7 29.1

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#86
p-Isopropyltoluene
Concen: 77.393 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21



Tgt Ion:119 Resp: 457898
Ion Ratio Lower Upper
119 100
134 24.6 12.5 37.5
91 26.8 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 70.449 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion:146 Resp: 21975

Ion Ratio Lower Upper

146 100

111 44.1 22.1 66.3





148 63.5 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 71.190 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:146 Resp: 217613


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.9

148 63.2 31.9 95.5





#89

n-Butylbenzene

Concen: 80.996 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC100

Tgt Ion: 91 Resp: 416602

Ion Ratio Lower Upper

91 100

92 54.0 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 74.978 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046043.D




Acq: 05 May 2025 12:21


Tgt Ion:117 Resp: 82528

Ion Ratio Lower Upper

117 100


201 63.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 71.367 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:146 Resp: 216834
Ion Ratio Lower Upper
146 100
111 46.2 22.4 67.3
148 64.0 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 74.512 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion: 75 Resp: 43673
Ion Ratio Lower Upper
75 100
155 69.4 34.9 104.8
157 90.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 82.605 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC100

Tgt Ion:180 Resp: 137430

Ion Ratio Lower Upper

180 100

182 93.7 47.9 143.8





145 34.6 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 71.314 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. -0.000 min

Lab File: VX046043.D

Acq: 05 May 2025 12:21




Tgt Ion:225 Resp: 55494


Ion Ratio Lower Upper

225 100

223 62.3 30.8 92.4


227 64.9 30.9 92.7





#95
Naphthalene
Concen: 80.842 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100



Tgt Ion:128 Resp: 489778
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 10.9 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 79.981 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046043.D
Acq: 05 May 2025 12:21

Tgt Ion:180 Resp: 139525
Ion Ratio Lower Upper
180 100
182 93.7 47.8 143.3
145 35.3 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 187.960%#	
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 197.120%#	
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 202.500%#	
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 218.340%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	213744m	84.960	ug/l	
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:168 Resp: 8296
Ion Ratio Lower Upper
168 100
99 67.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 118.929 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 85 Resp: 217793
Ion Ratio Lower Upper
85 100
87 32.5 15.7 47.1





#3
Chloromethane
Concen: 109.161 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 50 Resp: 19684
Ion Ratio Lower Upper
50 100
52 32.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 118.953 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 62 Resp: 187924
Ion Ratio Lower Upper
62 100
64 31.7 25.2 37.8





#5
Bromomethane
Concen: 103.625 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 94 Resp: 83008
Ion Ratio Lower Upper
94 100
96 95.8 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 96.822 ug/l
RT: 1.660 min Scan# 94
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 64 Resp: 78806
Ion Ratio Lower Upper
64 100
66 31.3 24.3 36.5





#7
Trichlorofluoromethane
Concen: 101.851 ug/l
RT: 1.868 min Scan# 129
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:101 Resp: 245108
Ion Ratio Lower Upper
101 100
103 66.4 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 111.791 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 74 Resp: 88441
Ion Ratio Lower Upper
74 100
45 104.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 112.243 ug/l
RT: 2.313 min Scan# 202
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:101 Resp: 161404
Ion Ratio Lower Upper
101 100
85 46.8 38.6 58.0
151 69.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 112.441 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:142 Resp: 189843
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.4 11.5 17.3





#11

Tert butyl alcohol

Concen: 616.119 ug/l

RT: 3.002 min Scan# 310

Delta R.T. 0.024 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 59 Resp: 181238

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 111.039 ug/l

RT: 2.307 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 96 Resp: 155451


Ion Ratio Lower Upper

96 100

61 184.5 146.2 219.2

98 64.3 51.0 76.6





#13

Acrolein

Concen: 594.872 ug/l

RT: 2.239 min Scan# 188

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 56 Resp: 20325

Ion Ratio Lower Upper

56 100





55 72.1 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 112.561 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 41 Resp: 297776


Ion Ratio Lower Upper

41 100

39 73.7 60.6 90.8

76 31.8 24.9 37.3





#15

Acrylonitrile

Concen: 550.973 ug/l

RT: 3.069 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC150

Tgt Ion:	53	Resp:	484854
Ion	Ratio	Lower	Upper
53	100		
52	84.2	65.3	97.9
51	37.5	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 545.754 ug/l




RT: 2.392 min Scan# 214


Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Tgt Ion:	43	Resp:	460820
Ion	Ratio	Lower	Upper
43	100		
58	27.3	21.2	31.8





#17

Carbon Disulfide

Concen: 122.499 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 39735

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 108.099 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 43 Resp: 217885

Ion Ratio Lower Upper

43 100

74 21.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 114.112 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 73 Resp: 557210
Ion Ratio Lower Upper
73 100
57 22.0 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 100.579 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 84 Resp: 172035
Ion Ratio Lower Upper
84 100
49 136.0 113.9 170.9
51 41.5 33.5 50.3
86 63.7 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 108.241 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 96 Resp: 154694
Ion Ratio Lower Upper
96 100
61 154.6 119.5 179.3
98 64.3 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 116.858 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 45 Resp: 577610
Ion Ratio Lower Upper
45 100
43 70.3 66.6 100.0
87 24.7 19.8 29.6
59 10.6 8.6 12.8





#23
Vinyl Acetate
Concen: 608.005 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150




Tgt Ion: 43 Resp: 265575
Ion Ratio Lower Upper
43 100
86 9.6 7.5 11.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#24
1,1-Dichloroethane
Concen: 110.182 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45



Tgt Ion: 63 Resp: 320083
Ion Ratio Lower Upper
63 100
98 7.0 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 581.292 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 70846

Ion Ratio Lower Upper

43 100





72 24.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 117.548 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 77 Resp: 258698

Ion Ratio Lower Upper

77 100

97 21.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 108.994 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 96 Resp: 18783

Ion Ratio Lower Upper

96 100

61 159.8 0.0 322.8





98 64.8 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 95.108 ug/l

RT: 4.898 min Scan# 625

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 49 Resp: 146923


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 569.081 ug/l

RT: 5.007 min Scan# 642

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 42 Resp: 450450

Ion Ratio Lower Upper

42 100

72 41.6 33.7 50.5





71 38.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 107.208 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 83 Resp: 323659

Ion Ratio Lower Upper

83 100


85 65.1 49.3 73.9





#31
Cyclohexane
Concen: 116.934 ug/l
RT: 5.458 min Scan# 717
Delta R.T. -0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 56 Resp: 28631
Ion Ratio Lower Upper
56 100
69 31.2 24.4 36.6
84 82.8 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 113.572 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 97 Resp: 295702
Ion Ratio Lower Upper
97 100
99 64.3 51.8 77.6
61 48.0 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 93.983 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 65 Resp: 231952

Ion Ratio Lower Upper

65 100





67 49.4 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:114 Resp: 144975


Ion Ratio Lower Upper

114 100

63 23.9 0.0 49.2

88 18.3 0.0 33.6





#35

Dibromofluoromethane

Concen: 98.562 ug/l

RT: 5.385 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:113 Resp: 160168

Ion Ratio Lower Upper

113 100

111 103.0 83.1 124.7





192 16.9 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 116.462 ug/l

RT: 5.684 min Scan# 754

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 219731


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5





#37

Ethyl Acetate

Concen: 115.889 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 274312

Ion Ratio Lower Upper

43 100

61 12.5 10.3 15.5





70 10.2 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 115.760 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:117 Resp: 250916


Ion Ratio Lower Upper

117 100

119 94.3 75.2 112.8

121 29.7 24.2 36.4





#39

Methylcyclohexane

Concen: 121.107 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 83 Resp: 28631

Ion Ratio Lower Upper

83 100

55 83.8 64.7 97.1





98 47.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 109.813 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 78 Resp: 642178

Ion Ratio Lower Upper

78 100


77 23.4 19.0 28.4





#41
Methacrylonitrile
Concen: 115.139 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 41 Resp: 151424
Ion Ratio Lower Upper
41 100
39 60.0 47.2 70.8
67 65.9 50.7 76.1
52 30.9 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 112.492 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 62 Resp: 271992
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 123.780 ug/l

RT: 6.342 min Scan# 862

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Tgt Ion: 43 Resp: 44787

Ion Ratio Lower Upper

43 100

61 18.1 14.3 21.5





87 11.7 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 114.032 ug/l

RT: 7.117 min Scan# 989

Delta R.T. -0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:130 Resp: 157641

Ion Ratio Lower Upper

130 100

95 105.9 0.0 204.2





#45

1,2-Dichloropropane

Concen: 113.217 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 63 Resp: 16450

Ion Ratio Lower Upper

63 100





65 30.8 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 109.869 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 93 Resp: 125846


Ion Ratio Lower Upper

93 100

95 82.4 65.6 98.4


174 85.0 68.2 102.2





#47
Bromodichloromethane
Concen: 116.365 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion: 83 Resp: 258320
Ion Ratio Lower Upper
83 100
85 64.8 50.5 75.7
127 8.0 6.5 9.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#48
Methyl methacrylate
Concen: 124.356 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 41 Resp: 231040
Ion Ratio Lower Upper
41 100
69 73.0 58.5 87.7
39 63.8 51.7 77.5





#49

1,4-Dioxane

Concen: 2241.695 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 88 Resp: 82800

Ion Ratio Lower Upper

88 100

43 40.0 33.4 50.2





58 74.8 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 101.252 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 98 Resp: 554390

Ion Ratio Lower Upper

98 100

100 65.4 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 577.831 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 43 Resp: 137127
Ion Ratio Lower Upper
43 100
58 36.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 113.430 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 92 Resp: 393323
Ion Ratio Lower Upper
92 100
91 171.0 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 138.001 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 75 Resp: 25723

Ion Ratio Lower Upper

75 100





77 31.7 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 124.658 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion: 75 Resp: 271054


Ion Ratio Lower Upper

75 100

77 31.4 25.4 38.0


39 60.0 52.2 78.4





#55
1,1,2-Trichloroethane
Concen: 110.551 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion: 97 Resp: 154932
Ion Ratio Lower Upper
97 100
83 88.5 69.0 103.6
85 56.0 44.3 66.5
99 64.5 47.8 71.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#56
Ethyl methacrylate
Concen: 127.968 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 69 Resp: 277921
Ion Ratio Lower Upper
69 100
41 75.4 60.8 91.2
39 47.2 39.0 58.6





#57

1,3-Dichloropropane

Concen: 109.813 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 76 Resp: 272814

Ion Ratio Lower Upper

76 100





78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 697.344 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion: 63 Resp: 680608

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 571.094 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 43 Resp: 102859

Ion Ratio Lower Upper

43 100





58 50.2 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 122.915 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:129 Resp: 187497

Ion Ratio Lower Upper

129 100


127 77.1 39.3 117.8





#61
1,2-Dibromoethane
Concen: 115.087 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:107 Resp: 165634
Ion Ratio Lower Upper
107 100
109 94.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 109.174 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 95 Resp: 217536
Ion Ratio Lower Upper
95 100
174 67.5 0.0 135.8
176 64.4 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:117 Resp: 12855
Ion Ratio Lower Upper
117 100
82 61.4 50.6 76.0
119 31.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 100.918 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:164 Resp: 132488
Ion Ratio Lower Upper
164 100
166 123.3 105.0 157.6
129 101.6 83.5 125.3
131 97.3 76.5 114.7









#65
Chlorobenzene
Concen: 108.615 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion:112 Resp: 429650
Ion Ratio Lower Upper
112 100
114 31.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 115.333 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:131 Resp: 152173
Ion Ratio Lower Upper
131 100
133 95.7 47.3 141.9
119 64.0 31.6 95.0









#67
Ethyl Benzene
Concen: 116.421 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 91 Resp: 78512
Ion Ratio Lower Upper
91 100
106 29.9 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#68
m/p-Xylenes
Concen: 234.349 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:106 Resp: 570913
Ion Ratio Lower Upper
106 100
91 214.2 171.2 256.8









#69
o-Xylene
Concen: 113.990 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:106 Resp: 279900
Ion Ratio Lower Upper
106 100
91 226.9 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 120.694 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:104 Resp: 474254
Ion Ratio Lower Upper
104 100
78 57.2 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 127.128 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:173 Resp: 125980
Ion Ratio Lower Upper
173 100
175 47.3 23.4 70.0
254 0.1 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion:152 Resp: 60345
Ion Ratio Lower Upper
152 100
115 150.7 46.9 140.7#
150 211.6 0.0 351.0





#73
Isopropylbenzene
Concen: 114.312 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:105 Resp: 75233
Ion Ratio Lower Upper
105 100
120 25.3 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 120.944 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 43 Resp: 396877
Ion Ratio Lower Upper
43 100
70 38.5 30.9 46.3
55 22.5 18.7 28.1
61 22.5 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 104.018 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion: 83 Resp: 24353
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 64.4 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 84.960 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45


Tgt Ion: 75 Resp: 213744
Ion Ratio Lower Upper
75 100
77 42.4 20.5 61.5





#77
Bromobenzene
Concen: 110.297 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:156 Resp: 167934
Ion Ratio Lower Upper
156 100
77 187.1 92.8 278.3
158 97.3 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 119.448 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.006 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 881218
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 107.514 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 534560

Ion Ratio Lower Upper

91 100





126 31.0 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 112.825 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:105 Resp: 616477

Ion Ratio Lower Upper

105 100

120 46.0 23.1 69.2









#81
trans-1,4-Dichloro-2-butene
Concen: 134.609 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150

Tgt Ion: 75 Resp: 81372
Ion Ratio Lower Upper
75 100
53 114.9 100.5 150.7
89 46.5 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 111.948 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 91 Resp: 611736
Ion Ratio Lower Upper
91 100
126 26.7 13.3 39.8









#83
tert-Butylbenzene
Concen: 115.065 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Tgt Ion:119 Resp: 61747
Ion Ratio Lower Upper
119 100
91 65.6 32.9 98.7
134 22.6 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 114.236 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:105 Resp: 623456
Ion Ratio Lower Upper
105 100
120 43.1 21.2 63.6





#85

sec-Butylbenzene

Concen: 118.023 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:105 Resp: 78629

Ion Ratio Lower Upper

105 100





134 19.1 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 121.113 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:119 Resp: 651531


Ion Ratio Lower Upper

119 100

134 25.4 12.5 37.5

91 27.0 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 110.394 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:146 Resp: 313100

Ion Ratio Lower Upper

146 100

111 44.5 22.1 66.3





148 64.0 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 112.167 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:146 Resp: 311752


Ion Ratio Lower Upper

146 100

111 43.0 21.3 63.9

148 63.9 31.9 95.5





#89

n-Butylbenzene

Concen: 129.533 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion: 91 Resp: 60578

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7





134 24.0 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 123.091 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046044.D




Acq: 05 May 2025 12:45


Tgt Ion:117 Resp: 123188

Ion Ratio Lower Upper

117 100


201 63.2 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 111.506 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC150



Tgt Ion:146 Resp: 308039
Ion Ratio Lower Upper
146 100
111 46.3 22.4 67.3
148 63.6 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 120.915 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion: 75 Resp: 64438
Ion Ratio Lower Upper
75 100
155 69.6 34.9 104.8
157 88.4 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 134.366 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. 0.000 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC150

Tgt Ion:180 Resp: 20325

Ion Ratio Lower Upper

180 100

182 94.1 47.9 143.8





145 34.5 17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 113.857 ug/l

RT: 13.719 min Scan# 2072

Delta R.T. -0.006 min

Lab File: VX046044.D

Acq: 05 May 2025 12:45




Tgt Ion:225 Resp: 80558


Ion Ratio Lower Upper

225 100

223 62.5 30.8 92.4


227 63.1 30.9 92.7





#95
Naphthalene
Concen: 130.929 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150



Tgt Ion:128 Resp: 721236
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 11.0 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 126.331 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046044.D
Acq: 05 May 2025 12:45

Tgt Ion:180 Resp: 200379
Ion Ratio Lower Upper
180 100
182 96.5 47.8 143.3
145 36.7 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	6.280%#		
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	6.320%#		
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	6.460%#		
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	6.760%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9806m	3.565	ug/l	
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	7928m	2.799	ug/l	
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On    : 05 May 2025  16:04
Operator  : JC/MD
Sample    : VSTDIC005
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 10   Sample Multiplier: 1
```


Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:168 Resp: 9696
Ion Ratio Lower Upper
168 100
99 74.0 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 2.894 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 85 Resp: 6195
Ion Ratio Lower Upper
85 100
87 28.7 15.7 47.1









#3
Chloromethane
Concen: 3.125 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 50 Resp: 658
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 3.249 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 62 Resp: 5999
Ion Ratio Lower Upper
62 100
64 29.1 25.2 37.8





#5
Bromomethane
Concen: 3.164 ug/l
RT: 1.593 min Scan# 83
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 94 Resp: 2962
Ion Ratio Lower Upper
94 100
96 102.2 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 3.750 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 64 Resp: 3567
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 3.413 ug/l
RT: 1.874 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:101 Resp: 959
Ion Ratio Lower Upper
101 100
103 54.0 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 3.266 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 74 Resp: 3020
Ion Ratio Lower Upper
74 100
45 115.1 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 3.517 ug/l
RT: 2.325 min Scan# 202
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion:101 Resp: 591
Ion Ratio Lower Upper
101 100
85 46.3 38.6 58.0
151 68.8 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 2.990 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:142 Resp: 5900
Ion Ratio Lower Upper
142 100
127 52.8 41.7 62.5
141 14.3 11.5 17.3





#11

Tert butyl alcohol

Concen: 16.111 ug/l

RT: 2.965 min Scan# 30

Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 59 Resp: 553

Ion Ratio Lower Upper

59 100





57 10.0 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 3.359 ug/l

RT: 2.306 min Scan# 200

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 96 Resp: 5496


Ion Ratio Lower Upper

96 100

61 178.7 146.2 219.2

98 59.5 51.0 76.6





#13

Acrolein

Concen: 14.206 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 567

Ion Ratio Lower Upper

56 100





55 74.5 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carbone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 3.319 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 41 Resp: 10262


Ion Ratio Lower Upper

41 100

39 74.1 60.6 90.8

76 29.1 24.9 37.3





#15

Acrylonitrile

Concen: 16.282 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	53	Resp:	16740
Ion	Ratio	Lower	Upper
53	100		
52	84.3	65.3	97.9
51	38.5	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 20.036 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	43	Resp:	19773
Ion	Ratio	Lower	Upper
43	100		
58	27.7	21.2	31.8





#17

Carbon Disulfide

Concen: 2.919 ug/l

RT: 2.501 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 76 Resp: 1106

Ion Ratio Lower Upper

76 100





78 10.2 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 3.357 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 43 Resp: 7909

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 3.241 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04

Tgt Ion: 73 Resp: 1849

Ion Ratio Lower Upper

73 100

57 20.5 17.7 26.5



Instrument :

MSVOA_X

ClientSampleId :


VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 3.341 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 84 Resp: 6680




Ion Ratio Lower Upper


84 100

49 137.0 113.9 170.9

51 43.8 33.5 50.3


86 67.6 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 3.234 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 96 Resp: 540
Ion Ratio Lower Upper
96 100
61 149.9 119.5 179.3
98 62.9 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 3.230 ug/l
RT: 3.751 min Scan# 437
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 45 Resp: 18657
Ion Ratio Lower Upper
45 100
43 128.0 66.6 100.0#
87 24.7 19.8 29.6
59 10.7 8.6 12.8









#23
Vinyl Acetate
Concen: 16.127 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 82328
Ion Ratio Lower Upper
43 100
86 9.1 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 3.297 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 63 Resp: 11193
Ion Ratio Lower Upper
63 100
98 6.9 3.6 10.8
100 4.7 2.1 6.3





#25

2-Butanone

Concen: 18.340 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 2612

Ion Ratio Lower Upper

43 100





72 21.8 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 3.205 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 77 Resp: 8245

Ion Ratio Lower Upper

77 100

97 23.1 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 3.089 ug/l

RT: 4.483 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 96 Resp: 622

Ion Ratio Lower Upper

96 100

61 168.5 0.0 322.8





98 67.6 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 2.969 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 49 Resp: 5360


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.0

130 70.5 56.2 84.2





#29

Tetrahydrofuran

Concen: 16.680 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 42 Resp: 1543

Ion Ratio Lower Upper

42 100

72 42.2 33.7 50.5





71 37.4 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 3.294 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 83 Resp: 11624

Ion Ratio Lower Upper

83 100

85 69.9 49.3 73.9





#31

Cyclohexane

Concen: 3.588 ug/l

RT: 5.458 min Scan# 718

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 56 Resp: 10269

Ion Ratio Lower Upper

56 100

69 31.6 24.4 36.6





84 79.1 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#32

1,1,1-Trichloroethane

Concen: 3.229 ug/l

RT: 5.367 min Scan# 702

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 97 Resp: 9827


Ion Ratio Lower Upper

97 100

99 62.6 51.8 77.6

61 51.8 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 3.143 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 65 Resp: 906

Ion Ratio Lower Upper

65 100





67 50.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion:114 Resp: 168484


Ion Ratio Lower Upper

114 100

63 24.5 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 3.161 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	113	Resp:	596
Ion	Ratio	Lower	Upper
113	100		
111	104.4	83.1	124.7
192	17.5	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 3.551 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	75	Resp:	7787
Ion	Ratio	Lower	Upper
75	100		
110	30.6	16.3	48.9
77	30.5	24.3	36.5





#37

Ethyl Acetate

Concen: 3.565 ug/l m

RT: 4.721 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	43	Resp:	9800
Ion	Ratio	Lower	Upper
43	100		
61	11.0	10.3	15.5
70	8.4	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 3.376 ug/l




RT: 5.659 min Scan# 750


Delta R.T. -0.012 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	117	Resp:	8505
Ion	Ratio	Lower	Upper
117	100		
119	90.6	75.2	112.8
121	29.5	24.2	36.4





#39

Methylcyclohexane

Concen: 3.597 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC005

Tgt Ion:	83	Resp:	988
Ion	Ratio	Lower	Upper
83	100		
55	73.3	64.7	97.1
98	50.3	37.4	56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 3.315 ug/l




RT: 6.031 min Scan# 811


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:	78	Resp:	22528
Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.4





#41
Methacrylonitrile
Concen: 3.174 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 41 Resp: 485
Ion Ratio Lower Upper
41 100
39 59.7 47.2 70.8
67 0.0 50.7 76.1
52 28.8 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 3.563 ug/l
RT: 6.080 min Scan# 819
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 62 Resp: 10011
Ion Ratio Lower Upper
62 100
98 7.8 0.0 15.2





#43

Isopropyl Acetate

Concen: 3.309 ug/l

RT: 6.348 min Scan# 80

Delta R.T. 0.012 min

Lab File: VX046046.D


Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005






Tgt Ion: 43	Resp: 13914
Ion Ratio	Lower Upper
43	100
61	16.2 14.3 21.5
87	10.6 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene


Concen: 3.307 ug/l

RT: 7.123 min Scan# 990



Delta R.T. -0.000 min


Lab File: VX046046.D

Acq: 05 May 2025 16:04



Tgt Ion:130	Resp: 5313
Ion Ratio	Lower Upper
130	100
95	108.1 0.0 204.2





#45

1,2-Dichloropropane

Concen: 3.228 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 63 Resp: 545

Ion Ratio Lower Upper

63 100





65 36.2 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 3.318 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 93 Resp: 4417


Ion Ratio Lower Upper

93 100

95 80.1 65.6 98.4

174 87.6 68.2 102.2





#47

Bromodichloromethane

Concen: 3.254 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	83	Resp:	8394
Ion	Ratio	Lower	Upper
83	100		
85	61.9	50.5	75.7
127	9.0	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 3.321 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	41	Resp:	7171
Ion	Ratio	Lower	Upper
41	100		
69	68.4	58.5	87.7
39	65.1	51.7	77.5





#49

1,4-Dioxane

Concen: 67.721 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	88	Resp:	290
Ion	Ratio	Lower	Upper
88	100		
43	45.3	33.4	50.2
58	71.0	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 3.232 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion:	98	Resp:	20566
Ion	Ratio	Lower	Upper
98	100		
100	66.1	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 16.967 ug/l
RT: 8.567 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 43 Resp: 46794
Ion Ratio Lower Upper
43 100
58 34.8 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 3.505 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 92 Resp: 14126
Ion Ratio Lower Upper
92 100
91 167.2 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 3.155 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 75 Resp: 6834

Ion Ratio Lower Upper

75 100





77 31.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 3.129 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 75 Resp: 7908


Ion Ratio Lower Upper

75 100

77 31.9 25.4 38.0

39 55.3 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 3.490 ug/l

RT: 9.153 min Scan# 1323

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Tgt Ion: 97 Resp: 568

Ion Ratio Lower Upper

97 100

83 88.0 69.0 103.6

85 54.4 44.3 66.5

99 62.2 47.8 71.6

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC005

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 3.390 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04




Tgt Ion: 69 Resp: 8555


Ion Ratio Lower Upper

69 100

41 73.7 60.8 91.2

39 49.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 3.508 ug/l

RT: 9.305 min Scan# 1134

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC005

Tgt Ion: 76 Resp: 1012

Ion Ratio Lower Upper

76 100





78 29.5 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 18.346 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion: 63 Resp: 20809

Ion Ratio Lower Upper

63 100

106 26.4 21.5 32.3





#59

2-Hexanone

Concen: 16.651 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 43 Resp: 3485

Ion Ratio Lower Upper

43 100





58 47.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 3.102 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:129 Resp: 5500

Ion Ratio Lower Upper

129 100

127 80.4 39.3 117.8









#61
1,2-Dibromoethane
Concen: 3.356 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:107 Resp: 5614
Ion Ratio Lower Upper
107 100
109 91.0 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 3.378 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 95 Resp: 7822
Ion Ratio Lower Upper
95 100
174 65.0 0.0 135.8
176 66.6 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:117 Resp: 14716
Ion Ratio Lower Upper
117 100
82 61.1 50.6 76.0
119 32.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 3.162 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:164 Resp: 4752
Ion Ratio Lower Upper
164 100
166 127.8 105.0 157.6
129 105.8 83.5 125.3
131 104.1 76.5 114.7









#65
Chlorobenzene
Concen: 3.399 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:112 Resp: 1539
Ion Ratio Lower Upper
112 100
114 32.6 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 3.326 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:131 Resp: 5024
Ion Ratio Lower Upper
131 100
133 95.0 47.3 141.9
119 64.8 31.6 95.0





#67

Ethyl Benzene

Concen: 3.463 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion: 91 Resp: 2673

Ion Ratio Lower Upper

91 100





106 29.9 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 7.156 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:106 Resp: 19957

Ion Ratio Lower Upper

106 100

91 210.1 171.2 256.8









#69
o-Xylene
Concen: 3.343 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion:106 Resp: 9398
Ion Ratio Lower Upper
106 100
91 233.6 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 3.312 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:104 Resp: 14900
Ion Ratio Lower Upper
104 100
78 57.9 45.7 68.5
103 55.6 43.7 65.5





#71
Bromoform
Concen: 3.059 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:173 Resp: 3470
Ion Ratio Lower Upper
173 100
175 45.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:152 Resp: 67939
Ion Ratio Lower Upper
152 100
115 66.8 46.9 140.7
150 157.5 0.0 351.0









#73
Isopropylbenzene
Concen: 3.266 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 105 Resp: 2420
Ion Ratio Lower Upper
105 100
120 26.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 3.038 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 43 Resp: 11224
Ion Ratio Lower Upper
43 100
70 35.4 30.9 46.3
55 25.0 18.7 28.1
61 19.6 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 3.479 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005

Tgt Ion: 83 Resp: 9170
Ion Ratio Lower Upper
83 100
131 8.3 5.0 14.9
85 61.7 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 2.799 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion: 75 Resp: 7928
Ion Ratio Lower Upper
75 100
77 40.6 20.5 61.5





#77
Bromobenzene
Concen: 3.417 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:156 Resp: 585
Ion Ratio Lower Upper
156 100
77 177.0 92.8 278.3
158 97.7 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#78
n-propylbenzene
Concen: 3.424 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 28436
Ion Ratio Lower Upper
91 100
120 21.7 10.8 32.4









#79
2-Chlorotoluene
Concen: 3.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Tgt Ion: 91 Resp: 18672
Ion Ratio Lower Upper
91 100
126 29.7 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 3.372 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20742
Ion Ratio Lower Upper
105 100
120 45.9 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 2.686 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion: 75 Resp: 1828
Ion Ratio Lower Upper
75 100
53 125.8 100.5 150.7
89 47.9 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 3.245 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 91 Resp: 19966
Ion Ratio Lower Upper
91 100
126 27.3 13.3 39.8





#83

tert-Butylbenzene

Concen: 3.484 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005

Tgt Ion:119 Resp: 21049

Ion Ratio Lower Upper

119 100

91 67.0 32.9 98.7





134 23.8 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 3.355 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046046.D




Acq: 05 May 2025 16:04


Tgt Ion:105 Resp: 20613

Ion Ratio Lower Upper

105 100


120 44.5 21.2 63.6





#85
sec-Butylbenzene
Concen: 3.412 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:105 Resp: 25590
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 3.460 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04


Tgt Ion:119 Resp: 20955
Ion Ratio Lower Upper
119 100
134 25.6 12.5 37.5
91 28.0 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 3.315 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005




Tgt Ion:146 Resp: 1058
Ion Ratio Lower Upper
146 100
111 46.2 22.1 66.3
148 62.4 32.1 96.5

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#88
1,4-Dichlorobenzene
Concen: 3.486 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04



Tgt Ion:146 Resp: 10908
Ion Ratio Lower Upper
146 100
111 52.1 21.3 63.9
148 69.4 31.9 95.5





#89

n-Butylbenzene

Concen: 3.419 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:	91	Resp:	18004
Ion Ratio	Lower	Upper	
91	100		
92	49.5	26.9	80.7
134	21.2	11.8	35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 3.153 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:	117	Resp:	3553
Ion Ratio	Lower	Upper	
117	100		
201	58.5	31.6	94.7





#91
1,2-Dichlorobenzene
Concen: 3.444 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC005



Tgt Ion:146 Resp: 10712
Ion Ratio Lower Upper
146 100
111 44.7 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 2.812 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion: 75 Resp: 1687
Ion Ratio Lower Upper
75 100
155 71.7 34.9 104.8
157 87.0 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 3.359 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC005





Tgt Ion:180	Resp:	5720
Ion Ratio	Lower	Upper
180	100	
182	91.6	47.9 143.8
145	35.7	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 3.535 ug/l




RT: 13.719 min Scan# 2072


Delta R.T. -0.006 min

Lab File: VX046046.D

Acq: 05 May 2025 16:04


Tgt Ion:225	Resp:	2816
Ion Ratio	Lower	Upper
225	100	
223	65.1	30.8 92.4
227	58.8	30.9 92.7





#95
Naphthalene
Concen: 3.209 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005



Tgt Ion:128 Resp: 1990
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 11.5 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carbone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 3.217 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046046.D
Acq: 05 May 2025 16:04

Tgt Ion:180 Resp: 5745
Ion Ratio Lower Upper
180 100
182 99.1 47.8 143.3
145 34.8 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	1061m	0.619	ug/l	
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1956m	0.717	ug/l	
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDIC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1235m	0.577	ug/l	
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1661m	0.674	ug/l	
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	2149m	0.789	ug/l	
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:168 Resp: 92040

Ion Ratio Lower Upper

168 100





99 71.5 54.9 82.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#2

Dichlorodifluoromethane

Concen: 0.596 ug/l

RT: 1.166 min Scan# 13

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 85 Resp: 1211

Ion Ratio Lower Upper

85 100

87 32.4 15.7 47.1









#3
Chloromethane
Concen: 0.639 ug/l
RT: 1.307 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 50 Resp: 1278
Ion Ratio Lower Upper
50 100
52 30.3 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 0.707 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 62 Resp: 1239
Ion Ratio Lower Upper
62 100
64 41.7 25.2 37.8#





#6
Chloroethane
Concen: 0.951 ug/l
RT: 1.678 min Scan# 91
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001



Tgt Ion: 64 Resp: 859
Ion Ratio Lower Upper
64 100
66 35.2 24.3 36.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#7
Trichlorofluoromethane
Concen: 0.734 ug/l
RT: 1.880 min Scan# 130
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1959
Ion Ratio Lower Upper
101 100
103 67.8 51.0 76.4









#8
Diethyl Ether
Concen: 0.845 ug/l
RT: 2.130 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 74 Resp: 74
Ion Ratio Lower Upper
74 100
45 123.6 54.9 164.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#9
1,1,2-Trichlorotrifluoroethane
Concen: 0.731 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:101 Resp: 1166
Ion Ratio Lower Upper
101 100
85 47.3 38.6 58.0
151 73.1 55.2 82.8





#12

1,1-Dichloroethene

Concen: 0.704 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	96	Resp:	109
Ion	Ratio	Lower	Upper
96	100		
61	199.6	146.2	219.2
98	49.6	51.0	76.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 0.660 ug/l




RT: 2.654 min Scan# 257


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	41	Resp:	1936
Ion	Ratio	Lower	Upper
41	100		
39	85.5	60.6	90.8
76	26.4	24.9	37.3





#15

Acrylonitrile

Concen: 3.424 ug/l

RT: 3.075 min Scan# 31

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 53	Resp: 334
Ion Ratio	Lower Upper
53 100	
52 82.9	65.3 97.9
51 34.4	29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#16

Acetone

Concen: 3.736 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 43	Resp: 3500
Ion Ratio	Lower Upper
43 100	
58 28.1	21.2 31.8





#17

Carbon Disulfide

Concen: 0.728 ug/l

RT: 2.508 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 76 Resp: 2619

Ion Ratio Lower Upper

76 100





78 8.0 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 0.828 ug/l

RT: 2.709 min Scan# 266

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 1852

Ion Ratio Lower Upper

43 100

74 22.0 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 0.662 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 73 Resp: 3588
Ion Ratio Lower Upper
73 100
57 20.3 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 0.828 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 84 Resp: 1571
Ion Ratio Lower Upper
84 100
49 140.0 113.9 170.9
51 51.3 33.5 50.3#
86 85.2 53.8 80.8#





#21

trans-1,2-Dichloroethene

Concen: 0.701 ug/l

RT: 3.081 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 111

Ion Ratio Lower Upper

96 100

61 119.0 119.5 179.3





98 30.1 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#22

Diisopropyl ether

Concen: 0.703 ug/l

RT: 3.757 min Scan# 438

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 45 Resp: 3857




Ion Ratio Lower Upper


45 100

43 176.2 66.6 100.0

87 28.9 19.8 29.6


59 9.1 8.6 12.8





#23
Vinyl Acetate
Concen: 3.154 ug/l
RT: 3.727 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion: 43 Resp: 1528
Ion Ratio Lower Upper
43 100
86 10.8 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 0.637 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 63 Resp: 2054
Ion Ratio Lower Upper
63 100
98 6.3 3.6 10.8
100 5.8 2.1 6.3





#25

2-Butanone

Concen: 3.371 ug/l

RT: 4.593 min Scan# 51

Delta R.T. 0.037 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 43 Resp: 4558

Ion Ratio Lower Upper

43

100

72

19.2

18.4





27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 0.728 ug/l

RT: 4.458 min Scan# 553

Delta R.T. -0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 77 Resp: 1777

Ion Ratio Lower Upper

77




100


97

13.7

10.5

31.5





#27

cis-1,2-Dichloroethene

Concen: 0.693 ug/l

RT: 4.483 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 96 Resp: 1324

Ion Ratio Lower Upper

96 100

61 145.5 0.0 322.8





98 63.1 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 0.619 ug/l m

RT: 4.910 min Scan# 627

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 49 Resp: 1061


Ion Ratio Lower Upper

49 100

129 0.0 0.0 4.0

130 65.3 56.2 84.2





#29

Tetrahydrofuran

Concen: 3.338 ug/l

RT: 5.019 min Scan# 64

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 42 Resp: 293

Ion Ratio Lower Upper

42 100

72 41.9 33.7 50.5





71 37.7 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 0.695 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 83 Resp: 2328

Ion Ratio Lower Upper

83 100

85 60.4 49.3 73.9





#32

1,1,1-Trichloroethane

Concen: 0.647 ug/l

RT: 5.379 min Scan# 70

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001

Tgt Ion: 97 Resp: 186

Ion Ratio Lower Upper

97 100

99 0.0 51.8 77.6





61 55.0 40.1 60.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

Expected RT: 5.95 min

Lab File: VX046047.D


Acq: 05 May 2025 16:27


Tgt Ion: 65

Sig Exp Ratio

65 100

67 49.5





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:114 Resp: 16702

Ion Ratio Lower Upper

114 100

63 24.2 0.0 49.2





88 17.1 0.0 33.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#35

Dibromofluoromethane

Concen: 0.000 ug/l

Expected RT: 5.38 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 113


Sig Exp Ratio

113 100

111 103.9

192 16.6





#36

1,1-Dichloropropene

Concen: 0.758 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	75	Resp:	164
Ion	Ratio	Lower	Upper
75	100		
110	28.2	16.3	48.9
77	15.8	24.3	36.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#37

Ethyl Acetate

Concen: 0.717 ug/l m




RT: 4.733 min Scan# 598


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	43	Resp:	1956
Ion	Ratio	Lower	Upper
43	100		
61	0.0	10.3	15.5#
70	1.0	7.9	11.9#





#38

Carbon Tetrachloride

Concen: 0.724 ug/l

RT: 5.678 min Scan# 752

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:117 Resp: 180

Ion Ratio Lower Upper

117 100

119 67.0 75.2 112.8





121 28.5 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#39

Methylcyclohexane

Concen: 0.770 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 83 Resp: 2096


Ion Ratio Lower Upper

83 100

55 80.1 64.7 97.1

98 41.9 37.4 56.2





#40

Benzene

Concen: 0.668 ug/l

RT: 6.044 min Scan# 811

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 78 Resp: 450

Ion Ratio Lower Upper

78 100





77 29.1 19.0 28.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#41

Methacrylonitrile

Concen: 0.513 ug/l

RT: 4.946 min Scan# 633

Delta R.T. 0.030 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 41 Resp: 778




Ion Ratio Lower Upper


41 100

39 51.7 47.2 70.8

67 0.0 50.7 76.1#

52 13.6 24.6 37.0#





#42

1,2-Dichloroethane

Concen: 0.694 ug/l

RT: 6.092 min Scan# 819

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 62 Resp: 1934

Ion Ratio Lower Upper

62 100





98 1.1 0.0 15.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#43

Isopropyl Acetate

Concen: 0.612 ug/l

RT: 6.354 min Scan# 864

Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion: 43 Resp: 2553


Ion Ratio Lower Upper

43 100

61 10.7 14.3 21.5#


87 3.2 9.5 14.3#





#44
Trichloroethene
Concen: 0.680 ug/l
RT: 7.135 min Scan# 990
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:130 Resp: 1083
Ion Ratio Lower Upper
130 100
95 117.3 0.0 204.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#45
1,2-Dichloropropane
Concen: 0.633 ug/l
RT: 7.440 min Scan# 1042
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 63 Resp: 1059
Ion Ratio Lower Upper
63 100
65 36.1 24.1 36.1





#46

Dibromomethane

Concen: 0.667 ug/l

RT: 7.586 min Scan# 1106

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	93	Resp:	880
Ion	Ratio	Lower	Upper
93	100		
95	84.5	65.6	98.4
174	88.5	68.2	102.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#47

Bromodichloromethane

Concen: 0.633 ug/l




RT: 7.818 min Scan# 1104


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	83	Resp:	1620
Ion	Ratio	Lower	Upper
83	100		
85	57.3	50.5	75.7
127	0.0	6.5	9.7





#48

Methyl methacrylate

Concen: 0.577 ug/l

RT: 7.714 min Scan# 1083

Delta R.T. 0.024 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion: 41	Resp: 123
Ion Ratio	Lower Upper
41	100
69	70.6 58.5 87.7
39	84.8 51.7 77.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#49

1,4-Dioxane

Concen: 11.703 ug/l




RT: 7.677 min Scan# 1081


Delta R.T. 0.018 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 88	Resp: 498
Ion Ratio	Lower Upper
88	100
43	44.4 33.4 50.2
58	30.3 58.6 88.0





#50
Toluene-d8
Concen: 0.000 ug/l
Expected RT: 8.65 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 98
Sig Exp Ratio
98 100
100 66.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#51
4-Methyl-2-Pentanone
Concen: 3.430 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 43 Resp: 9378
Ion Ratio Lower Upper
43 100
58 34.4 28.9 43.3









#52
Toluene
Concen: 0.672 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion: 92 Resp: 2684
Ion Ratio Lower Upper
92 100
91 165.7 136.6 205.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#53
t-1,3-Dichloropropene
Concen: 0.576 ug/l
RT: 8.988 min Scan# 1296
Delta R.T. 0.012 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27


Tgt Ion: 75 Resp: 1238
Ion Ratio Lower Upper
75 100
77 42.7 25.0 37.6#





#54
 cis-1,3-Dichloropropene
 Concen: 0.564 ug/l
 RT: 8.379 min Scan# 1196
 Delta R.T. 0.018 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDIC001



Tgt Ion: 75 Resp: 1414
 Ion Ratio Lower Upper
 75 100
 77 44.2 25.4 38.0
 39 59.0 52.2 78.4




Manual Integrations
 APPROVED


Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025



#55
 1,1,2-Trichloroethane
 Concen: 0.637 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. 0.006 min
 Lab File: VX046047.D
 Acq: 05 May 2025 16:27

Tgt Ion: 97 Resp: 1029
 Ion Ratio Lower Upper
 97 100
 83 74.8 69.0 103.6
 85 59.5 44.3 66.5
 99 63.0 47.8 71.6





#56

Ethyl methacrylate

Concen: 0.504 ug/l

RT: 9.128 min Scan# 1319

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 69 Resp: 1260

Ion Ratio Lower Upper

69 100

41 86.6 60.8 91.2





39 78.0 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#57

1,3-Dichloropropane

Concen: 0.711 ug/l

RT: 9.311 min Scan# 1349

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 76 Resp: 2036

Ion Ratio Lower Upper

76 100

78 31.4 26.3 39.5





#58

2-Chloroethyl Vinyl ether

Concen: 3.420 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 63 Resp: 3840

Ion Ratio Lower Upper

63 100





106 24.0 21.5 32.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#59

2-Hexanone

Concen: 3.096 ug/l

RT: 9.439 min Scan# 1370

Delta R.T. 0.012 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 43 Resp: 6424

Ion Ratio Lower Upper

43 100

58 41.6 24.9 74.6





#60

Dibromochloromethane

Concen: 0.582 ug/l

RT: 9.525 min Scan# 11

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:129 Resp: 102

Ion Ratio Lower Upper

129 100





127 72.5 39.3 117.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#61

1,2-Dibromoethane

Concen: 0.650 ug/l

RT: 9.610 min Scan# 1398

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:107 Resp: 1077

Ion Ratio Lower Upper

107 100

109 94.0 73.3 109.9





#62
4-Bromofluorobenzene
Concen: 0.000 ug/l
Expected RT: 11.08 min



Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 95
Sig Exp Ratio
95 100
174 67.9
176 65.7




Manual Integrations APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:117 Resp: 140108
Ion Ratio Lower Upper
117 100
82 61.0 50.6 76.0
119 32.1 25.8 38.6





#64
Tetrachloroethene

Concen: 0.679 ug/l

RT: 9.275 min Scan# 1343

Delta R.T. 0.006 min

Lab File: VX046047.D


Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001



Tgt Ion:164 Resp: 972

Ion Ratio Lower Upper

164 100

166 119.1 105.0 157.6

129 103.4 83.5 125.3




131 95.4 76.5 114.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#65

Chlorobenzene

Concen: 0.735 ug/l

RT: 10.079 min Scan# 1475

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:112 Resp: 3170

Ion Ratio Lower Upper

112 100


114 30.6 25.4 38.2





#66
1,1,1,2-Tetrachloroethane
Concen: 0.720 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001



Tgt Ion:131 Resp: 1035
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 0.0 31.6 95.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#67
Ethyl Benzene
Concen: 0.687 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion: 91 Resp: 5052
Ion Ratio Lower Upper
91 100
106 32.5 23.4 35.2









#68
m/p-Xylenes
Concen: 1.367 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001

Tgt Ion:106 Resp: 3630
Ion Ratio Lower Upper
106 100
91 209.8 171.2 256.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#69
o-Xylene
Concen: 0.672 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:106 Resp: 1799
Ion Ratio Lower Upper
106 100
91 214.9 112.7 338.1





#70

Styrene

Concen: 0.622 ug/l

RT: 10.659 min Scan# 1569

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDIC001

Tgt Ion:104	Resp:	266
Ion Ratio	Lower	Upper
104	100	
78	54.3	45.7 68.5
103	53.9	43.7 65.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#71

Bromoform

Concen: 0.608 ug/l




RT: 10.799 min Scan# 1593


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:173	Resp:	657
Ion Ratio	Lower	Upper
173	100	
175	46.6	23.4 70.0
254	0.0	0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:152 Resp: 5912

Ion Ratio Lower Upper

152 100

115 66.6 46.9 140.7





150 156.3 0.0 351.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#73

Isopropylbenzene

Concen: 0.695 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion:105 Resp: 4480

Ion Ratio Lower Upper

105 100

120 23.7 12.8 38.4





#74

N-amy1 acetate

Concen: 0.631 ug/l

RT: 10.854 min Scan# 1602

Delta R.T. 0.012 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion: 43	Resp: 2028
Ion Ratio	Lower Upper
43 100	
70 31.5	30.9 46.3
55 24.5	18.7 28.1
61 15.4	17.1 25.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#75

1,1,2,2-Tetrachloroethane

Concen: 0.800 ug/l




RT: 11.213 min Scan# 1661


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion: 83	Resp: 1835
Ion Ratio	Lower Upper
83 100	
131 7.5	5.0 14.9
85 58.7	31.9 95.7









#76
1,2,3-Trichloropropane
Concen: 0.674 ug/l m
RT: 11.244 min Scan# 1665
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Tgt Ion: 75 Resp: 1665
Ion Ratio Lower Upper
75 100
77 29.7 20.5 61.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#77
Bromobenzene
Concen: 0.734 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27

Tgt Ion:156 Resp: 1095
Ion Ratio Lower Upper
156 100
77 179.5 92.8 278.3
158 93.5 48.8 146.3





#78

n-propylbenzene

Concen: 0.699 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion: 91 Resp: 5052

Ion Ratio Lower Upper

91 100





120 20.1 10.8 32.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#79

2-Chlorotoluene

Concen: 0.773 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3765

Ion Ratio Lower Upper

91 100

126 28.0 15.6 46.7





#80

1,3,5-Trimethylbenzene

Concen: 0.671 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 3590

Ion Ratio Lower Upper

105 100





120 43.4 23.1 69.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#82

4-Chlorotoluene

Concen: 0.713 ug/l

RT: 11.457 min Scan# 1701

Delta R.T. 0.006 min

Lab File: VX046047.D




Acq: 05 May 2025 16:27


Tgt Ion: 91 Resp: 3815

Ion Ratio Lower Upper

91 100

126 24.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 0.751 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:	119	Resp:	395
Ion	Ratio	Lower	Upper
119	100		
91	63.9	32.9	98.7
134	25.2	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 0.697 ug/l




RT: 11.756 min Scan# 1750


Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:	105	Resp:	3725
Ion	Ratio	Lower	Upper
105	100		
120	40.3	21.2	63.6





#85

sec-Butylbenzene

Concen: 0.672 ug/l

RT: 11.890 min Scan# 1772

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001

Tgt Ion:105 Resp: 438

Ion Ratio Lower Upper

105 100





134 19.2 9.7 29.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#86

p-Isopropyltoluene

Concen: 0.679 ug/l

RT: 12.006 min Scan# 1791

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27




Tgt Ion:119 Resp: 3577


Ion Ratio Lower Upper

119 100

134 25.0 12.5 37.5

91 33.1 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 0.689 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDIC001





Tgt Ion:146	Resp:	1914
Ion Ratio	Lower	Upper
146	100	
111	43.9	22.1 66.3
148	65.3	32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 0.789 ug/l m




RT: 12.036 min Scan# 1796


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Tgt Ion:146	Resp:	2149
Ion Ratio	Lower	Upper
146	100	
111	39.1	21.3 63.9
148	58.2	31.9 95.5





#89

n-Butylbenzene

Concen: 0.631 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :





VSTDICC001

Tgt Ion: 91	Resp: 288
Ion Ratio	Lower Upper
91	100
92	48.1 26.9 80.7
134	20.1 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 0.616 ug/l




RT: 12.536 min Scan# 1878


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion: 117	Resp: 604
Ion Ratio	Lower Upper
117	100
201	64.2 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 0.747 ug/l
RT: 12.341 min Scan# 1846
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001




Tgt Ion:146 Resp: 202
Ion Ratio Lower Upper
146 100
111 48.8 22.4 67.3
148 59.4 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 0.586 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion: 75 Resp: 306
Ion Ratio Lower Upper
75 100
155 49.3 34.9 104.8
157 78.1 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 0.688 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.006 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC001





Tgt Ion:180	Resp:	1019
Ion Ratio	Lower	Upper
180	100	
182	100.6	47.9 143.8
145	29.5	17.1 51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#94

Hexachlorobutadiene

Concen: 0.671 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX046047.D

Acq: 05 May 2025 16:27


Tgt Ion:225	Resp:	465
Ion Ratio	Lower	Upper
225	100	
223	69.7	30.8 92.4
227	70.5	30.9 92.7





#95
Naphthalene
Concen: 0.767 ug/l
RT: 13.780 min Scan# 2081
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27




Instrument :
MSVOA_X
ClientSampleId :
VSTDIC001




Tgt Ion:128 Resp: 4138
Ion Ratio Lower Upper
128 100
127 12.4 10.4 15.6
129 8.9 8.6 13.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 0.716 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046047.D
Acq: 05 May 2025 16:27



Tgt Ion:180 Resp: 1113
Ion Ratio Lower Upper
180 100
182 92.5 47.8 143.3
145 32.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.020%		
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.360%		
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.960%		
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.200%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	79763m	48.568	ug/l	
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion:168 Resp: 9258

Ion Ratio Lower Upper

168 100





99 74.5 54.9 82.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#2

Dichlorodifluoromethane

Concen: 55.594 ug/l

RT: 1.167 min Scan# 13

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 85 Resp: 78783

Ion Ratio Lower Upper

85 100

87 32.2 15.7 47.1









#3
Chloromethane
Concen: 52.592 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 50 Resp: 72275
Ion Ratio Lower Upper
50 100
52 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#4
Vinyl Chloride
Concen: 51.598 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 62 Resp: 65994
Ion Ratio Lower Upper
62 100
64 32.3 25.2 37.8





#5
Bromomethane
Concen: 50.627 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 94 Resp: 3003
Ion Ratio Lower Upper
94 100
96 93.5 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 52.271 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 64 Resp: 35690
Ion Ratio Lower Upper
64 100
66 33.7 24.3 36.5





#7
Trichlorofluoromethane
Concen: 52.560 ug/l
RT: 1.874 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion:101 Resp: 99352
Ion Ratio Lower Upper
101 100
103 62.6 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#8
Diethyl Ether
Concen: 49.536 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 74 Resp: 31875
Ion Ratio Lower Upper
74 100
45 107.1 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 50.822 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	101	Resp:	5945
Ion	Ratio	Lower	Upper
101	100		
85	47.8	38.6	58.0
151	69.4	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#10

Methyl Iodide

Concen: 53.419 ug/l




RT: 2.447 min Scan# 223


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	142	Resp:	73940
Ion	Ratio	Lower	Upper
142	100		
127	52.9	41.7	62.5
141	14.4	11.5	17.3





#11

Tert butyl alcohol

Concen: 252.025 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 59 Resp: 6106

Ion Ratio Lower Upper

59 100





57 11.1 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#12

1,1-Dichloroethene

Concen: 51.097 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 96 Resp: 56097


Ion Ratio Lower Upper

96 100

61 179.7 146.2 219.2

98 62.1 51.0 76.6





#13

Acrolein

Concen: 232.505 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 56 Resp: 6415

Ion Ratio Lower Upper

56 100





55 70.8 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#14

Allyl chloride

Concen: 51.833 ug/l

RT: 2.660 min Scan# 258

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 41 Resp: 108757


Ion Ratio Lower Upper

41 100

39 75.6 60.6 90.8

76 31.1 24.9 37.3









#15
Acrylonitrile
Concen: 262.412 ug/l
RT: 3.063 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 53 Resp: 181808
Ion Ratio Lower Upper
53 100
52 82.9 65.3 97.9
51 36.5 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16
Acetone
Concen: 245.518 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 43 Resp: 169923
Ion Ratio Lower Upper
43 100
58 27.3 21.2 31.8





#17

Carbon Disulfide

Concen: 50.981 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 76 Resp: 13269

Ion Ratio Lower Upper

76 100





78 9.1 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 49.968 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 80251

Ion Ratio Lower Upper

43 100

74 22.1 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 50.777 ug/l

RT: 3.111 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 73 Resp: 19544

Ion Ratio Lower Upper

73 100





57 21.8 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#20

Methylene Chloride

Concen: 47.779 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion: 84 Resp: 63369




Ion Ratio Lower Upper


84 100

49 142.7 113.9 170.9

51 43.3 33.5 50.3


86 66.8 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 50.513 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 96 Resp: 55770
Ion Ratio Lower Upper
96 100
61 149.3 119.5 179.3
98 64.6 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 51.800 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 45 Resp: 209946
Ion Ratio Lower Upper
45 100
43 86.9 66.6 100.0
87 24.8 19.8 29.6
59 10.8 8.6 12.8









#23
Vinyl Acetate
Concen: 265.585 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVX050525

Tgt Ion: 43 Resp: 946740
Ion Ratio Lower Upper
43 100
86 9.4 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#24
1,1-Dichloroethane
Concen: 50.825 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 63 Resp: 114733
Ion Ratio Lower Upper
63 100
98 6.4 3.6 10.8
100 4.8 2.1 6.3





#25

2-Butanone

Concen: 257.110 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 43 Resp: 258339

Ion Ratio Lower Upper

43 100





72 24.3 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#26

2,2-Dichloropropane

Concen: 49.934 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 77 Resp: 88229

Ion Ratio Lower Upper

77 100

97 21.0 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 50.956 ug/l

RT: 4.483 min Scan# 51

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 96 Resp: 6772

Ion Ratio Lower Upper

96 100

61 160.7 0.0 322.8





98 64.4 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#28

Bromochloromethane

Concen: 50.975 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 49 Resp: 55390


Ion Ratio Lower Upper

49 100

129 1.8 0.0 4.0

130 68.7 56.2 84.2





#29

Tetrahydrofuran

Concen: 259.958 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 42 Resp: 163679

Ion Ratio Lower Upper

42 100

72 42.0 33.7 50.5





71 38.5 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 51.332 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.006 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 83 Resp: 120782

Ion Ratio Lower Upper

83 100


85 63.6 49.3 73.9





#31
Cyclohexane
Concen: 50.147 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 56 Resp: 103158
Ion Ratio Lower Upper
56 100
69 27.8 24.4 36.6
84 80.5 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#32
1,1,1-Trichloroethane
Concen: 51.108 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 97 Resp: 104243
Ion Ratio Lower Upper
97 100
99 64.5 51.8 77.6
61 50.2 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 48.509 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 65 Resp: 8373

Ion Ratio Lower Upper

65 100





67 50.0 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion:114 Resp: 162651


Ion Ratio Lower Upper

114 100

63 23.1 0.0 49.2

88 17.2 0.0 33.6





#35

Dibromofluoromethane

Concen: 49.683 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	113	Resp:	5819
Ion	Ratio	Lower	Upper
113	100		
111	102.1	83.1	124.7
192	16.2	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#36

1,1-Dichloropropene

Concen: 49.521 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	75	Resp:	77931
Ion	Ratio	Lower	Upper
75	100		
110	32.7	16.3	48.9
77	31.1	24.3	36.5





#37

Ethyl Acetate

Concen: 50.714 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :





ICVX050525

Tgt Ion: 43 Resp: 98604
Ion Ratio Lower Upper
43 100
61 12.6 10.3 15.5
70 10.0 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#38

Carbon Tetrachloride

Concen: 49.825 ug/l




RT: 5.666 min Scan# 751


Delta R.T. -0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:117 Resp: 88100
Ion Ratio Lower Upper
117 100
119 90.7 75.2 112.8
121 29.4 24.2 36.4





#39

Methylcyclohexane

Concen: 49.757 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 83 Resp: 100808

Ion Ratio Lower Upper

83 100

55 82.7 64.7 97.1





98 48.3 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 50.759 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 78 Resp: 233977

Ion Ratio Lower Upper

78 100

77 24.0 19.0 28.4









#41
Methacrylonitrile
Concen: 55.076 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:	41	Resp:	5601
Ion	Ratio	Lower	Upper
41	100		
39	58.7	47.2	70.8
67	63.7	50.7	76.1
52	30.5	24.6	37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 50.861 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:	62	Resp:	101185
Ion	Ratio	Lower	Upper
62	100		
98	7.1	0.0	15.2





#43

Isopropyl Acetate

Concen: 52.195 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 43 Resp: 154818

Ion Ratio Lower Upper

43 100

61 17.9 14.3 21.5





87 11.4 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#44

Trichloroethene

Concen: 50.018 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:130 Resp: 55491

Ion Ratio Lower Upper

130 100

95 102.6 0.0 204.2





#45

1,2-Dichloropropane

Concen: 51.983 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 63 Resp: 5958

Ion Ratio Lower Upper

63 100





65 31.4 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#46

Dibromomethane

Concen: 49.978 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 93 Resp: 45181


Ion Ratio Lower Upper

93 100

95 83.0 65.6 98.4

174 84.8 68.2 102.2





#47

Bromodichloromethane

Concen: 51.889 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	83	Resp:	9238
Ion	Ratio	Lower	Upper
83	100		
85	63.0	50.5	75.7
127	8.8	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#48

Methyl methacrylate

Concen: 54.226 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	41	Resp:	82144
Ion	Ratio	Lower	Upper
41	100		
69	70.4	58.5	87.7
39	64.2	51.7	77.5





#49

1,4-Dioxane

Concen: 1059.599 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525





Tgt Ion:	88	Resp:	3047
Ion	Ratio	Lower	Upper
88	100		
43	39.9	33.4	50.2
58	76.2	58.6	88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#50

Toluene-d8

Concen: 48.984 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	98	Resp:	198575
Ion	Ratio	Lower	Upper
98	100		
100	65.6	53.5	80.3









#51
4-Methyl-2-Pentanone
Concen: 259.133 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion: 43 Resp: 510200
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 51.150 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 92 Resp: 144574
Ion Ratio Lower Upper
92 100
91 171.4 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 52.920 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 75 Resp: 8375

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 52.892 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 75 Resp: 92517


Ion Ratio Lower Upper

75 100

77 31.5 25.4 38.0

39 63.2 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 50.887 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 97 Resp: 5671

Ion Ratio Lower Upper

97 100

83 89.8 69.0 103.6

85 58.4 44.3 66.5





99 63.9 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#56

Ethyl methacrylate

Concen: 53.918 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50




Tgt Ion: 69 Resp: 95772


Ion Ratio Lower Upper

69 100

41 76.1 60.8 91.2

39 47.4 39.0 58.6





#57

1,3-Dichloropropane

Concen: 49.793 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525

Tgt Ion: 76 Resp: 9966

Ion Ratio Lower Upper

76 100





78 31.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 265.780 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion: 63 Resp: 240681

Ion Ratio Lower Upper

63 100

106 26.6 21.5 32.3





#59

2-Hexanone

Concen: 268.218 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 43 Resp: 39070

Ion Ratio Lower Upper

43 100





58 49.6 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#60

Dibromochloromethane

Concen: 52.075 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:129 Resp: 63738

Ion Ratio Lower Upper

129 100

127 77.9 39.3 117.8









#61
1,2-Dibromoethane
Concen: 51.050 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:107 Resp: 59134
Ion Ratio Lower Upper
107 100
109 93.8 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 48.596 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 95 Resp: 75568
Ion Ratio Lower Upper
95 100
174 68.2 0.0 135.8
176 66.5 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:117 Resp: 145078
Ion Ratio Lower Upper
117 100
82 62.1 50.6 76.0
119 32.9 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 46.795 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:164 Resp: 48034
Ion Ratio Lower Upper
164 100
166 129.9 105.0 157.6
129 105.3 83.5 125.3
131 98.8 76.5 114.7





#65
Chlorobenzene
Concen: 48.699 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:112 Resp: 154640
Ion Ratio Lower Upper
112 100
114 31.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.557 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:131 Resp: 53735
Ion Ratio Lower Upper
131 100
133 94.9 47.3 141.9
119 63.7 31.6 95.0





#67

Ethyl Benzene

Concen: 50.245 ug/l

RT: 10.189 min Scan# 1493

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 91 Resp: 281239

Ion Ratio Lower Upper

91 100





106 29.1 23.4 35.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#68

m/p-Xylenes

Concen: 101.022 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:106 Resp: 206813

Ion Ratio Lower Upper

106 100

91 215.6 171.2 256.8









#69
o-Xylene
Concen: 50.841 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:106 Resp: 101469
Ion Ratio Lower Upper
106 100
91 227.7 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 51.907 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:104 Resp: 169703
Ion Ratio Lower Upper
104 100
78 57.6 45.7 68.5
103 55.5 43.7 65.5





#71
Bromoform
Concen: 49.583 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:173 Resp: 40364
Ion Ratio Lower Upper
173 100
175 48.8 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:152 Resp: 68229
Ion Ratio Lower Upper
152 100
115 93.3 46.9 140.7
150 175.3 0.0 351.0





#73
Isopropylbenzene
Concen: 50.733 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:105 Resp: 269484
Ion Ratio Lower Upper
105 100
120 25.4 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 52.407 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 43 Resp: 137559
Ion Ratio Lower Upper
43 100
70 37.9 30.9 46.3
55 22.7 18.7 28.1
61 21.8 17.1 25.7





#75
1,1,2,2-Tetrachloroethane
Concen: 47.827 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion: 83 Resp: 8902
Ion Ratio Lower Upper
83 100
131 9.7 5.0 14.9
85 65.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 48.568 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 75 Resp: 79763
Ion Ratio Lower Upper
75 100
77 40.8 20.5 61.5





#77

Bromobenzene

Concen: 48.549 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525





Tgt Ion:	156	Resp:	5987
Ion	Ratio	Lower	Upper
156	100		
77	189.0	92.8	278.3
158	98.9	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#78

n-propylbenzene

Concen: 50.562 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Tgt Ion:	91	Resp:	312285
Ion	Ratio	Lower	Upper
91	100		
120	21.4	10.8	32.4





#79

2-Chlorotoluene

Concen: 49.107 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046048.D

Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVVX050525

Tgt Ion: 91 Resp: 195630

Ion Ratio Lower Upper

91 100





126 30.8 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#80

1,3,5-Trimethylbenzene

Concen: 51.500 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 228538

Ion Ratio Lower Upper

105 100


120 46.3 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 49.276 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVX050525



Tgt Ion: 75 Resp: 2485
Ion Ratio Lower Upper
75 100
53 122.3 100.5 150.7
89 46.2 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 50.661 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion: 91 Resp: 223811
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8









#83
tert-Butylbenzene
Concen: 50.511 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Tgt Ion:119 Resp: 22578
Ion Ratio Lower Upper
119 100
91 64.9 32.9 98.7
134 22.6 11.4 34.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#84
1,2,4-Trimethylbenzene
Concen: 51.485 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:105 Resp: 231371
Ion Ratio Lower Upper
105 100
120 42.3 21.2 63.6





#85
sec-Butylbenzene
Concen: 51.122 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:105 Resp: 280578
Ion Ratio Lower Upper
105 100
134 19.1 9.7 29.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#86
p-Isopropyltoluene
Concen: 51.198 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:119 Resp: 231942
Ion Ratio Lower Upper
119 100
134 25.0 12.5 37.5
91 27.5 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 50.272 ug/l
RT: 11.963 min Scan# 1784
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:146 Resp: 113140
Ion Ratio Lower Upper
146 100
111 44.1 22.1 66.3
148 63.5 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#88
1,4-Dichlorobenzene
Concen: 48.368 ug/l
RT: 12.037 min Scan# 1796
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:146 Resp: 111173
Ion Ratio Lower Upper
146 100
111 43.9 21.3 63.9
148 63.3 31.9 95.5





#89

n-Butylbenzene

Concen: 51.804 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046048.D


Acq: 05 May 2025 16:50

Instrument :

MSVOA_X

ClientSampleId :

ICVX050525



Tgt Ion: 91 Resp: 20585

Ion Ratio Lower Upper

91 100

92 53.7 26.9 80.7




134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 49.895 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046048.D




Acq: 05 May 2025 16:50


Tgt Ion:117 Resp: 39823

Ion Ratio Lower Upper

117 100


201 62.4 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 49.361 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:146 Resp: 11148
Ion Ratio Lower Upper
146 100
111 45.8 22.4 67.3
148 64.2 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.458 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion: 75 Resp: 21220
Ion Ratio Lower Upper
75 100
155 68.6 34.9 104.8
157 86.7 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 49.662 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:180 Resp: 6442
Ion Ratio Lower Upper
180 100
182 95.5 47.9 143.8
145 34.4 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 48.408 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50


Tgt Ion:225 Resp: 27425
Ion Ratio Lower Upper
225 100
223 63.1 30.8 92.4
227 63.7 30.9 92.7





#95
Naphthalene
Concen: 50.404 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50




Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525



Tgt Ion:128 Resp: 239809
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 48.525 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046048.D
Acq: 05 May 2025 16:50

Tgt Ion:180 Resp: 64950
Ion Ratio Lower Upper
180 100
182 97.0 47.8 143.3
145 37.5 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.765	0.851	-11.2	94	0.00
3 P	Chloromethane	0.742	0.781	-5.3	96	0.00
4 C	Vinyl Chloride	0.691	0.713	-3.2#	96	0.00
5 T	Bromomethane	0.320	0.324	-1.3	95	0.00
6 T	Chloroethane	0.369	0.385	-4.3	97	0.00
7 T	Trichlorofluoromethane	1.021	1.073	-5.1	96	0.00
8 T	Diethyl Ether	0.347	0.344	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.642	-1.6	96	0.00
10 T	Methyl Iodide	0.747	0.799	-7.0	94	0.00
11 T	Tert butyl alcohol	0.131	0.132	-0.8	98	0.00
12 CM	1,1-Dichloroethene	0.593	0.606	-2.2#	96	0.00
13 T	Acrolein	0.149	0.139	6.7	87	0.00
14 T	Allyl chloride	1.133	1.175	-3.7	95	0.00
15 T	Acrylonitrile	0.374	0.393	-5.1	97	0.00
16 T	Acetone	0.374	0.367	1.9	97	0.00
17 T	Carbon Disulfide	1.406	1.433	-1.9	94	0.00
18 T	Methyl Acetate	0.867	0.867	0.0	97	0.00
19 T	Methyl tert-butyl Ether	2.079	2.111	-1.5	93	0.00
20 T	Methylene Chloride	0.716	0.684	4.5	95	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.602	-1.0	94	0.00
22 T	Diisopropyl ether	2.189	2.268	-3.6	95	0.00
23 T	Vinyl Acetate	1.925	2.045	-6.2	95	0.00
24 P	1,1-Dichloroethane	1.219	1.239	-1.6	94	0.00
25 T	2-Butanone	0.543	0.558	-2.8	96	0.00
26 T	2,2-Dichloropropane	0.954	0.953	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.731	-1.8	95	0.00
28 T	Bromochloromethane	0.587	0.598	-1.9	99	0.00
29 T	Tetrahydrofuran	0.340	0.354	-4.1	96	0.00
30 C	Chloroform	1.271	1.305	-2.7#	96	0.00
31 T	Cyclohexane	1.111	1.114	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	1.101	1.126	-2.3	95	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.904	3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.360	0.358	0.6	98	0.00
36 T	1,1-Dichloropropene	0.484	0.479	1.0	94	0.00
37 T	Ethyl Acetate	0.598	0.606	-1.3	96	0.00
38 T	Carbon Tetrachloride	0.544	0.542	0.4	94	0.00
39 T	Methylcyclohexane	0.623	0.620	0.5	94	0.00
40 TM	Benzene	1.417	1.439	-1.6	95	0.00
41 T	Methacrylonitrile	0.313	0.344	-9.9	96	0.00
42 TM	1,2-Dichloroethane	0.612	0.622	-1.6	96	0.00
43 T	Isopropyl Acetate	0.912	0.952	-4.4	96	0.00
44 TM	Trichloroethene	0.341	0.341	0.0	93	0.00
45 C	1,2-Dichloropropane	0.352	0.366	-4.0#	96	0.00
46 T	Dibromomethane	0.278	0.278	0.0	94	0.00
47 T	Bromodichloromethane	0.547	0.568	-3.8	95	0.00
48 T	Methyl methacrylate	0.466	0.505	-8.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.246	1.221	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.627	-3.6	96	0.00
52 CM	Toluene	0.869	0.889	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.487	0.515	-5.7	94	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.569	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	0.343	0.349	-1.7	95	0.00
56 T	Ethyl methacrylate	0.546	0.589	-7.9	96	0.00
57 T	1,3-Dichloropropane	0.615	0.613	0.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.296	-6.5	93	0.00
59 T	2-Hexanone	0.448	0.480	-7.1	98	0.00
60 T	Dibromochloromethane	0.376	0.392	-4.3	95	0.00
61 T	1,2-Dibromoethane	0.356	0.364	-2.2	94	0.00
62 S	4-Bromofluorobenzene	0.478	0.465	2.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.354	0.331	6.5	88	0.00
65 PM	Chlorobenzene	1.094	1.066	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.370	1.1	94	0.00
67 C	Ethyl Benzene	1.929	1.939	-0.5#	95	0.00
68 T	m/p-Xylenes	0.706	0.713	-1.0	95	0.00
69 T	o-Xylene	0.688	0.699	-1.6	95	0.00
70 T	Styrene	1.127	1.170	-3.8	95	0.00
71 P	Bromoform	0.281	0.278	1.1	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.893	3.950	-1.5	96	0.00
74 T	N-amyl acetate	1.924	2.016	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.305	4.3	98	0.00
76 T	1,2,3-Trichloropropane	1.204	1.169	2.9	99	0.00
77 T	Bromobenzene	0.904	0.878	2.9	95	0.00
78 T	n-propylbenzene	4.526	4.577	-1.1	95	0.00
79 T	2-Chlorotoluene	2.919	2.867	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.350	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.364	1.6	95	0.00
82 T	4-Chlorotoluene	3.238	3.280	-1.3	96	0.00
83 T	tert-Butylbenzene	3.276	3.309	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.391	-3.0	97	0.00
85 T	sec-Butylbenzene	4.022	4.112	-2.2	96	0.00
86 T	p-Isopropyltoluene	3.320	3.399	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	1.649	1.658	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	1.684	1.629	3.3	97	0.00
89 T	n-Butylbenzene	2.912	3.017	-3.6	96	0.00
90 T	Hexachloroethane	0.585	0.584	0.2	94	0.00
91 T	1,2-Dichlorobenzene	1.655	1.634	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.311	-3.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.944	0.7	97	0.00
94 T	Hexachlorobutadiene	0.415	0.402	3.1	95	0.00
95 T	Naphthalene	3.487	3.515	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.952	3.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	55.594	-11.2	94	0.00
3 P	Chloromethane	50.000	52.592	-5.2	96	0.00
4 C	Vinyl Chloride	50.000	51.598	-3.2#	96	0.00
5 T	Bromomethane	50.000	50.627	-1.3	95	0.00
6 T	Chloroethane	50.000	52.271	-4.5	97	0.00
7 T	Trichlorofluoromethane	50.000	52.560	-5.1	96	0.00
8 T	Diethyl Ether	50.000	49.536	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.822	-1.6	96	0.00
10 T	Methyl Iodide	50.000	53.419	-6.8	94	0.00
11 T	Tert butyl alcohol	250.000	252.025	-0.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	51.097	-2.2#	96	0.00
13 T	Acrolein	250.000	232.505	7.0	87	0.00
14 T	Allyl chloride	50.000	51.833	-3.7	95	0.00
15 T	Acrylonitrile	250.000	262.412	-5.0	97	0.00
16 T	Acetone	250.000	245.518	1.8	97	0.00
17 T	Carbon Disulfide	50.000	50.981	-2.0	94	0.00
18 T	Methyl Acetate	50.000	49.968	0.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.777	-1.6	93	0.00
20 T	Methylene Chloride	50.000	47.779	4.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.513	-1.0	94	0.00
22 T	Diisopropyl ether	50.000	51.800	-3.6	95	0.00
23 T	Vinyl Acetate	250.000	265.585	-6.2	95	0.00
24 P	1,1-Dichloroethane	50.000	50.825	-1.7	94	0.00
25 T	2-Butanone	250.000	257.110	-2.8	96	0.00
26 T	2,2-Dichloropropane	50.000	49.934	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.956	-1.9	95	0.00
28 T	Bromochloromethane	50.000	50.975	-2.0	99	0.00
29 T	Tetrahydrofuran	250.000	259.958	-4.0	96	0.00
30 C	Chloroform	50.000	51.332	-2.7#	96	0.00
31 T	Cyclohexane	50.000	50.147	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.108	-2.2	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.509	3.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.683	0.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.521	1.0	94	0.00
37 T	Ethyl Acetate	50.000	50.714	-1.4	96	0.00
38 T	Carbon Tetrachloride	50.000	49.825	0.3	94	0.00
39 T	Methylcyclohexane	50.000	49.757	0.5	94	0.00
40 TM	Benzene	50.000	50.759	-1.5	95	0.00
41 T	Methacrylonitrile	50.000	55.076	-10.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.861	-1.7	96	0.00
43 T	Isopropyl Acetate	50.000	52.195	-4.4	96	0.00
44 TM	Trichloroethene	50.000	50.018	-0.0	93	0.00
45 C	1,2-Dichloropropane	50.000	51.983	-4.0#	96	0.00
46 T	Dibromomethane	50.000	49.978	0.0	94	0.00
47 T	Bromodichloromethane	50.000	51.889	-3.8	95	0.00
48 T	Methyl methacrylate	50.000	54.226	-8.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1059.599	-6.0	100	0.00
50 S	Toluene-d8	50.000	48.984	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.133	-3.7	96	0.00
52 CM	Toluene	50.000	51.150	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.920	-5.8	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.892	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	50.000	50.887	-1.8	95	0.00
56 T	Ethyl methacrylate	50.000	53.918	-7.8	96	0.00
57 T	1,3-Dichloropropane	50.000	49.793	0.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.780	-6.3	93	0.00
59 T	2-Hexanone	250.000	268.218	-7.3	98	0.00
60 T	Dibromochloromethane	50.000	52.075	-4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.050	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.596	2.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	46.795	6.4	88	0.00
65 PM	Chlorobenzene	50.000	48.699	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.557	0.9	94	0.00
67 C	Ethyl Benzene	50.000	50.245	-0.5#	95	0.00
68 T	m/p-Xylenes	100.000	101.022	-1.0	95	0.00
69 T	o-Xylene	50.000	50.841	-1.7	95	0.00
70 T	Styrene	50.000	51.907	-3.8	95	0.00
71 P	Bromoform	50.000	49.583	0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.733	-1.5	96	0.00
74 T	N-ethyl acetate	50.000	52.407	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.827	4.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.568	2.9	99	0.00
77 T	Bromobenzene	50.000	48.549	2.9	95	0.00
78 T	n-propylbenzene	50.000	50.562	-1.1	95	0.00
79 T	2-Chlorotoluene	50.000	49.107	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.500	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.276	1.4	95	0.00
82 T	4-Chlorotoluene	50.000	50.661	-1.3	96	0.00
83 T	tert-Butylbenzene	50.000	50.511	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.485	-3.0	97	0.00
85 T	sec-Butylbenzene	50.000	51.122	-2.2	96	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.272	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.368	3.3	97	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	96	0.00
90 T	Hexachloroethane	50.000	49.895	0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.361	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.458	-2.9	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.662	0.7	97	0.00
94 T	Hexachlorobutadiene	50.000	48.408	3.2	95	0.00
95 T	Naphthalene	50.000	50.404	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.525	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ARAM01

Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG No.: Q2149

Instrument ID: MSVOA_X Calibration Date/Time: 05/30/2025 10:14

Lab File ID: VX046409.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.691	0.722		4.49	20
1,1-Dichloroethene	0.593	0.644		8.6	20
2-Butanone	0.543	0.600		10.5	20
Carbon Tetrachloride	0.544	0.593		9.01	20
Chloroform	1.271	1.387		9.13	20
Benzene	1.417	1.556		9.81	20
1,2-Dichloroethane	0.612	0.666		8.82	20
Trichloroethene	0.341	0.373		9.38	20
Tetrachloroethene	0.354	0.397		12.15	20
Chlorobenzene	1.094	1.183	0.3	8.14	20
1,2-Dichloroethane-d4	0.932	0.920		-1.29	20
Dibromofluoromethane	0.360	0.375		4.17	20
Toluene-d8	1.246	1.233		-1.04	20
4-Bromofluorobenzene	0.478	0.485		1.46	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046409.D
 Acq On : 30 May 2025 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/02/2025
 Supervised By : Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:22:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	91458	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	155659	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	84132	49.342	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.680%		
35) Dibromofluoromethane	5.373	113	58349	52.055	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.120%		
50) Toluene-d8	8.641	98	191933	49.472	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.940%		
62) 4-Bromofluorobenzene	11.079	95	75501	50.734	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	101.460%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	74890	53.499	ug/l	96
3) Chloromethane	1.301	50	73284	53.985	ug/l	98
4) Vinyl Chloride	1.374	62	66063	52.290	ug/l	99
5) Bromomethane	1.593	94	29432	50.227	ug/l	100
6) Chloroethane	1.666	64	36390	53.955	ug/l	98
7) Trichlorofluoromethane	1.874	101	102733	55.020	ug/l	99
8) Diethyl Ether	2.130	74	32852	51.685	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	64429	55.758	ug/l	98
10) Methyl Iodide	2.441	142	73946	54.083	ug/l	100
11) Tert butyl alcohol	2.971	59	60217	251.596	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58940	54.349	ug/l	96
13) Acrolein	2.233	56	73354	269.119	ug/l	98
14) Allyl chloride	2.654	41	118430	57.140	ug/l	97
15) Acrylonitrile	3.056	53	191001	279.087	ug/l	98
16) Acetone	2.380	43	192269	281.238	ug/l	99
17) Carbon Disulfide	2.502	76	136592	53.127	ug/l	100
18) Methyl Acetate	2.703	43	107277	67.621	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	206503	54.313	ug/l	100
20) Methylene Chloride	2.782	84	65272	49.822	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	59084	54.176	ug/l	95
22) Diisopropyl ether	3.751	45	224967	56.192	ug/l	99
23) Vinyl Acetate	3.715	43	959795	272.572	ug/l	100
24) 1,1-Dichloroethane	3.599	63	123217	55.257	ug/l	99
25) 2-Butanone	4.544	43	274497	276.566	ug/l	100
26) 2,2-Dichloropropane	4.465	77	95191	54.540	ug/l	100
27) cis-1,2-Dichloroethene	4.477	96	71161	54.202	ug/l	99
28) Bromochloromethane	4.885	49	54104	50.407	ug/l	98
29) Tetrahydrofuran	4.995	42	170240	273.726	ug/l	98
30) Chloroform	5.080	83	126809	54.560	ug/l	99
31) Cyclohexane	5.458	56	107618	52.962	ug/l	99
32) 1,1,1-Trichloroethane	5.367	97	108862	54.032	ug/l	99
36) 1,1-Dichloropropene	5.678	75	81871	54.361	ug/l	99
37) Ethyl Acetate	4.702	43	102307	54.982	ug/l	99
38) Carbon Tetrachloride	5.666	117	92271	54.528	ug/l	97
39) Methylcyclohexane	7.373	83	105648	54.488	ug/l	98
40) Benzene	6.025	78	242267	54.919	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046409.D
 Acq On : 30 May 2025 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/02/2025
 Supervised By : Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:22:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.910	41	59869	61.507	ug/l	98
42) 1,2-Dichloroethane	6.074	62	103627	54.429	ug/l	99
43) Isopropyl Acetate	6.330	43	159524	56.197	ug/l	99
44) Trichloroethene	7.117	130	58068	54.691	ug/l	94
45) 1,2-Dichloropropane	7.421	63	61502	56.068	ug/l	95
46) Dibromomethane	7.574	93	47561	54.974	ug/l	99
47) Bromodichloromethane	7.812	83	96856	56.841	ug/l	99
48) Methyl methacrylate	7.684	41	83228	57.410	ug/l	99
49) 1,4-Dioxane	7.653	88	28493	1035.118	ug/l	97
51) 4-Methyl-2-Pentanone	8.568	43	527696	280.055	ug/l	99
52) Toluene	8.714	92	146396	54.121	ug/l	99
53) t-1,3-Dichloropropene	8.970	75	86367	57.024	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	96608	57.712	ug/l	94
55) 1,1,2-Trichloroethane	9.147	97	58592	54.934	ug/l	98
56) Ethyl methacrylate	9.110	69	97109	57.126	ug/l	99
57) 1,3-Dichloropropane	9.305	76	102354	53.435	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.232	63	239108	275.903	ug/l	100
59) 2-Hexanone	9.427	43	391463	280.812	ug/l	100
60) Dibromochloromethane	9.519	129	67824	57.902	ug/l	97
61) 1,2-Dibromoethane	9.604	107	61152	55.163	ug/l	99
64) Tetrachloroethene	9.269	164	51973	56.142	ug/l	96
65) Chlorobenzene	10.073	112	154770	54.044	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	55431	56.685	ug/l	99
67) Ethyl Benzene	10.189	91	289173	57.284	ug/l	100
68) m/p-Xylenes	10.299	106	209150	113.281	ug/l	98
69) o-Xylene	10.634	106	102312	56.842	ug/l	98
70) Styrene	10.653	104	173337	58.788	ug/l	99
71) Bromoform	10.793	173	42852	58.367	ug/l #	98
73) Isopropylbenzene	10.957	105	274433	54.621	ug/l	100
74) N-amyl acetate	10.842	43	136906	55.143	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	90062	51.151	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	84137m	54.163	ug/l	
77) Bromobenzene	11.195	156	62435	53.525	ug/l	99
78) n-propylbenzene	11.299	91	326575	55.901	ug/l	100
79) 2-Chlorotoluene	11.360	91	198028	52.554	ug/l	100
80) 1,3,5-Trimethylbenzene	11.445	105	230250	54.855	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.012	75	26906	56.390	ug/l	98
82) 4-Chlorotoluene	11.451	91	230030	55.048	ug/l	99
83) tert-Butylbenzene	11.713	119	227701	53.855	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	235959	55.511	ug/l	99
85) sec-Butylbenzene	11.884	105	291277	56.108	ug/l	100
86) p-Isopropyltoluene	12.006	119	243737	56.880	ug/l	99
87) 1,3-Dichlorobenzene	11.963	146	116500	54.725	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	115980	53.346	ug/l	100
89) n-Butylbenzene	12.329	91	224309	59.677	ug/l	100
90) Hexachloroethane	12.536	117	43197	57.220	ug/l	99
91) 1,2-Dichlorobenzene	12.329	146	114383	53.544	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21197	54.344	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	69193	56.393	ug/l	99
94) Hexachlorobutadiene	13.719	225	30357	56.649	ug/l	97
95) Naphthalene	13.774	128	246917	54.869	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	69047	54.538	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046409.D
Acq On : 30 May 2025 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:22:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046409.D
Acq On : 30 May 2025 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:22:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.538 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 91458
Ion Ratio Lower Upper
168 100
99 70.1 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#2
Dichlorodifluoromethane
Concen: 53.499 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 85 Resp: 74890
Ion Ratio Lower Upper
85 100
87 33.4 15.7 47.1









#3
Chloromethane
Concen: 53.985 ug/l
RT: 1.301 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 73284
Ion Ratio Lower Upper
50 100
52 32.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#4
Vinyl Chloride
Concen: 52.290 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 62 Resp: 66063
Ion Ratio Lower Upper
62 100
64 31.9 25.2 37.8









#5
Bromomethane
Concen: 50.227 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 94 Resp: 29432
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#6
Chloroethane
Concen: 53.955 ug/l
RT: 1.666 min Scan# 95
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14


Tgt Ion: 64 Resp: 36390
Ion Ratio Lower Upper
64 100
66 31.4 24.3 36.5





#7
Trichlorofluoromethane
Concen: 55.020 ug/l
RT: 1.874 min Scan# 110
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:101 Resp: 10273
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#8
Diethyl Ether
Concen: 51.685 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 74 Resp: 32852
Ion Ratio Lower Upper
74 100
45 106.7 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 55.758 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 64429
Ion Ratio Lower Upper
101 100
85 46.6 38.6 58.0
151 70.0 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#10
Methyl Iodide
Concen: 54.083 ug/l
RT: 2.441 min Scan# 222
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion:142 Resp: 73946
Ion Ratio Lower Upper
142 100
127 52.4 41.7 62.5
141 14.3 11.5 17.3





#11

Tert butyl alcohol

Concen: 251.596 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 6021

Ion Ratio Lower Upper

59 100





57 10.8 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#12

1,1-Dichloroethene

Concen: 54.349 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 96 Resp: 58940


Ion Ratio Lower Upper

96 100

61 176.0 146.2 219.2

98 62.7 51.0 76.6





#13

Acrolein

Concen: 269.119 ug/l

RT: 2.233 min Scan# 188

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 56 Resp: 73354

Ion Ratio Lower Upper

56 100





55 71.9 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#14

Allyl chloride

Concen: 57.140 ug/l

RT: 2.654 min Scan# 257

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 41 Resp: 118430


Ion Ratio Lower Upper

41 100

39 73.2 60.6 90.8

76 29.4 24.9 37.3









#15
Acrylonitrile
Concen: 279.087 ug/l
RT: 3.056 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 53 Resp: 19100
Ion Ratio Lower Upper
53 100
52 83.7 65.3 97.9
51 37.7 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#16
Acetone
Concen: 281.238 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 43 Resp: 192269
Ion Ratio Lower Upper
43 100
58 27.0 21.2 31.8





#17

Carbon Disulfide

Concen: 53.127 ug/l

RT: 2.502 min Scan# 21

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 136592

Ion Ratio Lower Upper

76 100





78 8.8 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#18

Methyl Acetate

Concen: 67.621 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 43 Resp: 107277

Ion Ratio Lower Upper

43 100

74 21.3 16.7 25.1





#19

Methyl tert-butyl Ether

Concen: 54.313 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 73 Resp: 20650

Ion Ratio Lower Upper

73 100





57 22.1 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#20

Methylene Chloride

Concen: 49.822 ug/l

RT: 2.782 min Scan# 278

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Tgt Ion: 84 Resp: 65272




Ion Ratio Lower Upper


84 100

49 138.8 113.9 170.9

51 41.9 33.5 50.3

86 62.5 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 54.176 ug/l

RT: 3.081 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 59084

Ion Ratio Lower Upper

96 100

61 157.0 119.5 179.3





98 65.0 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#22

Diisopropyl ether

Concen: 56.192 ug/l

RT: 3.751 min Scan# 437

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Tgt Ion: 45 Resp: 224967




Ion Ratio Lower Upper


45 100

43 84.8 66.6 100.0

87 24.9 19.8 29.6

59 10.6 8.6 12.8









#23
Vinyl Acetate
Concen: 272.572 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 95979
Ion Ratio Lower Upper
43 100
86 9.2 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#24
1,1-Dichloroethane
Concen: 55.257 ug/l
RT: 3.599 min Scan# 412
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 63 Resp: 123217
Ion Ratio Lower Upper
63 100
98 7.0 3.6 10.8
100 4.0 2.1 6.3





#25

2-Butanone

Concen: 276.566 ug/l

RT: 4.544 min Scan# 50

Delta R.T. -0.012 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 27449

Ion Ratio Lower Upper

43 100





72 22.9 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#26

2,2-Dichloropropane

Concen: 54.540 ug/l

RT: 4.465 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 77 Resp: 95191

Ion Ratio Lower Upper

77 100

97 20.9 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 54.202 ug/l

RT: 4.477 min Scan# 511

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 7116

Ion Ratio Lower Upper

96 100

61 162.5 0.0 322.8





98 65.2 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#28

Bromochloromethane

Concen: 50.407 ug/l

RT: 4.885 min Scan# 623

Delta R.T. -0.012 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 49 Resp: 54104


Ion Ratio Lower Upper

49 100

129 1.6 0.0 4.0

130 68.9 56.2 84.2





#29

Tetrahydrofuran

Concen: 273.726 ug/l

RT: 4.995 min Scan# 641

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 42 Resp: 170240

Ion Ratio Lower Upper

42 100

72 41.2 33.7 50.5





71 37.2 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#30

Chloroform

Concen: 54.560 ug/l

RT: 5.080 min Scan# 655

Delta R.T. -0.006 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 83 Resp: 126809

Ion Ratio Lower Upper

83 100


85 62.6 49.3 73.9





#31
Cyclohexane
Concen: 52.962 ug/l
RT: 5.458 min Scan# 718
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 56 Resp: 107618
Ion Ratio Lower Upper
56 100
69 29.7 24.4 36.6
84 82.3 66.9 100.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#32
1,1,1-Trichloroethane
Concen: 54.032 ug/l
RT: 5.367 min Scan# 702
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 97 Resp: 108862
Ion Ratio Lower Upper
97 100
99 65.0 51.8 77.6
61 51.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 49.342 ug/l

RT: 5.946 min Scan# 798

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 8413

Ion Ratio Lower Upper

65 100





67 49.8 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.751 min Scan# 929

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion:114 Resp: 155659


Ion Ratio Lower Upper

114 100

63 24.8 0.0 49.2

88 17.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 52.055 ug/l

RT: 5.373 min Scan# 704

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:113 Resp: 58349

Ion Ratio Lower Upper

113 100

111 103.3 83.1 124.7





192 16.6 13.3 19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#36

1,1-Dichloropropene

Concen: 54.361 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 75 Resp: 81871


Ion Ratio Lower Upper

75 100

110 32.9 16.3 48.9

77 30.8 24.3 36.5





#37

Ethyl Acetate

Concen: 54.982 ug/l

RT: 4.702 min Scan# 593

Delta R.T. -0.012 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 10230

Ion Ratio Lower Upper

43 100

61 12.4 10.3 15.5





70 9.7 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#38

Carbon Tetrachloride

Concen: 54.528 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion:117 Resp: 92271


Ion Ratio Lower Upper

117 100

119 98.1 75.2 112.8

121 30.5 24.2 36.4





#39

Methylcyclohexane

Concen: 54.488 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 105648

Ion Ratio Lower Upper

83 100

55 82.8 64.7 97.1





98 48.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#40

Benzene

Concen: 54.919 ug/l

RT: 6.025 min Scan# 810

Delta R.T. -0.006 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 78 Resp: 242267

Ion Ratio Lower Upper

78 100


77 23.9 19.0 28.4





#41
Methacrylonitrile
Concen: 61.507 ug/l
RT: 4.910 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 5986
Ion Ratio Lower Upper
41 100
39 56.4 47.2 70.8
67 62.3 50.7 76.1
52 30.4 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#42
1,2-Dichloroethane
Concen: 54.429 ug/l
RT: 6.074 min Scan# 818
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 62 Resp: 103627
Ion Ratio Lower Upper
62 100
98 7.4 0.0 15.2





#43

Isopropyl Acetate

Concen: 56.197 ug/l

RT: 6.330 min Scan# 861

Delta R.T. -0.006 min

Lab File: VX046409.D


Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050






Tgt Ion: 43	Resp: 159524
Ion Ratio	Lower Upper
43	100
61	17.4 14.3 21.5
87	11.3 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#44

Trichloroethene

Concen: 54.691 ug/l




RT: 7.117 min Scan# 989


Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Tgt Ion:130	Resp: 58068
Ion Ratio	Lower Upper
130	100
95	108.4 0.0 204.2





#45

1,2-Dichloropropane

Concen: 56.068 ug/l

RT: 7.421 min Scan# 1039

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 61502

Ion Ratio Lower Upper

63 100





65 33.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#46

Dibromomethane

Concen: 54.974 ug/l

RT: 7.574 min Scan# 1064

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 93 Resp: 47561


Ion Ratio Lower Upper

93 100

95 81.8 65.6 98.4

174 83.1 68.2 102.2





#47

Bromodichloromethane

Concen: 56.841 ug/l

RT: 7.812 min Scan# 1104

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:	83	Resp:	9685
Ion	Ratio	Lower	Upper
83	100		
85	62.6	50.5	75.7
127	8.2	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#48

Methyl methacrylate

Concen: 57.410 ug/l




RT: 7.684 min Scan# 1082


Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Tgt Ion:	41	Resp:	83228
Ion	Ratio	Lower	Upper
41	100		
69	71.8	58.5	87.7
39	64.3	51.7	77.5





#49

1,4-Dioxane

Concen: 1035.118 ug/l

RT: 7.653 min Scan# 1078

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 88 Resp: 2849

Ion Ratio Lower Upper

88 100

43 43.2 33.4 50.2





58 75.9 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#50

Toluene-d8

Concen: 49.472 ug/l

RT: 8.641 min Scan# 1239

Delta R.T. -0.006 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 98 Resp: 191933

Ion Ratio Lower Upper

98 100


100 64.7 53.5 80.3





#51
4-Methyl-2-Pentanone
Concen: 280.055 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 527690
Ion Ratio Lower Upper
43 100
58 35.5 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#52
Toluene
Concen: 54.121 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 92 Resp: 146396
Ion Ratio Lower Upper
92 100
91 172.6 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 57.024 ug/l

RT: 8.970 min Scan# 1193

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 75 Resp: 8636

Ion Ratio Lower Upper

75 100





77 31.5 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#54

cis-1,3-Dichloropropene

Concen: 57.712 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 75 Resp: 96608


Ion Ratio Lower Upper

75 100

77 30.1 25.4 38.0

39 59.9 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 54.934 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 97 Resp: 58592

Ion Ratio Lower Upper

97 100

83 87.6 69.0 103.6

85 55.7 44.3 66.5





99 63.2 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#56

Ethyl methacrylate

Concen: 57.126 ug/l

RT: 9.110 min Scan# 1316

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion: 69 Resp: 97109


Ion Ratio Lower Upper

69 100

41 76.9 60.8 91.2

39 48.3 39.0 58.6





#57

1,3-Dichloropropane

Concen: 53.435 ug/l

RT: 9.305 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 76 Resp: 102354

Ion Ratio Lower Upper

76 100





78 32.8 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#58

2-Chloroethyl Vinyl ether

Concen: 275.903 ug/l

RT: 8.232 min Scan# 1172

Delta R.T. -0.006 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 63 Resp: 239108

Ion Ratio Lower Upper

63 100

106 26.7 21.5 32.3





#59

2-Hexanone

Concen: 280.812 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 39146

Ion Ratio Lower Upper

43 100





58 49.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#60

Dibromochloromethane

Concen: 57.902 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion:129 Resp: 67824

Ion Ratio Lower Upper

129 100

127 76.3 39.3 117.8









#61
1,2-Dibromoethane
Concen: 55.163 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:107 Resp: 61152
Ion Ratio Lower Upper
107 100
109 92.8 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#62
4-Bromofluorobenzene
Concen: 50.734 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14


Tgt Ion: 95 Resp: 75501
Ion Ratio Lower Upper
95 100
174 66.6 0.0 135.8
176 64.6 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:117 Resp: 130840
Ion Ratio Lower Upper
117 100
82 61.2 50.6 76.0
119 31.7 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#64
Tetrachloroethene
Concen: 56.142 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14


Tgt Ion:164 Resp: 51973
Ion Ratio Lower Upper
164 100
166 132.6 105.0 157.6
129 107.2 83.5 125.3
131 103.9 76.5 114.7





#65
Chlorobenzene
Concen: 54.044 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 154770
Ion Ratio Lower Upper
112 100
114 32.7 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 56.685 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14


Tgt Ion:131 Resp: 55431
Ion Ratio Lower Upper
131 100
133 95.5 47.3 141.9
119 64.1 31.6 95.0





#67
Ethyl Benzene
Concen: 57.284 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 28917
Ion Ratio Lower Upper
91 100
106 29.1 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#68
m/p-Xylenes
Concen: 113.281 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion:106 Resp: 209150
Ion Ratio Lower Upper
106 100
91 217.8 171.2 256.8









#69
o-Xylene
Concen: 56.842 ug/l
RT: 10.634 min Scan# 1566
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:106 Resp: 102312
Ion Ratio Lower Upper
106 100
91 229.3 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#70
Styrene
Concen: 58.788 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion:104 Resp: 173337
Ion Ratio Lower Upper
104 100
78 58.1 45.7 68.5
103 54.6 43.7 65.5





#71

Bromoform

Concen: 58.367 ug/l

RT: 10.793 min Scan# 1593

Delta R.T. -0.006 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:173 Resp: 42852

Ion Ratio Lower Upper

173 100

175 48.0 23.4 70.0





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14




Tgt Ion:152 Resp: 64536


Ion Ratio Lower Upper

152 100

115 96.2 46.9 140.7

150 174.1 0.0 351.0





#73

Isopropylbenzene

Concen: 54.621 ug/l

RT: 10.957 min Scan# 1619

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 27443

Ion Ratio Lower Upper

105 100





120 25.4 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#74

N-ethyl acetate

Concen: 55.143 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Tgt Ion: 43 Resp: 136906




Ion Ratio Lower Upper


43 100

70 37.4 30.9 46.3

55 23.4 18.7 28.1

61 21.5 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 51.151 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 83 Resp: 90062

Ion Ratio Lower Upper

83 100

131 9.8 5.0 14.9





85 64.1 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#76

1,2,3-Trichloropropane

Concen: 54.163 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 75 Resp: 84137

Ion Ratio Lower Upper

75 100

77 38.8 20.5 61.5





#77

Bromobenzene

Concen: 53.525 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:156 Resp: 62435

Ion Ratio Lower Upper

156 100

77 186.0 92.8 278.3





158 95.4 48.8 146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#78

n-propylbenzene

Concen: 55.901 ug/l

RT: 11.299 min Scan# 1675

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion: 91 Resp: 326575

Ion Ratio Lower Upper

91 100

120 21.7 10.8 32.4





#79

2-Chlorotoluene

Concen: 52.554 ug/l

RT: 11.360 min Scan# 1685

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 198028

Ion Ratio Lower Upper

91 100





126 30.9 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#80

1,3,5-Trimethylbenzene

Concen: 54.855 ug/l

RT: 11.445 min Scan# 1699

Delta R.T. -0.006 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion:105 Resp: 230250

Ion Ratio Lower Upper

105 100


120 46.0 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 56.390 ug/l
RT: 11.012 min Scan# 1628
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion: 75 Resp: 26900
Ion Ratio Lower Upper
75 100
53 123.5 100.5 150.7
89 48.1 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#82
4-Chlorotoluene
Concen: 55.048 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion: 91 Resp: 230030
Ion Ratio Lower Upper
91 100
126 26.9 13.3 39.8





#83

tert-Butylbenzene

Concen: 53.855 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion:119 Resp: 22770

Ion Ratio Lower Upper

119 100

91 65.1 32.9 98.7





134 22.9 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#84

1,2,4-Trimethylbenzene

Concen: 55.511 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion:105 Resp: 235959

Ion Ratio Lower Upper

105 100


120 42.1 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 56.108 ug/l
 RT: 11.884 min Scan# 1771
 Delta R.T. -0.006 min
 Lab File: VX046409.D
 Acq: 30 May 2025 10:14




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050



Tgt Ion:105 Resp: 29127
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 29.1




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/02/2025
 Supervised By :Mahesh Dadoda 06/02/2025



#86
 p-Isopropyltoluene
 Concen: 56.880 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046409.D
 Acq: 30 May 2025 10:14


Tgt Ion:119 Resp: 243737
 Ion Ratio Lower Upper
 119 100
 134 24.8 12.5 37.5
 91 27.1 13.8 41.4





#87
1,3-Dichlorobenzene
Concen: 54.725 ug/l
RT: 11.963 min Scan# 1784
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:146 Resp: 116500
Ion Ratio Lower Upper
146 100
111 44.8 22.1 66.3
148 63.1 32.1 96.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#88
1,4-Dichlorobenzene
Concen: 53.346 ug/l
RT: 12.036 min Scan# 1796
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion:146 Resp: 115980
Ion Ratio Lower Upper
146 100
111 42.6 21.3 63.9
148 63.1 31.9 95.5





#89

n-Butylbenzene

Concen: 59.677 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046409.D

Acq: 30 May 2025 10:14

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 224309

Ion Ratio Lower Upper

91 100

92 53.6 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#90

Hexachloroethane

Concen: 57.220 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046409.D




Acq: 30 May 2025 10:14


Tgt Ion:117 Resp: 43197

Ion Ratio Lower Upper

117 100


201 62.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 53.544 ug/l
RT: 12.329 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 11438
Ion Ratio Lower Upper
146 100
111 46.0 22.4 67.3
148 63.8 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 54.344 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Tgt Ion: 75 Resp: 21197
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 88.4 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 56.393 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:180 Resp: 6919
Ion Ratio Lower Upper
180 100
182 95.0 47.9 143.8
145 34.0 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#94
Hexachlorobutadiene
Concen: 56.649 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14


Tgt Ion:225 Resp: 30357
Ion Ratio Lower Upper
225 100
223 63.8 30.8 92.4
227 64.0 30.9 92.7





#95
Naphthalene
Concen: 54.869 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050



Tgt Ion:128 Resp: 246917
Ion Ratio Lower Upper
128 100
127 12.9 10.4 15.6
129 10.5 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 54.538 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX046409.D
Acq: 30 May 2025 10:14

Tgt Ion:180 Resp: 69047
Ion Ratio Lower Upper
180 100
182 95.1 47.8 143.3
145 37.4 18.1 54.3



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046409.D
 Acq On : 30 May 2025 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 30 23:22:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.765	0.819	-7.1	89	0.00
3 P	Chloromethane	0.742	0.801	-8.0	97	0.00
4 C	Vinyl Chloride	0.691	0.722	-4.5#	96	0.00
5 T	Bromomethane	0.320	0.322	-0.6	93	0.00
6 T	Chloroethane	0.369	0.398	-7.9	99	0.00
7 T	Trichlorofluoromethane	1.021	1.123	-10.0	99	0.00
8 T	Diethyl Ether	0.347	0.359	-3.5	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.704	-11.4	103	0.00
10 T	Methyl Iodide	0.747	0.809	-8.3	94	0.00
11 T	Tert butyl alcohol	0.131	0.132	-0.8	96	0.00
12 CM	1,1-Dichloroethene	0.593	0.644	-8.6#	101	0.00
13 T	Acrolein	0.149	0.160	-7.4	99	0.00
14 T	Allyl chloride	1.133	1.295	-14.3	103	0.00
15 T	Acrylonitrile	0.374	0.418	-11.8	101	0.00
16 T	Acetone	0.374	0.420	-12.3	109	0.00
17 T	Carbon Disulfide	1.406	1.493	-6.2	97	0.00
18 T	Methyl Acetate	0.867	1.173	-35.3#	130	0.00
19 T	Methyl tert-butyl Ether	2.079	2.258	-8.6	98	0.00
20 T	Methylene Chloride	0.716	0.714	0.3	98	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.646	-8.4	100	0.00
22 T	Diisopropyl ether	2.189	2.460	-12.4	102	0.00
23 T	Vinyl Acetate	1.925	2.099	-9.0	97	0.00
24 P	1,1-Dichloroethane	1.219	1.347	-10.5	101	0.00
25 T	2-Butanone	0.543	0.600	-10.5	102	-0.01
26 T	2,2-Dichloropropane	0.954	1.041	-9.1	103	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.778	-8.4	100	0.00
28 T	Bromochloromethane	0.587	0.592	-0.9	96	-0.01
29 T	Tetrahydrofuran	0.340	0.372	-9.4	100	0.00
30 C	Chloroform	1.271	1.387	-9.1#	101	0.00
31 T	Cyclohexane	1.111	1.177	-5.9	98	0.00
32 T	1,1,1-Trichloroethane	1.101	1.190	-8.1	99	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.920	1.3	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.360	0.375	-4.2	98	0.00
36 T	1,1-Dichloropropene	0.484	0.526	-8.7	98	0.00
37 T	Ethyl Acetate	0.598	0.657	-9.9	100	-0.01
38 T	Carbon Tetrachloride	0.544	0.593	-9.0	98	0.00
39 T	Methylcyclohexane	0.623	0.679	-9.0	98	0.00
40 TM	Benzene	1.417	1.556	-9.8	98	0.00
41 T	Methacrylonitrile	0.313	0.385	-23.0	103	0.00
42 TM	1,2-Dichloroethane	0.612	0.666	-8.8	98	0.00
43 T	Isopropyl Acetate	0.912	1.025	-12.4	99	0.00
44 TM	Trichloroethene	0.341	0.373	-9.4	97	0.00
45 C	1,2-Dichloropropane	0.352	0.395	-12.2#	99	0.00
46 T	Dibromomethane	0.278	0.306	-10.1	99	0.00
47 T	Bromodichloromethane	0.547	0.622	-13.7	100	0.00
48 T	Methyl methacrylate	0.466	0.535	-14.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046409.D
 Acq On : 30 May 2025 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 30 23:22:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	93	0.00
50 S	Toluene-d8	1.246	1.233	1.0	93	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.678	-12.1	99	0.00
52 CM	Toluene	0.869	0.940	-8.2#	97	0.00
53 T	t-1,3-Dichloropropene	0.487	0.555	-14.0	97	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.621	-15.4	100	0.00
55 T	1,1,2-Trichloroethane	0.343	0.376	-9.6	98	0.00
56 T	Ethyl methacrylate	0.546	0.624	-14.3	97	0.00
57 T	1,3-Dichloropropane	0.615	0.658	-7.0	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.307	-10.4	93	0.00
59 T	2-Hexanone	0.448	0.503	-12.3	99	0.00
60 T	Dibromochloromethane	0.376	0.436	-16.0	101	0.00
61 T	1,2-Dibromoethane	0.356	0.393	-10.4	98	0.00
62 S	4-Bromofluorobenzene	0.478	0.485	-1.5	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	89	0.00
64 T	Tetrachloroethene	0.354	0.397	-12.1	95	0.00
65 PM	Chlorobenzene	1.094	1.183	-8.1	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.424	-13.4	97	0.00
67 C	Ethyl Benzene	1.929	2.210	-14.6#	98	0.00
68 T	m/p-Xylenes	0.706	0.799	-13.2	97	0.00
69 T	o-Xylene	0.688	0.782	-13.7	96	0.00
70 T	Styrene	1.127	1.325	-17.6	97	0.00
71 P	Bromoform	0.281	0.328	-16.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.893	4.252	-9.2	98	0.00
74 T	N-amyl acetate	1.924	2.121	-10.2	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.396	-2.3	99	0.00
76 T	1,2,3-Trichloropropane	1.204	1.304	-8.3	104	0.00
77 T	Bromobenzene	0.904	0.967	-7.0	99	0.00
78 T	n-propylbenzene	4.526	5.060	-11.8	99	0.00
79 T	2-Chlorotoluene	2.919	3.068	-5.1	97	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.568	-9.7	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.417	-12.7	103	0.00
82 T	4-Chlorotoluene	3.238	3.564	-10.1	99	0.00
83 T	tert-Butylbenzene	3.276	3.528	-7.7	98	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.656	-11.0	99	0.00
85 T	sec-Butylbenzene	4.022	4.513	-12.2	100	0.00
86 T	p-Isopropyltoluene	3.320	3.777	-13.8	101	0.00
87 T	1,3-Dichlorobenzene	1.649	1.805	-9.5	101	0.00
88 T	1,4-Dichlorobenzene	1.684	1.797	-6.7	101	0.00
89 T	n-Butylbenzene	2.912	3.476	-19.4	105	0.00
90 T	Hexachloroethane	0.585	0.669	-14.4	102	0.00
91 T	1,2-Dichlorobenzene	1.655	1.772	-7.1	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.328	-8.6	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	1.072	-12.7	104	0.00
94 T	Hexachlorobutadiene	0.415	0.470	-13.3	105	0.00
95 T	Naphthalene	3.487	3.826	-9.7	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046409.D
Acq On : 30 May 2025 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 30 23:22:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	1.070	-9.1	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046409.D
 Acq On : 30 May 2025 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 30 23:22:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	53.499	-7.0	89	0.00
3 P	Chloromethane	50.000	53.985	-8.0	97	0.00
4 C	Vinyl Chloride	50.000	52.290	-4.6#	96	0.00
5 T	Bromomethane	50.000	50.227	-0.5	93	0.00
6 T	Chloroethane	50.000	53.955	-7.9	99	0.00
7 T	Trichlorofluoromethane	50.000	55.020	-10.0	99	0.00
8 T	Diethyl Ether	50.000	51.685	-3.4	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	55.758	-11.5	103	0.00
10 T	Methyl Iodide	50.000	54.083	-8.2	94	0.00
11 T	Tert butyl alcohol	250.000	251.596	-0.6	96	0.00
12 CM	1,1-Dichloroethene	50.000	54.349	-8.7#	101	0.00
13 T	Acrolein	250.000	269.119	-7.6	99	0.00
14 T	Allyl chloride	50.000	57.140	-14.3	103	0.00
15 T	Acrylonitrile	250.000	279.087	-11.6	101	0.00
16 T	Acetone	250.000	281.238	-12.5	109	0.00
17 T	Carbon Disulfide	50.000	53.127	-6.3	97	0.00
18 T	Methyl Acetate	50.000	67.621	-35.2#	130	0.00
19 T	Methyl tert-butyl Ether	50.000	54.313	-8.6	98	0.00
20 T	Methylene Chloride	50.000	49.822	0.4	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	54.176	-8.4	100	0.00
22 T	Diisopropyl ether	50.000	56.192	-12.4	102	0.00
23 T	Vinyl Acetate	250.000	272.572	-9.0	97	0.00
24 P	1,1-Dichloroethane	50.000	55.257	-10.5	101	0.00
25 T	2-Butanone	250.000	276.566	-10.6	102	-0.01
26 T	2,2-Dichloropropane	50.000	54.540	-9.1	103	0.00
27 T	cis-1,2-Dichloroethene	50.000	54.202	-8.4	100	0.00
28 T	Bromochloromethane	50.000	50.407	-0.8	96	-0.01
29 T	Tetrahydrofuran	250.000	273.726	-9.5	100	0.00
30 C	Chloroform	50.000	54.560	-9.1#	101	0.00
31 T	Cyclohexane	50.000	52.962	-5.9	98	0.00
32 T	1,1,1-Trichloroethane	50.000	54.032	-8.1	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.342	1.3	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	93	0.00
35 S	Dibromofluoromethane	50.000	52.055	-4.1	98	0.00
36 T	1,1-Dichloropropene	50.000	54.361	-8.7	98	0.00
37 T	Ethyl Acetate	50.000	54.982	-10.0	100	-0.01
38 T	Carbon Tetrachloride	50.000	54.528	-9.1	98	0.00
39 T	Methylcyclohexane	50.000	54.488	-9.0	98	0.00
40 TM	Benzene	50.000	54.919	-9.8	98	0.00
41 T	Methacrylonitrile	50.000	61.507	-23.0	103	0.00
42 TM	1,2-Dichloroethane	50.000	54.429	-8.9	98	0.00
43 T	Isopropyl Acetate	50.000	56.197	-12.4	99	0.00
44 TM	Trichloroethene	50.000	54.691	-9.4	97	0.00
45 C	1,2-Dichloropropane	50.000	56.068	-12.1#	99	0.00
46 T	Dibromomethane	50.000	54.974	-9.9	99	0.00
47 T	Bromodichloromethane	50.000	56.841	-13.7	100	0.00
48 T	Methyl methacrylate	50.000	57.410	-14.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046409.D
 Acq On : 30 May 2025 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 30 23:22:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1035.118	-3.5	93	0.00
50 S	Toluene-d8	50.000	49.472	1.1	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	280.055	-12.0	99	0.00
52 CM	Toluene	50.000	54.121	-8.2#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	57.024	-14.0	97	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.712	-15.4	100	0.00
55 T	1,1,2-Trichloroethane	50.000	54.934	-9.9	98	0.00
56 T	Ethyl methacrylate	50.000	57.126	-14.3	97	0.00
57 T	1,3-Dichloropropane	50.000	53.435	-6.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	275.903	-10.4	93	0.00
59 T	2-Hexanone	250.000	280.812	-12.3	99	0.00
60 T	Dibromochloromethane	50.000	57.902	-15.8	101	0.00
61 T	1,2-Dibromoethane	50.000	55.163	-10.3	98	0.00
62 S	4-Bromofluorobenzene	50.000	50.734	-1.5	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	89	0.00
64 T	Tetrachloroethene	50.000	56.142	-12.3	95	0.00
65 PM	Chlorobenzene	50.000	54.044	-8.1	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	56.685	-13.4	97	0.00
67 C	Ethyl Benzene	50.000	57.284	-14.6#	98	0.00
68 T	m/p-Xylenes	100.000	113.281	-13.3	97	0.00
69 T	o-Xylene	50.000	56.842	-13.7	96	0.00
70 T	Styrene	50.000	58.788	-17.6	97	0.00
71 P	Bromoform	50.000	58.367	-16.7	96	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	95	0.00
73 T	Isopropylbenzene	50.000	54.621	-9.2	98	0.00
74 T	N-amyl acetate	50.000	55.143	-10.3	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	51.151	-2.3	99	0.00
76 T	1,2,3-Trichloropropane	50.000	54.163	-8.3	104	0.00
77 T	Bromobenzene	50.000	53.525	-7.0	99	0.00
78 T	n-propylbenzene	50.000	55.901	-11.8	99	0.00
79 T	2-Chlorotoluene	50.000	52.554	-5.1	97	0.00
80 T	1,3,5-Trimethylbenzene	50.000	54.855	-9.7	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	56.390	-12.8	103	0.00
82 T	4-Chlorotoluene	50.000	55.048	-10.1	99	0.00
83 T	tert-Butylbenzene	50.000	53.855	-7.7	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	55.511	-11.0	99	0.00
85 T	sec-Butylbenzene	50.000	56.108	-12.2	100	0.00
86 T	p-Isopropyltoluene	50.000	56.880	-13.8	101	0.00
87 T	1,3-Dichlorobenzene	50.000	54.725	-9.5	101	0.00
88 T	1,4-Dichlorobenzene	50.000	53.346	-6.7	101	0.00
89 T	n-Butylbenzene	50.000	59.677	-19.4	105	0.00
90 T	Hexachloroethane	50.000	57.220	-14.4	102	0.00
91 T	1,2-Dichlorobenzene	50.000	53.544	-7.1	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.344	-8.7	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	56.393	-12.8	104	0.00
94 T	Hexachlorobutadiene	50.000	56.649	-13.3	105	0.00
95 T	Naphthalene	50.000	54.869	-9.7	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046409.D
Acq On : 30 May 2025 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 30 23:22:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	54.538	-9.1	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046038.D
Acq On : 05 May 2025 09:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	8676	PASS
75	95	30	60	56.2	22052	PASS
95	95	100	100	100.0	39213	PASS
96	95	5	9	6.4	2505	PASS
173	174	0.00	2	0.7	185	PASS
174	95	50	100	68.8	26963	PASS
175	174	5	9	7.3	1980	PASS
176	174	95	101	97.0	26147	PASS
177	176	5	9	6.9	1802	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046408.D
Acq On : 30 May 2025 08:20
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.6	7633	PASS
75	95	30	60	57.6	19408	PASS
95	95	100	100	100.0	33717	PASS
96	95	5	9	6.2	2087	PASS
173	174	0.00	2	1.3	293	PASS
174	95	50	100	68.3	23021	PASS
175	174	5	9	7.7	1769	PASS
176	174	95	101	95.2	21917	PASS
177	176	5	9	6.5	1422	PASS

Report of Analysis

Client:	Aramark Uniforms			Date Collected:	
Project:	Waste Characterization			Date Received:	
Client Sample ID:	VX0530WBL01			SDG No.:	Q2149
Lab Sample ID:	VX0530WBL01			Matrix:	TCLP
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	TCLP VOA Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046411.D	1		05/30/25 11:06	VX053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.0		74 - 125	108%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	50.1		86 - 113	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.8		77 - 121	102%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	64500	5.55			
540-36-3	1,4-Difluorobenzene	130000	6.757			
3114-55-4	Chlorobenzene-d5	123000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	51200	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046411.D
 Acq On : 30 May 2025 11:06
 Operator : JC/MD
 Sample : VX0530WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBL01

Quant Time: May 30 23:23:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64533	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129905	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	123011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51208	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	64990	54.019	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.040%
35) Dibromofluoromethane	5.379	113	46982	50.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.440%
50) Toluene-d8	8.647	98	162314	50.132	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.260%
62) 4-Bromofluorobenzene	11.079	95	63110	50.815	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.640%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046411.D
Acq On : 30 May 2025 11:06
Operator : JC/MD
Sample : VX0530WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBL01

Quant Time: May 30 23:23:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 711
Delta R.T. 0.006 min
Lab File: VX046411.D
Acq: 30 May 2025 11:06




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBL01


Tgt Ion:168 Resp: 64533
Ion Ratio Lower Upper
168 100
99 67.1 54.9 82.3



#33
1,2-Dichloroethane-d4
Concen: 54.019 ug/l
RT: 5.946 min Scan# 797
Delta R.T. -0.006 min
Lab File: VX046411.D
Acq: 30 May 2025 11:06

Tgt Ion: 65 Resp: 64990
Ion Ratio Lower Upper
65 100
67 48.5 0.0 99.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 91

Delta R.T. 0.000 min

Lab File: VX046411.D

Acq: 30 May 2025 11:06

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBL01





Tgt Ion:114 Resp: 129905

Ion Ratio Lower Upper

114 100

63 23.8 0.0 49.2

88 17.1 0.0 33.6



#35

Dibromofluoromethane

Concen: 50.224 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046411.D

Acq: 30 May 2025 11:06




Tgt Ion:113 Resp: 46982


Ion Ratio Lower Upper

113 100

111 103.8 83.1 124.7

192 17.1 13.3 19.9





#50

Toluene-d8

Concen: 50.132 ug/l

RT: 8.647 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046411.D

Acq: 30 May 2025 11:06

Instrument :

MSVOA_X

ClientSampleId :





VX0530WBL01

Tgt Ion: 98 Resp: 162314

Ion Ratio Lower Upper

98 100

100 65.7 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 50.815 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX046411.D

Acq: 30 May 2025 11:06




Tgt Ion: 95 Resp: 63110


Ion Ratio Lower Upper

95 100

174 66.6 0.0 135.8

176 62.3 0.0 131.4





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. 0.000 min

Lab File: VX046411.D

Acq: 30 May 2025 11:06

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBL01





Tgt Ion:117 Resp: 123011

Ion Ratio Lower Upper

117 100

82 62.6 50.6 76.0

119 31.1 25.8 38.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046411.D

Acq: 30 May 2025 11:06




Tgt Ion:152 Resp: 51208

Ion Ratio Lower Upper

152 100

115 66.9 46.9 140.7

150 157.0 0.0 351.0



Report of Analysis

Client:	Aramark Uniforms	Date Collected:	
Project:	Waste Characterization	Date Received:	
Client Sample ID:	VX0530WBS01	SDG No.:	Q2149
Lab Sample ID:	VX0530WBS01	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046412.D	1		05/30/25 11:29	VX053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	16.5		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.9		0.23	1.00	ug/L
78-93-3	2-Butanone	110		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.7		0.25	1.00	ug/L
67-66-3	Chloroform	20.5		0.25	1.00	ug/L
71-43-2	Benzene	19.3		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.9		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.8		0.090	1.00	ug/L
127-18-4	Tetrachloroethene	18.4		0.23	1.00	ug/L
108-90-7	Chlorobenzene	19.1		0.12	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.9		74 - 125	100%	SPK: 50
1868-53-7	Dibromofluoromethane	50.4		75 - 124	101%	SPK: 50
2037-26-5	Toluene-d8	47.8		86 - 113	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		77 - 121	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	86600	5.549			
540-36-3	1,4-Difluorobenzene	153000	6.756			
3114-55-4	Chlorobenzene-d5	135000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	64100	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046412.D
 Acq On : 30 May 2025 11:29
 Operator : JC/MD
 Sample : VX0530WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/02/2025
 Supervised By : Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:23:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	86619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	153415	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	134836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	64120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80579	49.899	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.800%		
35) Dibromofluoromethane	5.379	113	55654	50.377	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.760%		
50) Toluene-d8	8.646	98	182778	47.801	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	95.600%		
62) 4-Bromofluorobenzene	11.079	95	72518	49.443	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.880%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	22046	16.629	ug/l	94
3) Chloromethane	1.306	50	21447	16.682	ug/l	94
4) Vinyl Chloride	1.373	62	19698	16.462	ug/l	97
5) Bromomethane	1.599	94	9497	17.112	ug/l	98
6) Chloroethane	1.672	64	11661	18.255	ug/l	95
7) Trichlorofluoromethane	1.879	101	32574	18.420	ug/l	99
8) Diethyl Ether	2.135	74	10810	17.957	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.318	101	20795	19.002	ug/l	98
10) Methyl Iodide	2.446	142	22001	16.990	ug/l	100
11) Tert butyl alcohol	2.971	59	24227	106.879	ug/l	99
12) 1,1-Dichloroethene	2.312	96	18348	17.864	ug/l	98
13) Acrolein	2.233	56	16928	65.574	ug/l	98
14) Allyl chloride	2.660	41	38498	19.612	ug/l	95
15) Acrylonitrile	3.062	53	68823	106.181	ug/l	98
16) Acetone	2.379	43	68128	105.220	ug/l	100
17) Carbon Disulfide	2.507	76	34583	14.202	ug/l	96
18) Methyl Acetate	2.702	43	39508	26.295	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	71509	19.859	ug/l	98
20) Methylene Chloride	2.788	84	22533	18.160	ug/l	93
21) trans-1,2-Dichloroethene	3.086	96	18558	17.967	ug/l	98
22) Diisopropyl ether	3.763	45	77266	20.378	ug/l	93
23) Vinyl Acetate	3.720	43	320325	96.051	ug/l	99
24) 1,1-Dichloroethane	3.605	63	41445	19.625	ug/l	99
25) 2-Butanone	4.556	43	102333	108.864	ug/l	100
26) 2,2-Dichloropropane	4.464	77	30832	18.652	ug/l	99
27) cis-1,2-Dichloroethene	4.482	96	23968	19.276	ug/l	99
28) Bromochloromethane	4.897	49	21089	20.745	ug/l	98
29) Tetrahydrofuran	5.001	42	62399	105.935	ug/l	99
30) Chloroform	5.086	83	45096	20.486	ug/l	93
31) Cyclohexane	5.464	56	33070	17.184	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	37073	19.429	ug/l	98
36) 1,1-Dichloropropene	5.684	75	26188	17.643	ug/l	98
37) Ethyl Acetate	4.720	43	35095	19.137	ug/l	99
38) Carbon Tetrachloride	5.671	117	31244	18.734	ug/l	97
39) Methylcyclohexane	7.378	83	32348	16.928	ug/l	98
40) Benzene	6.031	78	84048	19.331	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046412.D
 Acq On : 30 May 2025 11:29
 Operator : JC/MD
 Sample : VX0530WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/02/2025
 Supervised By : Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:23:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	20721	21.599	ug/l	98
42) 1,2-Dichloroethane	6.086	62	37398	19.930	ug/l	100
43) Isopropyl Acetate	6.342	43	56552	20.214	ug/l	98
44) Trichloroethene	7.122	130	19629	18.758	ug/l	100
45) 1,2-Dichloropropane	7.427	63	22287	20.615	ug/l	97
46) Dibromomethane	7.580	93	16946	19.874	ug/l	98
47) Bromodichloromethane	7.817	83	33719	20.078	ug/l	97
48) Methyl methacrylate	7.695	41	28679	20.072	ug/l	97
49) 1,4-Dioxane	7.659	88	12612	464.881	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	200919	108.190	ug/l	99
52) Toluene	8.713	92	51245	19.222	ug/l	96
53) t-1,3-Dichloropropene	8.976	75	27848	18.656	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	31975	19.381	ug/l	95
55) 1,1,2-Trichloroethane	9.146	97	21705	20.648	ug/l	94
56) Ethyl methacrylate	9.116	69	34157	20.387	ug/l	99
57) 1,3-Dichloropropane	9.305	76	37433	19.828	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	94881	111.083	ug/l	98
59) 2-Hexanone	9.427	43	150933	109.854	ug/l	99
60) Dibromochloromethane	9.518	129	23960	20.754	ug/l	97
61) 1,2-Dibromoethane	9.604	107	21935	20.076	ug/l	99
64) Tetrachloroethene	9.268	164	17515	18.359	ug/l	95
65) Chlorobenzene	10.073	112	56306	19.079	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.158	131	20489	20.331	ug/l	99
67) Ethyl Benzene	10.189	91	101092	19.432	ug/l	100
68) m/p-Xylenes	10.299	106	72256	37.976	ug/l	97
69) o-Xylene	10.640	106	36463	19.657	ug/l	97
70) Styrene	10.652	104	60271	19.835	ug/l	97
71) Bromoform	10.798	173	14805	19.568	ug/l #	98
73) Isopropylbenzene	10.957	105	98609	19.754	ug/l	99
74) N-amyl acetate	10.841	43	48523	19.671	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	34441	19.688	ug/l	97
76) 1,2,3-Trichloropropane	11.237	75	31277m	20.265	ug/l	
77) Bromobenzene	11.195	156	22245	19.194	ug/l	98
78) n-propylbenzene	11.298	91	111193	19.157	ug/l	100
79) 2-Chlorotoluene	11.359	91	71240	19.029	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	82619	19.811	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8673	18.295	ug/l	98
82) 4-Chlorotoluene	11.451	91	79525	19.154	ug/l	98
83) tert-Butylbenzene	11.713	119	81221	19.335	ug/l	99
84) 1,2,4-Trimethylbenzene	11.749	105	82870	19.622	ug/l	99
85) sec-Butylbenzene	11.890	105	100632	19.510	ug/l	100
86) p-Isopropyltoluene	12.006	119	82888	19.469	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	40311	19.059	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	40286	18.650	ug/l	97
89) n-Butylbenzene	12.329	91	70173	18.790	ug/l	99
90) Hexachloroethane	12.536	117	14438	19.249	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	42227	19.895	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.944	75	7914	20.421	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	22315	18.305	ug/l	98
94) Hexachlorobutadiene	13.725	225	9922	18.636	ug/l	96
95) Naphthalene	13.774	128	83733	18.727	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	23105	18.368	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046412.D
Acq On : 30 May 2025 11:29
Operator : JC/MD
Sample : VX0530WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:23:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046412.D
Acq On : 30 May 2025 11:29
Operator : JC/MD
Sample : VX0530WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Manual Integrations

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:23:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.549 min Scan# 71
Delta R.T. 0.005 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion:168 Resp: 86619
Ion Ratio Lower Upper
168 100
99 65.7 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#2
Dichlorodifluoromethane
Concen: 16.629 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 85 Resp: 22046
Ion Ratio Lower Upper
85 100
87 34.7 15.7 47.1









#3
Chloromethane
Concen: 16.682 ug/l
RT: 1.306 min Scan# 30
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 50 Resp: 2144
Ion Ratio Lower Upper
50 100
52 35.1 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#4
Vinyl Chloride
Concen: 16.462 ug/l
RT: 1.373 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29


Tgt Ion: 62 Resp: 19698
Ion Ratio Lower Upper
62 100
64 33.0 25.2 37.8





#5
Bromomethane
Concen: 17.112 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion: 94 Resp: 949
Ion Ratio Lower Upper
94 100
96 93.3 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#6
Chloroethane
Concen: 18.255 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29


Tgt Ion: 64 Resp: 11661
Ion Ratio Lower Upper
64 100
66 33.2 24.3 36.5





#7
Trichlorofluoromethane
Concen: 18.420 ug/l
RT: 1.879 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion:101 Resp: 32574
Ion Ratio Lower Upper
101 100
103 64.8 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#8
Diethyl Ether
Concen: 17.957 ug/l
RT: 2.135 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 74 Resp: 10810
Ion Ratio Lower Upper
74 100
45 107.4 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.002 ug/l
RT: 2.318 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion:101 Resp: 20795
Ion Ratio Lower Upper
101 100
85 46.4 38.6 58.0
151 68.0 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#10
Methyl Iodide
Concen: 16.990 ug/l
RT: 2.446 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion:142 Resp: 22001
Ion Ratio Lower Upper
142 100
127 52.2 41.7 62.5
141 13.8 11.5 17.3





#11

Tert butyl alcohol

Concen: 106.879 ug/l

RT: 2.971 min Scan# 30

Delta R.T. -0.007 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 59 Resp: 2422

Ion Ratio Lower Upper

59 100





57 11.1 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#12

1,1-Dichloroethene

Concen: 17.864 ug/l

RT: 2.312 min Scan# 201

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion: 96 Resp: 18348


Ion Ratio Lower Upper

96 100

61 179.6 146.2 219.2

98 62.4 51.0 76.6





#13

Acrolein

Concen: 65.574 ug/l

RT: 2.233 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion:	56	Resp:	16928
Ion Ratio	Lower	Upper	
56	100		
55	72.3	56.2	84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#14

Allyl chloride

Concen: 19.612 ug/l




RT: 2.660 min Scan# 258


Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion:	41	Resp:	38498
Ion Ratio	Lower	Upper	
41	100		
39	72.3	60.6	90.8
76	27.7	24.9	37.3





#15

Acrylonitrile

Concen: 106.181 ug/l

RT: 3.062 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion: 53	Resp: 6882
Ion Ratio	Lower Upper
53 100	
52 83.0	65.3 97.9
51 39.1	29.8 44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#16

Acetone

Concen: 105.220 ug/l




RT: 2.379 min Scan# 212


Delta R.T. -0.007 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion: 43	Resp: 68128
Ion Ratio	Lower Upper
43 100	
58 26.3	21.2 31.8









#17
Carbon Disulfide
Concen: 14.202 ug/l
RT: 2.507 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 76 Resp: 3458
Ion Ratio Lower Upper
76 100
78 10.0 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#18
Methyl Acetate
Concen: 26.295 ug/l
RT: 2.702 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 43 Resp: 39508
Ion Ratio Lower Upper
43 100
74 20.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 19.859 ug/l
RT: 3.117 min Scan# 311
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 73 Resp: 71509
Ion Ratio Lower Upper
73 100
57 21.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#20
Methylene Chloride
Concen: 18.160 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 84 Resp: 22533
Ion Ratio Lower Upper
84 100
49 131.2 113.9 170.9
51 39.3 33.5 50.3
86 64.2 53.8 80.8





#21

trans-1,2-Dichloroethene

Concen: 17.967 ug/l

RT: 3.086 min Scan# 31

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 96 Resp: 18558

Ion Ratio Lower Upper

96 100

61 150.8 119.5 179.3





98 65.5 50.0 75.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#22

Diisopropyl ether

Concen: 20.378 ug/l

RT: 3.763 min Scan# 439

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion: 45 Resp: 77266




Ion Ratio Lower Upper


45 100

43 75.0 66.6 100.0

87 23.0 19.8 29.6

59 9.9 8.6 12.8









#23
Vinyl Acetate
Concen: 96.051 ug/l
RT: 3.720 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 43 Resp: 320325
Ion Ratio Lower Upper
43 100
86 9.7 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#24
1,1-Dichloroethane
Concen: 19.625 ug/l
RT: 3.605 min Scan# 413
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 63 Resp: 41445
Ion Ratio Lower Upper
63 100
98 6.5 3.6 10.8
100 4.4 2.1 6.3





#25

2-Butanone

Concen: 108.864 ug/l

RT: 4.556 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 43 Resp: 10233

Ion Ratio Lower Upper

43 100





72 23.0 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#26

2,2-Dichloropropane

Concen: 18.652 ug/l

RT: 4.464 min Scan# 554

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion: 77 Resp: 30832

Ion Ratio Lower Upper

77 100

97 20.8 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 19.276 ug/l

RT: 4.482 min Scan# 557

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 96 Resp: 2396

Ion Ratio Lower Upper

96 100

61 160.0 0.0 322.8





98 66.3 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#28

Bromochloromethane

Concen: 20.745 ug/l

RT: 4.897 min Scan# 625

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion: 49 Resp: 21089


Ion Ratio Lower Upper

49 100

129 0.6 0.0 4.0

130 68.8 56.2 84.2





#29

Tetrahydrofuran

Concen: 105.935 ug/l

RT: 5.001 min Scan# 642

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 42 Resp: 62399

Ion Ratio Lower Upper

42 100

72 41.5 33.7 50.5





71 38.2 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#30

Chloroform

Concen: 20.486 ug/l

RT: 5.086 min Scan# 656

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion: 83 Resp: 45096

Ion Ratio Lower Upper

83 100

85 67.3 49.3 73.9





#31

Cyclohexane

Concen: 17.184 ug/l

RT: 5.464 min Scan# 718

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 56 Resp: 33070

Ion Ratio Lower Upper

56 100

69 28.6 24.4 36.6





84 82.7 66.9 100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#32

1,1,1-Trichloroethane

Concen: 19.429 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion: 97 Resp: 37073


Ion Ratio Lower Upper

97 100

99 63.1 51.8 77.6

61 51.6 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 49.899 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 65 Resp: 80579

Ion Ratio Lower Upper

65 100





67 49.3 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.756 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion:114 Resp: 153415


Ion Ratio Lower Upper

114 100

63 24.5 0.0 49.2

88 16.7 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.377 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion:	113	Resp:	5565
Ion	Ratio	Lower	Upper
113	100		
111	102.0	83.1	124.7
192	15.8	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#36

1,1-Dichloropropene

Concen: 17.643 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion:	75	Resp:	26188
Ion	Ratio	Lower	Upper
75	100		
110	34.3	16.3	48.9
77	31.2	24.3	36.5





#37

Ethyl Acetate

Concen: 19.137 ug/l

RT: 4.720 min Scan# 595

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 43 Resp: 3509





Ion	Ratio	Lower	Upper
43	100		
61	12.6	10.3	15.5
70	9.5	7.9	11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#38

Carbon Tetrachloride

Concen: 18.734 ug/l

RT: 5.671 min Scan# 752




Delta R.T. -0.000 min


Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion:117 Resp: 31244

Ion	Ratio	Lower	Upper
117	100		
119	90.7	75.2	112.8
121	29.6	24.2	36.4





#39

Methylcyclohexane

Concen: 16.928 ug/l

RT: 7.378 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 83 Resp: 3234

Ion Ratio Lower Upper

83 100

55 81.6 64.7 97.1





98 49.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#40

Benzene

Concen: 19.331 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion: 78 Resp: 84048

Ion Ratio Lower Upper

78 100


77 22.2 19.0 28.4





#41
Methacrylonitrile
Concen: 21.599 ug/l
RT: 4.915 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion: 41 Resp: 2072
Ion Ratio Lower Upper
41 100
39 58.3 47.2 70.8
67 65.5 50.7 76.1
52 31.4 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#42
1,2-Dichloroethane
Concen: 19.930 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 62 Resp: 37398
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43

Isopropyl Acetate

Concen: 20.214 ug/l

RT: 6.342 min Scan# 861

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion: 43	Resp: 5655
Ion Ratio	Lower Upper
43	100
61	17.0 14.3 21.5
87	11.1 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#44

Trichloroethene

Concen: 18.758 ug/l




RT: 7.122 min Scan# 990


Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion:130	Resp: 19629
Ion Ratio	Lower Upper
130	100
95	102.0 0.0 204.2





#45

1,2-Dichloropropane

Concen: 20.615 ug/l

RT: 7.427 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 63 Resp: 2228

Ion Ratio Lower Upper

63 100





65 32.0 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#46

Dibromomethane

Concen: 19.874 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion: 93 Resp: 16946


Ion Ratio Lower Upper

93 100

95 81.4 65.6 98.4

174 82.4 68.2 102.2





#47

Bromodichloromethane

Concen: 20.078 ug/l

RT: 7.817 min Scan# 1104

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion:	83	Resp:	33719
Ion	Ratio	Lower	Upper
83	100		
85	65.5	50.5	75.7
127	8.5	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#48

Methyl methacrylate

Concen: 20.072 ug/l




RT: 7.695 min Scan# 1084


Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion:	41	Resp:	28679
Ion	Ratio	Lower	Upper
41	100		
69	75.2	58.5	87.7
39	66.6	51.7	77.5





#49

1,4-Dioxane

Concen: 464.881 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 88 Resp: 12612

Ion Ratio Lower Upper

88 100

43 41.2 33.4 50.2





58 73.1 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#50

Toluene-d8

Concen: 47.801 ug/l

RT: 8.646 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion: 98 Resp: 182778

Ion Ratio Lower Upper

98 100

100 64.8 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 108.190 ug/l
RT: 8.573 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 43 Resp: 200919
Ion Ratio Lower Upper
43 100
58 35.2 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#52
Toluene
Concen: 19.222 ug/l
RT: 8.713 min Scan# 1251
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 92 Resp: 51245
Ion Ratio Lower Upper
92 100
91 175.7 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 18.656 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 75 Resp: 2784

Ion Ratio Lower Upper

75 100





77 33.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#54

cis-1,3-Dichloropropene

Concen: 19.381 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion: 75 Resp: 31975


Ion Ratio Lower Upper

75 100

77 33.8 25.4 38.0

39 60.9 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 20.648 ug/l

RT: 9.146 min Scan# 1322

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 97 Resp: 2170

Ion Ratio Lower Upper

97 100

83 88.8 69.0 103.6

85 62.4 44.3 66.5





99 65.2 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#56

Ethyl methacrylate

Concen: 20.387 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion: 69 Resp: 34157


Ion Ratio Lower Upper

69 100

41 76.6 60.8 91.2

39 48.1 39.0 58.6





#57

1,3-Dichloropropane

Concen: 19.828 ug/l

RT: 9.305 min Scan# 1348

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 76 Resp: 3743

Ion Ratio Lower Upper

76 100





78 32.4 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#58

2-Chloroethyl Vinyl ether

Concen: 111.083 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion: 63 Resp: 94881

Ion Ratio Lower Upper

63 100

106 26.1 21.5 32.3





#59

2-Hexanone

Concen: 109.854 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 43 Resp: 15093

Ion Ratio Lower Upper

43 100





58 48.8 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#60

Dibromochloromethane

Concen: 20.754 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion:129 Resp: 23960

Ion Ratio Lower Upper

129 100

127 76.2 39.3 117.8









#61
1,2-Dibromoethane
Concen: 20.076 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion:107 Resp: 21938
Ion Ratio Lower Upper
107 100
109 92.4 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#62
4-Bromofluorobenzene
Concen: 49.443 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29


Tgt Ion: 95 Resp: 72518
Ion Ratio Lower Upper
95 100
174 66.3 0.0 135.8
176 64.5 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion:117 Resp: 134830
Ion Ratio Lower Upper
117 100
82 62.0 50.6 76.0
119 31.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#64
Tetrachloroethene
Concen: 18.359 ug/l
RT: 9.268 min Scan# 1342
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion:164 Resp: 17515
Ion Ratio Lower Upper
164 100
166 123.8 105.0 157.6
129 108.1 83.5 125.3
131 99.1 76.5 114.7









#65
Chlorobenzene
Concen: 19.079 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion:112 Resp: 56300
Ion Ratio Lower Upper
112 100
114 31.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 20.331 ug/l
RT: 10.158 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion:131 Resp: 20489
Ion Ratio Lower Upper
131 100
133 94.0 47.3 141.9
119 62.2 31.6 95.0









#67
Ethyl Benzene
Concen: 19.432 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 91 Resp: 101092
Ion Ratio Lower Upper
91 100
106 29.4 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#68
m/p-Xylenes
Concen: 37.976 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion:106 Resp: 72256
Ion Ratio Lower Upper
106 100
91 219.0 171.2 256.8









#69
o-Xylene
Concen: 19.657 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion:106 Resp: 3646
Ion Ratio Lower Upper
106 100
91 230.2 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#70
Styrene
Concen: 19.835 ug/l
RT: 10.652 min Scan# 1569
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29


Tgt Ion:104 Resp: 60271
Ion Ratio Lower Upper
104 100
78 59.1 45.7 68.5
103 56.7 43.7 65.5





#71
Bromoform
Concen: 19.568 ug/l
RT: 10.798 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion:173 Resp: 1480
Ion Ratio Lower Upper
173 100
175 48.2 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion:152 Resp: 64120
Ion Ratio Lower Upper
152 100
115 73.8 46.9 140.7
150 160.6 0.0 351.0





#73

Isopropylbenzene

Concen: 19.754 ug/l

RT: 10.957 min Scan# 1619

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion:105 Resp: 98609

Ion Ratio Lower Upper

105 100





120 25.3 12.8 38.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#74

N-ethyl acetate

Concen: 19.671 ug/l

RT: 10.841 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Tgt Ion: 43 Resp: 48523




Ion Ratio Lower Upper


43 100

70 37.0 30.9 46.3

55 23.9 18.7 28.1

61 22.1 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 19.688 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01

Tgt Ion: 83 Resp: 3444
Ion Ratio Lower Upper
83 100
131 10.3 5.0 14.9
85 66.6 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#76
1,2,3-Trichloropropane
Concen: 20.265 ug/l m
RT: 11.237 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29


Tgt Ion: 75 Resp: 31277
Ion Ratio Lower Upper
75 100
77 39.7 20.5 61.5





#77
Bromobenzene
Concen: 19.194 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion:156 Resp: 22249
Ion Ratio Lower Upper
156 100
77 183.2 92.8 278.3
158 95.0 48.8 146.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#78
n-propylbenzene
Concen: 19.157 ug/l
RT: 11.298 min Scan# 1675
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 91 Resp: 111193
Ion Ratio Lower Upper
91 100
120 21.5 10.8 32.4





#79

2-Chlorotoluene

Concen: 19.029 ug/l

RT: 11.359 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 91 Resp: 71240

Ion Ratio Lower Upper

91 100





126 31.1 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#80

1,3,5-Trimethylbenzene

Concen: 19.811 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion:105 Resp: 82619

Ion Ratio Lower Upper

105 100


120 45.8 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 18.295 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion: 75 Resp: 867
Ion Ratio Lower Upper
75 100
53 128.2 100.5 150.7
89 45.5 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#82
4-Chlorotoluene
Concen: 19.154 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 91 Resp: 79525
Ion Ratio Lower Upper
91 100
126 27.8 13.3 39.8





#83

tert-Butylbenzene

Concen: 19.335 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion:	119	Resp:	8122
Ion Ratio	Lower	Upper	
119	100		
91	66.9	32.9	98.7
134	23.5	11.4	34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#84

1,2,4-Trimethylbenzene

Concen: 19.622 ug/l




RT: 11.749 min Scan# 1749


Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29


Tgt Ion:	105	Resp:	82870
Ion Ratio	Lower	Upper	
105	100		
120	43.2	21.2	63.6





#85
 sec-Butylbenzene
 Concen: 19.510 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX046412.D
 Acq: 30 May 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBS01




Tgt Ion:105 Resp: 100632
 Ion Ratio Lower Upper
 105 100
 134 19.3 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 06/02/2025
 Supervised By :Mahesh Dadoda 06/02/2025




#86
 p-Isopropyltoluene
 Concen: 19.469 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX046412.D
 Acq: 30 May 2025 11:29



Tgt Ion:119 Resp: 82888
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 27.9 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 19.059 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion:146 Resp: 4031

Ion Ratio Lower Upper

146 100

111 45.5 22.1 66.3





148 64.6 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#88

1,4-Dichlorobenzene

Concen: 18.650 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29




Tgt Ion:146 Resp: 40286


Ion Ratio Lower Upper

146 100

111 45.3 21.3 63.9

148 64.8 31.9 95.5





#89

n-Butylbenzene

Concen: 18.790 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01

Tgt Ion: 91 Resp: 7017

Ion Ratio Lower Upper

91 100

92 52.5 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#90

Hexachloroethane

Concen: 19.249 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX046412.D




Acq: 30 May 2025 11:29


Tgt Ion:117 Resp: 14438

Ion Ratio Lower Upper

117 100


201 62.3 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 19.895 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion:146 Resp: 4222
Ion Ratio Lower Upper
146 100
111 46.1 22.4 67.3
148 64.9 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.421 ug/l
RT: 12.944 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion: 75 Resp: 7914
Ion Ratio Lower Upper
75 100
155 65.4 34.9 104.8
157 90.2 43.8 131.4





#93

1,2,4-Trichlorobenzene

Concen: 18.305 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBS01





Tgt Ion:	180	Resp:	2231
Ion Ratio	Lower	Upper	
180	100		
182	93.2	47.9	143.8
145	33.7	17.1	51.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#94

Hexachlorobutadiene

Concen: 18.636 ug/l




RT: 13.725 min Scan# 2073


Delta R.T. -0.000 min

Lab File: VX046412.D

Acq: 30 May 2025 11:29


Tgt Ion:	225	Resp:	9922
Ion Ratio	Lower	Upper	
225	100		
223	67.5	30.8	92.4
227	62.7	30.9	92.7





#95
Naphthalene
Concen: 18.727 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBS01



Tgt Ion:128 Resp: 83733
Ion Ratio Lower Upper
128 100
127 13.4 10.4 15.6
129 10.9 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 18.368 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.006 min
Lab File: VX046412.D
Acq: 30 May 2025 11:29

Tgt Ion:180 Resp: 23105
Ion Ratio Lower Upper
180 100
182 94.7 47.8 143.3
145 38.4 18.1 54.3



Report of Analysis

Client:	Aramark Uniforms	Date Collected:	
Project:	Waste Characterization	Date Received:	
Client Sample ID:	VX0530WBSD01	SDG No.:	Q2149
Lab Sample ID:	VX0530WBSD01	Matrix:	TCLP
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	TCLP VOA Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046413.D	1		05/30/25 11:55	VX053025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-01-4	Vinyl Chloride	16.2		0.26	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.1		0.23	1.00	ug/L
78-93-3	2-Butanone	110		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.6		0.25	1.00	ug/L
67-66-3	Chloroform	20.9		0.25	1.00	ug/L
71-43-2	Benzene	19.2		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.0		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.7		0.090	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	1.00	ug/L
108-90-7	Chlorobenzene	19.5		0.12	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.5		74 - 125	101%	SPK: 50
1868-53-7	Dibromofluoromethane	50.3		75 - 124	101%	SPK: 50
2037-26-5	Toluene-d8	47.5		86 - 113	95%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.0		77 - 121	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	81600	5.544			
540-36-3	1,4-Difluorobenzene	146000	6.757			
3114-55-4	Chlorobenzene-d5	128000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	60900	12.018			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046413.D
 Acq On : 30 May 2025 11:55
 Operator : JC/MD
 Sample : VX0530WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBSD01

Manual Integrations
 APPROVED

Reviewed By :John Carlone 06/02/2025
 Supervised By :Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:24:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	81578	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145807	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127712	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60852	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76819	50.510	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 101.020%			
35) Dibromofluoromethane	5.379	113	52825	50.311	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 100.620%			
50) Toluene-d8	8.647	98	172547	47.480	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 94.960%			
62) 4-Bromofluorobenzene	11.079	95	68371	49.047	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 98.100%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	21494	17.214	ug/l	99
3) Chloromethane	1.307	50	19820	16.369	ug/l	96
4) Vinyl Chloride	1.374	62	18255	16.199	ug/l	99
5) Bromomethane	1.593	94	8990	17.200	ug/l	97
6) Chloroethane	1.666	64	11028	18.331	ug/l	95
7) Trichlorofluoromethane	1.874	101	30531	18.332	ug/l	95
8) Diethyl Ether	2.130	74	10561	18.627	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.319	101	19284	18.710	ug/l	98
10) Methyl Iodide	2.447	142	21500	17.629	ug/l	99
11) Tert butyl alcohol	2.971	59	24637	115.404	ug/l	100
12) 1,1-Dichloroethene	2.313	96	16500	17.058	ug/l	98
13) Acrolein	2.233	56	16506	67.891	ug/l	98
14) Allyl chloride	2.654	41	35994	19.470	ug/l	96
15) Acrylonitrile	3.063	53	67712	110.922	ug/l	97
16) Acetone	2.380	43	66624	109.256	ug/l	100
17) Carbon Disulfide	2.502	76	32386	14.122	ug/l #	95
18) Methyl Acetate	2.703	43	37969	26.832	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	69904	20.613	ug/l	98
20) Methylene Chloride	2.782	84	21739	18.603	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	17374	17.860	ug/l	98
22) Diisopropyl ether	3.758	45	74946	20.987	ug/l	95
23) Vinyl Acetate	3.715	43	311196	99.080	ug/l	99
24) 1,1-Dichloroethane	3.605	63	39878	20.049	ug/l	97
25) 2-Butanone	4.556	43	99207	112.061	ug/l	97
26) 2,2-Dichloropropane	4.471	77	27928	17.939	ug/l	99
27) cis-1,2-Dichloroethene	4.483	96	23211	19.820	ug/l	99
28) Bromochloromethane	4.891	49	20979	21.912	ug/l	95
29) Tetrahydrofuran	5.001	42	61848	111.488	ug/l	98
30) Chloroform	5.087	83	43358	20.914	ug/l	96
31) Cyclohexane	5.458	56	31693	17.486	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	34684	19.300	ug/l	100
36) 1,1-Dichloropropene	5.690	75	25597	18.144	ug/l	99
37) Ethyl Acetate	4.715	43	35682	20.472	ug/l	98
38) Carbon Tetrachloride	5.666	117	29548	18.641	ug/l	98
39) Methylcyclohexane	7.373	83	31033	17.087	ug/l	96
40) Benzene	6.032	78	79276	19.185	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046413.D
 Acq On : 30 May 2025 11:55
 Operator : JC/MD
 Sample : VX0530WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 06/02/2025
 Supervised By : Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:24:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	21300	23.361	ug/l	97
42) 1,2-Dichloroethane	6.080	62	35621	19.974	ug/l	100
43) Isopropyl Acetate	6.336	43	56506	21.251	ug/l	98
44) Trichloroethene	7.123	130	18642	18.744	ug/l	97
45) 1,2-Dichloropropane	7.428	63	21035	20.472	ug/l	95
46) Dibromomethane	7.580	93	16483	20.340	ug/l	98
47) Bromodichloromethane	7.818	83	32624	20.440	ug/l	96
48) Methyl methacrylate	7.690	41	29375	21.632	ug/l	98
49) 1,4-Dioxane	7.659	88	12171	472.035	ug/l	99
51) 4-Methyl-2-Pentanone	8.568	43	201687	114.270	ug/l	98
52) Toluene	8.714	92	48516	19.148	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27653	19.492	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	30122	19.210	ug/l	94
55) 1,1,2-Trichloroethane	9.147	97	21232	21.252	ug/l	96
56) Ethyl methacrylate	9.116	69	33915	21.299	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36382	20.277	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	93332	114.971	ug/l	100
59) 2-Hexanone	9.427	43	148657	113.843	ug/l	98
60) Dibromochloromethane	9.519	129	22470	20.479	ug/l	100
61) 1,2-Dibromoethane	9.610	107	21087	20.307	ug/l	99
64) Tetrachloroethene	9.269	164	17333	19.182	ug/l	93
65) Chlorobenzene	10.073	112	54427	19.471	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	19080	19.989	ug/l	97
67) Ethyl Benzene	10.189	91	95748	19.432	ug/l	100
68) m/p-Xylenes	10.299	106	71107	39.457	ug/l	100
69) o-Xylene	10.640	106	35032	19.939	ug/l	97
70) Styrene	10.653	104	58690	20.392	ug/l	99
71) Bromoform	10.799	173	14123	19.708	ug/l #	95
73) Isopropylbenzene	10.957	105	94263	19.897	ug/l	98
74) N-amyl acetate	10.842	43	47306	20.207	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	34310	20.666	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	30249m	20.651	ug/l	
77) Bromobenzene	11.195	156	22032	20.031	ug/l	99
78) n-propylbenzene	11.299	91	108214	19.645	ug/l	100
79) 2-Chlorotoluene	11.360	91	68808	19.366	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	80315	20.293	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	7998	17.777	ug/l	85
82) 4-Chlorotoluene	11.451	91	77928	19.778	ug/l	99
83) tert-Butylbenzene	11.713	119	78878	19.785	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	81005	20.211	ug/l	100
85) sec-Butylbenzene	11.890	105	98194	20.060	ug/l	99
86) p-Isopropyltoluene	12.006	119	81138	20.081	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	40059	19.957	ug/l	98
88) 1,4-Dichlorobenzene	12.037	146	40124	19.573	ug/l	98
89) n-Butylbenzene	12.329	91	68940	19.452	ug/l	99
90) Hexachloroethane	12.536	117	13761	19.332	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	40902	20.306	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7830	21.290	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	22080	19.085	ug/l	99
94) Hexachlorobutadiene	13.719	225	10190	20.167	ug/l	99
95) Naphthalene	13.774	128	83980	19.791	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	23113	19.361	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046413.D
Acq On : 30 May 2025 11:55
Operator : JC/MD
Sample : VX0530WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025

Quant Time: May 30 23:24:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046413.D
Acq On : 30 May 2025 11:55
Operator : JC/MD
Sample : VX0530WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1


Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Manual Integrations APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025


Quant Time: May 30 23:24:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion:168 Resp: 81578
Ion Ratio Lower Upper
168 100
99 71.9 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#2
Dichlorodifluoromethane
Concen: 17.214 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 85 Resp: 21494
Ion Ratio Lower Upper
85 100
87 31.0 15.7 47.1









#3
Chloromethane
Concen: 16.369 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion: 50 Resp: 19820
Ion Ratio Lower Upper
50 100
52 34.0 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#4
Vinyl Chloride
Concen: 16.199 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 62 Resp: 18255
Ion Ratio Lower Upper
62 100
64 30.8 25.2 37.8









#5
Bromomethane
Concen: 17.200 ug/l
RT: 1.593 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion: 94 Resp: 8990
Ion Ratio Lower Upper
94 100
96 93.6 72.8 109.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#6
Chloroethane
Concen: 18.331 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55


Tgt Ion: 64 Resp: 11028
Ion Ratio Lower Upper
64 100
66 27.9 24.3 36.5





#7
Trichlorofluoromethane
Concen: 18.332 ug/l
RT: 1.874 min Scan# 110
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01




Tgt Ion:101 Resp: 3053
Ion Ratio Lower Upper
101 100
103 60.0 51.0 76.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025




#8
Diethyl Ether
Concen: 18.627 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55



Tgt Ion: 74 Resp: 10561
Ion Ratio Lower Upper
74 100
45 111.0 54.9 164.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 18.710 ug/l

RT: 2.319 min Scan# 202

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01





Tgt Ion:	101	Resp:	19284
Ion	Ratio	Lower	Upper
101	100		
85	47.7	38.6	58.0
151	71.6	55.2	82.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#10

Methyl Iodide

Concen: 17.629 ug/l




RT: 2.447 min Scan# 223


Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Tgt Ion:	142	Resp:	21500
Ion	Ratio	Lower	Upper
142	100		
127	51.0	41.7	62.5
141	14.7	11.5	17.3





#11

Tert butyl alcohol

Concen: 115.404 ug/l

RT: 2.971 min Scan# 309

Delta R.T. -0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 59 Resp: 2463

Ion Ratio Lower Upper

59 100





57 10.6 8.6 12.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#12

1,1-Dichloroethene

Concen: 17.058 ug/l

RT: 2.313 min Scan# 201

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion: 96 Resp: 16500


Ion Ratio Lower Upper

96 100

61 186.1 146.2 219.2

98 64.0 51.0 76.6





#13

Acrolein

Concen: 67.891 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 56 Resp: 16500

Ion Ratio Lower Upper

56 100





55 72.0 56.2 84.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#14

Allyl chloride

Concen: 19.470 ug/l

RT: 2.654 min Scan# 257

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion: 41 Resp: 35994


Ion Ratio Lower Upper

41 100

39 73.2 60.6 90.8

76 27.8 24.9 37.3





#15

Acrylonitrile

Concen: 110.922 ug/l

RT: 3.063 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01





Tgt Ion:	53	Resp:	6771
Ion Ratio	Lower	Upper	
53	100		
52	84.3	65.3	97.9
51	38.3	29.8	44.8

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#16

Acetone

Concen: 109.256 ug/l




RT: 2.380 min Scan# 212


Delta R.T. -0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55


Tgt Ion:	43	Resp:	66624
Ion Ratio	Lower	Upper	
43	100		
58	26.4	21.2	31.8





#17
Carbon Disulfide
Concen: 14.122 ug/l
RT: 2.502 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion: 76 Resp: 32380
Ion Ratio Lower Upper
76 100
78 10.5 7.0 10.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#18
Methyl Acetate
Concen: 26.832 ug/l
RT: 2.703 min Scan# 265
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 43 Resp: 37969
Ion Ratio Lower Upper
43 100
74 21.8 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 20.613 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion: 73 Resp: 69904
Ion Ratio Lower Upper
73 100
57 22.8 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#20
Methylene Chloride
Concen: 18.603 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55


Tgt Ion: 84 Resp: 21739
Ion Ratio Lower Upper
84 100
49 139.2 113.9 170.9
51 43.4 33.5 50.3
86 66.1 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 17.860 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion: 96 Resp: 17374
Ion Ratio Lower Upper
96 100
61 153.5 119.5 179.3
98 62.1 50.0 75.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#22
Diisopropyl ether
Concen: 20.987 ug/l
RT: 3.758 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55


Tgt Ion: 45 Resp: 74946
Ion Ratio Lower Upper
45 100
43 89.9 66.6 100.0
87 24.8 19.8 29.6
59 11.4 8.6 12.8





#23
Vinyl Acetate
Concen: 99.080 ug/l
RT: 3.715 min Scan# 411
Delta R.T. -0.006 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion: 43 Resp: 311190
Ion Ratio Lower Upper
43 100
86 9.7 7.5 11.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#24
1,1-Dichloroethane
Concen: 20.049 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 63 Resp: 39878
Ion Ratio Lower Upper
63 100
98 5.9 3.6 10.8
100 4.3 2.1 6.3





#25

2-Butanone

Concen: 112.061 ug/l

RT: 4.556 min Scan# 50

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 43 Resp: 9920

Ion Ratio Lower Upper

43 100





72 24.3 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#26

2,2-Dichloropropane

Concen: 17.939 ug/l

RT: 4.471 min Scan# 555

Delta R.T. 0.006 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion: 77 Resp: 27928

Ion Ratio Lower Upper

77 100

97 21.6 10.5 31.5





#27

cis-1,2-Dichloroethene

Concen: 19.820 ug/l

RT: 4.483 min Scan# 557

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 96 Resp: 2321

Ion Ratio Lower Upper

96 100

61 160.5 0.0 322.8





98 63.3 0.0 129.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#28

Bromochloromethane

Concen: 21.912 ug/l

RT: 4.891 min Scan# 624

Delta R.T. -0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion: 49 Resp: 20979


Ion Ratio Lower Upper

49 100

129 1.4 0.0 4.0

130 66.3 56.2 84.2





#29

Tetrahydrofuran

Concen: 111.488 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 42 Resp: 6184

Ion Ratio Lower Upper

42 100

72 40.5 33.7 50.5





71 38.2 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#30

Chloroform

Concen: 20.914 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion: 83 Resp: 43358

Ion Ratio Lower Upper

83 100

85 64.5 49.3 73.9





#31

Cyclohexane

Concen: 17.486 ug/l

RT: 5.458 min Scan# 717

Delta R.T. -0.006 min

Lab File: VX046413.D


Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01






Tgt Ion:	56	Resp:	3169
Ion	Ratio	Lower	Upper
56	100		
69	30.0	24.4	36.6
84	79.3	66.9	100.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#32

1,1,1-Trichloroethane


Concen: 19.300 ug/l

RT: 5.373 min Scan# 703



Delta R.T. 0.000 min


Lab File: VX046413.D

Acq: 30 May 2025 11:55



Tgt Ion:	97	Resp:	34684
Ion	Ratio	Lower	Upper
97	100		
99	64.7	51.8	77.6
61	50.6	40.1	60.1





#33

1,2-Dichloroethane-d4

Concen: 50.510 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 65 Resp: 76819

Ion Ratio Lower Upper

65 100





67 49.1 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion:114 Resp: 145807


Ion Ratio Lower Upper

114 100

63 22.7 0.0 49.2

88 17.1 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.311 ug/l

RT: 5.379 min Scan# 704

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01





Tgt Ion:	113	Resp:	5282
Ion	Ratio	Lower	Upper
113	100		
111	101.2	83.1	124.7
192	16.6	13.3	19.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#36

1,1-Dichloropropene

Concen: 18.144 ug/l




RT: 5.690 min Scan# 755


Delta R.T. 0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Tgt Ion:	75	Resp:	25597
Ion	Ratio	Lower	Upper
75	100		
110	31.6	16.3	48.9
77	30.5	24.3	36.5





#37

Ethyl Acetate

Concen: 20.472 ug/l

RT: 4.715 min Scan# 595

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 43 Resp: 3568

Ion Ratio Lower Upper

43 100

61 12.3 10.3 15.5





70 9.2 7.9 11.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#38

Carbon Tetrachloride

Concen: 18.641 ug/l

RT: 5.666 min Scan# 751

Delta R.T. -0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion:117 Resp: 29548


Ion Ratio Lower Upper

117 100

119 96.5 75.2 112.8

121 30.2 24.2 36.4





#39

Methylcyclohexane

Concen: 17.087 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 83 Resp: 3103

Ion Ratio Lower Upper

83 100

55 83.9 64.7 97.1





98 49.5 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#40

Benzene

Concen: 19.185 ug/l

RT: 6.032 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion: 78 Resp: 79276

Ion Ratio Lower Upper

78 100


77 24.2 19.0 28.4





#41
Methacrylonitrile
Concen: 23.361 ug/l
RT: 4.916 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion: 41 Resp: 21300
Ion Ratio Lower Upper
41 100
39 54.9 47.2 70.8
67 62.6 50.7 76.1
52 31.4 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#42
1,2-Dichloroethane
Concen: 19.974 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 62 Resp: 35621
Ion Ratio Lower Upper
62 100
98 7.5 0.0 15.2





#43

Isopropyl Acetate

Concen: 21.251 ug/l

RT: 6.336 min Scan# 861

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 43 Resp: 56500

Ion Ratio Lower Upper

43 100

61 16.9 14.3 21.5





87 11.0 9.5 14.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#44

Trichloroethene

Concen: 18.744 ug/l

RT: 7.123 min Scan# 990

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion:130 Resp: 18642

Ion Ratio Lower Upper

130 100

95 99.4 0.0 204.2





#45

1,2-Dichloropropane

Concen: 20.472 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 63 Resp: 2103

Ion Ratio Lower Upper

63 100





65 32.6 24.1 36.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#46

Dibromomethane

Concen: 20.340 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. 0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion: 93 Resp: 16483


Ion Ratio Lower Upper

93 100

95 79.8 65.6 98.4

174 83.0 68.2 102.2





#47

Bromodichloromethane

Concen: 20.440 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01





Tgt Ion:	83	Resp:	32624
Ion Ratio	Lower	Upper	
83	100		
85	60.2	50.5	75.7
127	8.6	6.5	9.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#48

Methyl methacrylate

Concen: 21.632 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Tgt Ion:	41	Resp:	29375
Ion Ratio	Lower	Upper	
41	100		
69	71.1	58.5	87.7
39	63.6	51.7	77.5





#49

1,4-Dioxane

Concen: 472.035 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 88 Resp: 1217

Ion Ratio Lower Upper

88 100

43 42.3 33.4 50.2





58 71.9 58.6 88.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#50

Toluene-d8

Concen: 47.480 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion: 98 Resp: 172547

Ion Ratio Lower Upper

98 100

100 64.5 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 114.270 ug/l
RT: 8.568 min Scan# 1227
Delta R.T. -0.006 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion: 43 Resp: 20168
Ion Ratio Lower Upper
43 100
58 34.7 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#52
Toluene
Concen: 19.148 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 92 Resp: 48516
Ion Ratio Lower Upper
92 100
91 171.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 19.492 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 75 Resp: 2765

Ion Ratio Lower Upper

75 100





77 32.1 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#54

cis-1,3-Dichloropropene

Concen: 19.210 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.006 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion: 75 Resp: 30122


Ion Ratio Lower Upper

75 100

77 33.0 25.4 38.0

39 59.1 52.2 78.4





#55

1,1,2-Trichloroethane

Concen: 21.252 ug/l

RT: 9.147 min Scan# 1322

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 97 Resp: 2123

Ion Ratio Lower Upper

97 100

83 89.4 69.0 103.6

85 58.1 44.3 66.5





99 63.1 47.8 71.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#56

Ethyl methacrylate

Concen: 21.299 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion: 69 Resp: 33915


Ion Ratio Lower Upper

69 100

41 76.7 60.8 91.2

39 46.8 39.0 58.6





#57

1,3-Dichloropropane

Concen: 20.277 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 76 Resp: 3638

Ion Ratio Lower Upper

76 100





78 32.3 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#58

2-Chloroethyl Vinyl ether

Concen: 114.971 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion: 63 Resp: 93332

Ion Ratio Lower Upper

63 100

106 27.0 21.5 32.3





#59

2-Hexanone

Concen: 113.843 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 43 Resp: 14865

Ion Ratio Lower Upper

43 100





58 48.5 24.9 74.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#60

Dibromochloromethane

Concen: 20.479 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion:129 Resp: 22470

Ion Ratio Lower Upper

129 100

127 78.1 39.3 117.8









#61
1,2-Dibromoethane
Concen: 20.307 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion:107 Resp: 2108
Ion Ratio Lower Upper
107 100
109 92.5 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#62
4-Bromofluorobenzene
Concen: 49.047 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55


Tgt Ion: 95 Resp: 68371
Ion Ratio Lower Upper
95 100
174 67.7 0.0 135.8
176 65.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion:117 Resp: 127712
Ion Ratio Lower Upper
117 100
82 60.3 50.6 76.0
119 32.4 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#64
Tetrachloroethene
Concen: 19.182 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion:164 Resp: 17333
Ion Ratio Lower Upper
164 100
166 121.5 105.0 157.6
129 95.4 83.5 125.3
131 91.1 76.5 114.7









#65
Chlorobenzene
Concen: 19.471 ug/l
RT: 10.073 min Scan# 1475
Delta R.T. -0.006 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion:112 Resp: 5442
Ion Ratio Lower Upper
112 100
114 30.5 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 19.989 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion:131 Resp: 19080
Ion Ratio Lower Upper
131 100
133 96.6 47.3 141.9
119 66.2 31.6 95.0









#67
Ethyl Benzene
Concen: 19.432 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion: 91 Resp: 95748
Ion Ratio Lower Upper
91 100
106 29.6 23.4 35.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#68
m/p-Xylenes
Concen: 39.457 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion:106 Resp: 71107
Ion Ratio Lower Upper
106 100
91 214.5 171.2 256.8









#69
o-Xylene
Concen: 19.939 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion:106 Resp: 3503
Ion Ratio Lower Upper
106 100
91 230.4 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#70
Styrene
Concen: 20.392 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55


Tgt Ion:104 Resp: 58690
Ion Ratio Lower Upper
104 100
78 58.5 45.7 68.5
103 54.1 43.7 65.5





#71
Bromoform
Concen: 19.708 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion:173 Resp: 1412
Ion Ratio Lower Upper
173 100
175 50.2 23.4 70.0
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion:152 Resp: 60852
Ion Ratio Lower Upper
152 100
115 75.8 46.9 140.7
150 163.5 0.0 351.0









#73
Isopropylbenzene
Concen: 19.897 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion:105 Resp: 9426
Ion Ratio Lower Upper
105 100
120 24.7 12.8 38.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#74
N-ethyl acetate
Concen: 20.207 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 43 Resp: 47306
Ion Ratio Lower Upper
43 100
70 36.8 30.9 46.3
55 23.6 18.7 28.1
61 22.0 17.1 25.7





#75

1,1,2,2-Tetrachloroethane

Concen: 20.666 ug/l

RT: 11.207 min Scan# 1660

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 83 Resp: 34310

Ion Ratio Lower Upper

83 100

131 9.1 5.0 14.9





85 63.6 31.9 95.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#76

1,2,3-Trichloropropane

Concen: 20.651 ug/l m

RT: 11.238 min Scan# 1665

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion: 75 Resp: 30249

Ion Ratio Lower Upper

75 100

77 41.3 20.5 61.5





#77

Bromobenzene

Concen: 20.031 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01





Tgt Ion	Ratio	Lower	Upper
156	100		
77	184.6	92.8	278.3
158	95.8	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#78

n-propylbenzene

Concen: 19.645 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Tgt Ion	Ratio	Lower	Upper
91	100		
120	21.5	10.8	32.4





#79

2-Chlorotoluene

Concen: 19.366 ug/l

RT: 11.360 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 91 Resp: 6880

Ion Ratio Lower Upper

91 100





126 31.6 15.6 46.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#80

1,3,5-Trimethylbenzene

Concen: 20.293 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion:105 Resp: 80315

Ion Ratio Lower Upper

105 100


120 46.0 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 17.777 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01




Tgt Ion: 75 Resp: 7998
Ion Ratio Lower Upper
75 100
53 147.6 100.5 150.7
89 49.4 37.0 55.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025




#82
4-Chlorotoluene
Concen: 19.778 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55



Tgt Ion: 91 Resp: 77928
Ion Ratio Lower Upper
91 100
126 26.9 13.3 39.8





#83

tert-Butylbenzene

Concen: 19.785 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion:119 Resp: 78878

Ion Ratio Lower Upper

119 100

91 65.4 32.9 98.7





134 22.7 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#84

1,2,4-Trimethylbenzene

Concen: 20.211 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion:105 Resp: 81005

Ion Ratio Lower Upper

105 100

120 42.6 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 20.060 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046413.D
 Acq: 30 May 2025 11:55

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0530WBSD01

Tgt Ion:105 Resp: 98194
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.7 29.1




Manual Integrations
 APPROVED


Reviewed By :John Carlone 06/02/2025
 Supervised By :Mahesh Dadoda 06/02/2025



#86
 p-Isopropyltoluene
 Concen: 20.081 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046413.D
 Acq: 30 May 2025 11:55

Tgt Ion:119 Resp: 81138
 Ion Ratio Lower Upper
 119 100
 134 24.9 12.5 37.5
 91 27.9 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 19.957 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion:146 Resp: 40059

Ion Ratio Lower Upper

146 100

111 43.9 22.1 66.3





148 62.2 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#88

1,4-Dichlorobenzene

Concen: 19.573 ug/l

RT: 12.037 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55




Tgt Ion:146 Resp: 40124


Ion Ratio Lower Upper

146 100

111 44.3 21.3 63.9

148 62.9 31.9 95.5





#89

n-Butylbenzene

Concen: 19.452 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046413.D

Acq: 30 May 2025 11:55

Instrument :

MSVOA_X

ClientSampleId :

VX0530WBSD01

Tgt Ion: 91 Resp: 68940

Ion Ratio Lower Upper

91 100

92 53.1 26.9 80.7





134 23.7 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 06/02/2025

Supervised By :Mahesh Dadoda 06/02/2025



#90

Hexachloroethane

Concen: 19.332 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046413.D




Acq: 30 May 2025 11:55


Tgt Ion:117 Resp: 13761

Ion Ratio Lower Upper

117 100


201 61.9 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 20.306 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion:146 Resp: 40902
Ion Ratio Lower Upper
146 100
111 45.0 22.4 67.3
148 63.4 31.6 94.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.290 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion: 75 Resp: 7830
Ion Ratio Lower Upper
75 100
155 71.0 34.9 104.8
157 85.5 43.8 131.4









#93
1,2,4-Trichlorobenzene
Concen: 19.085 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01

Tgt Ion:180 Resp: 22080
Ion Ratio Lower Upper
180 100
182 96.3 47.9 143.8
145 35.3 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#94
Hexachlorobutadiene
Concen: 20.167 ug/l
RT: 13.719 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55


Tgt Ion:225 Resp: 10190
Ion Ratio Lower Upper
225 100
223 63.1 30.8 92.4
227 61.2 30.9 92.7





#95
Naphthalene
Concen: 19.791 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55




Instrument :
MSVOA_X
ClientSampleId :
VX0530WBSD01



Tgt Ion:128 Resp: 83980
Ion Ratio Lower Upper
128 100
127 12.8 10.4 15.6
129 10.9 8.6 13.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 06/02/2025
Supervised By :Mahesh Dadoda 06/02/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.361 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046413.D
Acq: 30 May 2025 11:55

Tgt Ion:180 Resp: 23113
Ion Ratio Lower Upper
180 100
182 93.6 47.8 143.3
145 36.9 18.1 54.3



Manual Integration Report

Sequence:	VX050525	Instrument	MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC020	VX046041.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:13 AM	MMDadoda	5/6/2025 12:42:46 PM	Peak Integrated by Software
VSTDICCC050	VX046042.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:18 AM	MMDadoda	5/6/2025 12:42:48 PM	Peak Integrated by Software
VSTDICC100	VX046043.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:22 AM	MMDadoda	5/6/2025 12:42:50 PM	Peak Integrated by Software
VSTDICC150	VX046044.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:27 AM	MMDadoda	5/6/2025 12:42:53 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	Ethyl Acetate	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,4-Dichlorobenzene	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Bromochloromethane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Ethyl Acetate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Methyl methacrylate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICV050	VX046048.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:45 AM	MMDadoda	5/6/2025 12:41:37 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VX050525

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

VX053025

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX046409.D	1,2,3-Trichloropropane	JOHN	6/2/2025 9:39:42 AM	MMDadoda	6/2/2025 1:33:37 PM	Peak Integrated by Software
VX0530WBS01	VX046412.D	1,2,3-Trichloropropane	JOHN	6/2/2025 9:39:46 AM	MMDadoda	6/2/2025 1:33:38 PM	Peak Integrated by Software
VX0530WBSD0 1	VX046413.D	1,2,3-Trichloropropane	JOHN	6/2/2025 9:39:51 AM	MMDadoda	6/2/2025 1:33:39 PM	Peak Integrated by Software
VSTDCCC050	VX046431.D	1,2,3-Trichloropropane	JOHN	6/2/2025 9:40:28 AM	MMDadoda	6/2/2025 1:33:47 PM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046038.D	05 May 2025 09:37	JC/MD	Ok
2	VSTDIC001	VX046039.D	05 May 2025 10:49	JC/MD	Not Ok
3	VSTDIC005	VX046040.D	05 May 2025 11:12	JC/MD	Not Ok
4	VSTDIC020	VX046041.D	05 May 2025 11:35	JC/MD	Ok,M
5	VSTDIC050	VX046042.D	05 May 2025 11:58	JC/MD	Ok,M
6	VSTDIC100	VX046043.D	05 May 2025 12:21	JC/MD	Ok,M
7	VSTDIC150	VX046044.D	05 May 2025 12:45	JC/MD	Ok,M
8	IBLK	VX046045.D	05 May 2025 13:08	JC/MD	Ok
9	VSTDIC005	VX046046.D	05 May 2025 16:04	JC/MD	Ok,M
10	VSTDIC001	VX046047.D	05 May 2025 16:27	JC/MD	Ok,M
11	VSTDICV050	VX046048.D	05 May 2025 16:50	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX053025

Review By	Mahesh Dadoda	Review On	6/2/2025 1:33:49 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/2/2025 1:35:05 PM
SubDirectory	VX053025	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134079		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134080,VP134081		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046408.D	30 May 2025 08:20	JC/MD	Ok
2	VSTDCCC050	VX046409.D	30 May 2025 10:14	JC/MD	Ok,M
3	VX0530MBL01	VX046410.D	30 May 2025 10:42	JC/MD	Ok
4	VX0530WBL01	VX046411.D	30 May 2025 11:06	JC/MD	Ok
5	VX0530WBS01	VX046412.D	30 May 2025 11:29	JC/MD	Ok,M
6	VX0530WBSD01	VX046413.D	30 May 2025 11:55	JC/MD	Ok,M
7	PB168191TB	VX046414.D	30 May 2025 12:18	JC/MD	Ok
8	Q2136-05	VX046415.D	30 May 2025 12:42	JC/MD	Ok,M
9	Q2146-04	VX046416.D	30 May 2025 13:05	JC/MD	Ok
10	Q2151-04	VX046417.D	30 May 2025 13:29	JC/MD	Ok
11	Q2149-01	VX046418.D	30 May 2025 13:52	JC/MD	Ok
12	IBLK	VX046419.D	30 May 2025 14:15	JC/MD	Ok
13	Q2163-02	VX046420.D	30 May 2025 14:39	JC/MD	Ok
14	Q2162-02	VX046421.D	30 May 2025 15:02	JC/MD	Ok
15	Q2162-10	VX046422.D	30 May 2025 15:26	JC/MD	ReRun
16	Q2162-07	VX046423.D	30 May 2025 15:49	JC/MD	Ok
17	Q2162-09	VX046424.D	30 May 2025 16:12	JC/MD	Ok
18	Q2163-01	VX046425.D	30 May 2025 16:36	JC/MD	Ok
19	IBLK	VX046426.D	30 May 2025 16:59	JC/MD	Ok
20	Q2162-10	VX046427.D	30 May 2025 17:23	JC/MD	Ok,M
21	Q2162-04	VX046428.D	30 May 2025 17:46	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX053025

Review By	Mahesh Dadoda	Review On	6/2/2025 1:33:49 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/2/2025 1:35:05 PM
SubDirectory	VX053025	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134079		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134080,VP134081		

22	Q2162-05MS	VX046429.D	30 May 2025 18:09	JC/MD	Ok,M
23	Q2162-06MSD	VX046430.D	30 May 2025 18:33	JC/MD	Ok,M
24	VSTDCCC050	VX046431.D	30 May 2025 18:56	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133811 VP133832,VP133833,VP133834,VP133835,VP133836,VP133837 VP133838

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX053025

Review By	Mahesh Dadoda	Review On	6/2/2025 1:33:49 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/2/2025 1:35:05 PM
SubDirectory	VX053025	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134079
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134080,VP134081

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046408.D	30 May 2025 08:20		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046409.D	30 May 2025 10:14	pH#Lot#V12668	JC/MD	Ok,M
3	VX0530MBL01	VX0530MBL01	VX046410.D	30 May 2025 10:42		JC/MD	Ok
4	VX0530WBL01	VX0530WBL01	VX046411.D	30 May 2025 11:06		JC/MD	Ok
5	VX0530WBS01	VX0530WBS01	VX046412.D	30 May 2025 11:29		JC/MD	Ok,M
6	VX0530WBSD01	VX0530WBSD01	VX046413.D	30 May 2025 11:55		JC/MD	Ok,M
7	PB168191TB	PB168191TB	VX046414.D	30 May 2025 12:18		JC/MD	Ok
8	Q2136-05	OR-646-COMP-52	VX046415.D	30 May 2025 12:42	vial A pH#5.0	JC/MD	Ok,M
9	Q2146-04	TP04-MHN-WC	VX046416.D	30 May 2025 13:05	vial A pH#5.0	JC/MD	Ok
10	Q2151-04	WC-1	VX046417.D	30 May 2025 13:29	vial A pH#5.0	JC/MD	Ok
11	Q2149-01	FILTER-CAKE	VX046418.D	30 May 2025 13:52	vial A pH#5.0 Sample very turbid	JC/MD	Ok
12	IBLK	IBLK	VX046419.D	30 May 2025 14:15		JC/MD	Ok
13	Q2163-02	250528060-03-TRIP-BL	VX046420.D	30 May 2025 14:39	vial A pH#6.0	JC/MD	Ok
14	Q2162-02	BP-VPB-182-TB-20250	VX046421.D	30 May 2025 15:02	vial A pH<2 TB	JC/MD	Ok
15	Q2162-10	BP-VPB-182-EB-20250	VX046422.D	30 May 2025 15:26	vial A pH<2 EB;Hit of comp.#16,20,25	JC/MD	ReRun
16	Q2162-07	BP-VPB-182-GW-620-6	VX046423.D	30 May 2025 15:49	vial A pH<2	JC/MD	Ok
17	Q2162-09	BP-VPB-182-DUP-2025	VX046424.D	30 May 2025 16:12	please change back to sample #08	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX053025

Review By	Maresh Dadoda	Review On	6/2/2025 1:33:49 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/2/2025 1:35:05 PM
SubDirectory	VX053025	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134079		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134080,VP134081		

18	Q2163-01	250528063-01	VX046425.D	30 May 2025 16:36	vial A pH#6.0	JC/MD	Ok
19	IBLK	IBLK	VX046426.D	30 May 2025 16:59		JC/MD	Ok
20	Q2162-10	BP-VPB-182-EB-20250	VX046427.D	30 May 2025 17:23	vial B pH<2 EB	JC/MD	Ok,M
21	Q2162-04	BP-VPB-182-GW-600-6	VX046428.D	30 May 2025 17:46	vial A+B pH<2	JC/MD	Ok
22	Q2162-05MS	BP-VPB-182-GW-600-6	VX046429.D	30 May 2025 18:09	vial A+B pH<2	JC/MD	Ok,M
23	Q2162-06MSD	BP-VPB-182-GW-600-6	VX046430.D	30 May 2025 18:33	vial A+B pH<2	JC/MD	Ok,M
24	VSTDCCC050	VSTDCCC050EC	VX046431.D	30 May 2025 18:56		JC/MD	Ok,M

M : Manual Integration

SOP ID :	M1311-TCLP-16	
SDG No :	N/A	Start Prep Date : 05/29/2025 Time : 14:00
Welgh By :	JP	End Prep Date : 05/30/2025 Time : 08:15
Balance ID :	WC SC-7	Combination Ratio : 20
pH Meter ID :	WC PH METER-1	ZHE Cleaning Batch : ¹⁸ N/A V2053025
Extraction By :	JP	Initial Room Temperature: 23 °C
Filter By :	JP	Final Room Temperature: 22 °C
Pippete ID :	WC	TCLP Technician Signature : ¹⁸
Tumbler ID :	ZHE-1	Supervisor By : 12
TCLP Filter ID :	50223706	

Standarded Name	MLS USED	STD REF. # FROM LOG
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE U	Lot Number
TCLP-FLUID-1	N/A	WP112795
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
40ml VOA Vials	N/A	MP84041

Extraction Conformance/Non-Conformance Comments:

ALL ZHE samples are extracted and given as vial A & B. Leak checked after 10 minutes if tumbling. TUMBLER ZHE-1 checked,30 rpm.

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
05/30/25 09:50	¹⁸ 1164 Room	S.Y 1VUC 245
	Preparation Group	Analysis Group

Sample ID	ClientID	ZHE Vessel ID	Sample Wt (g)	Volume Extraction Fluid #1 (mL)	Multi phasic	Phase Miscible	Phases Combined	Final Leachate PH	Metals Leachate Adj. PH	Prep Pos
PB168191TB	LEB191	05	N/A	500	N/A	N/A	N/A	4.94	N/A	ZHE-1
Q2136-05	OR-646-COMP-52	01	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2146-04	TP04-MHN-WC	02	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2149-01	FILTER-CAKE	03	25.02	500	N/A	N/A	N/A	N/A	N/A	ZHE-1
Q2151-04	WC-1	04	25.03	500	N/A	N/A	N/A	N/A	N/A	ZHE-1

SampleID	ClientID	Sample Weight (g)	Filter Weight (g)	Filtrate (mL)	Filter + Solid (After 100°C)	% solids	% Dry Solids
PB168191TB	LEB191	N/A	N/A	N/A	N/A	N/A	N/A
Q2136-05	OR-646-COMP-52	N/A	N/A	N/A	N/A	100	N/A
Q2146-04	TP04-MHN-WC	N/A	N/A	N/A	N/A	100	N/A
Q2149-01	FILTER-CAKE	N/A	N/A	N/A	N/A	100	N/A
Q2151-04	WC-1	N/A	N/A	N/A	N/A	100	N/A

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tclp zhe q2151 WorkList ID : 189816 Department : TCLP Extraction Date : 05-29-2025 12:50:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2136-05	OR-646-COMP-52	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	L41	05/27/2025	1311 ZHE
Q2146-04	TP04-MHN-WC	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	L31	05/28/2025	1311 ZHE
Q2149-01	FILTER-CAKE	Solid	TCLP ZHE Extraction	Cool 4 deg C	ARAM01	L31	05/28/2025	1311 ZHE
Q2151-04	WC-1	Solid	TCLP ZHE Extraction	Cool 4 deg C	PSEG03	L41	05/29/2025	1311 ZHE

Date/Time 05/29/25 13:15
 Raw Sample Received by: JTW
 Raw Sample Relinquished by: JTW

Date/Time 05/29/25 17:00
 Raw Sample Received by: JTW
 Raw Sample Relinquished by: JTW



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Aramark uniforms
ADDRESS: 740 Frellinghuysen Ave
CITY Newark STATE: NJ ZIP: 07114
ATTENTION: Jarrod Mills
PHONE: 973-824-1101 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Monthly
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

1. TAP Interstitial 2. TAP Tap metals 3. TAP mercury 4. Corrosivity 5. Ignitability 6. Reactivity 7. Reactivity 8. TAP Vol Group 9. TAP PAH Group

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	Filter Cake	S	✓	✓	5-28-25	1005	4	✓	✓	✓	✓	✓	✓	✓	✓	✓	
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>100</u> <u>5-28-25</u>	RECEIVED BY: 1. <u>[Signature]</u>	DATE/TIME: <u>100</u> <u>5-28-25</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>3.0</u> °C
RELINQUISHED BY SAMPLER: 2. <u>[Signature]</u>	DATE/TIME:	RECEIVED BY: 2. <u>[Signature]</u>		Comments:
RELINQUISHED BY SAMPLER: 3. <u>[Signature]</u>	DATE/TIME: <u>133</u> <u>5-28-25</u>	RECEIVED BY: 3. <u>[Signature]</u>		Client: ☐ Hand Delivered ☐ Other

Page ____ of

Shipment Complete
☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488