

Surrogate Summary

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2149-01	FILTER-CAKE	1,2-Dichloroethane-d4	50	54.9	110	74	125
		Dibromofluoromethane	50	51.9	104	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	51.3	103	77	121

Surrogate Summary

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0530WBL01	VX0530WBL01	1,2-Dichloroethane-d4	50	54.0	108	74	125
		Dibromofluoromethane	50	50.2	100	75	124
		Toluene-d8	50	50.1	100	86	113
		4-Bromofluorobenzene	50	50.8	102	77	121
VX0530WBS01	VX0530WBS01	1,2-Dichloroethane-d4	50	49.9	100	74	125
		Dibromofluoromethane	50	50.4	101	75	124
		Toluene-d8	50	47.8	96	86	113
		4-Bromofluorobenzene	50	49.4	99	77	121
VX0530WBSD0	VX0530WBSD01	1,2-Dichloroethane-d4	50	50.5	101	74	125
		Dibromofluoromethane	50	50.3	101	75	124
		Toluene-d8	50	47.5	95	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121