

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260-Low

Datafile : VX046412.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0530WBS01	Vinyl chloride	20	16.5	ug/L	83			65	117	
	1,1-Dichloroethene	20	17.9	ug/L	90			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	18.7	ug/L	94			77	113	
	Chloroform	20	20.5	ug/L	103			79	113	
	Benzene	20	19.3	ug/L	97			82	109	
	1,2-Dichloroethane	20	19.9	ug/L	100			80	115	
	Trichloroethene	20	18.8	ug/L	94			77	113	
	Tetrachloroethene	20	18.4	ug/L	92			67	123	
	Chlorobenzene	20	19.1	ug/L	96			82	109	