

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2149

Client: Aramark Uniforms

Analytical Method: SW8260-Low

Datafile : VX046413.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0530WBSD01	Vinyl chloride	20	16.2	ug/L	81	2		65	117	20
	1,1-Dichloroethene	20	17.1	ug/L	86	5		74	110	20
	2-Butanone	100	110	ug/L	110	0		65	122	20
	Carbon Tetrachloride	20	18.6	ug/L	93	1		77	113	20
	Chloroform	20	20.9	ug/L	104	1		79	113	20
	Benzene	20	19.2	ug/L	96	1		82	109	20
	1,2-Dichloroethane	20	20.0	ug/L	100	0		80	115	20
	Trichloroethene	20	18.7	ug/L	94	0		77	113	20
	Tetrachloroethene	20	19.2	ug/L	96	4		67	123	20
	Chlorobenzene	20	19.5	ug/L	98	2		82	109	20