

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ARAM01
 Lab Code: CHEM Case No.: Q2149 SAS No.: Q2149 SDG NO.: Q2149
 Lab File ID: VX046409.D Date Analyzed: 05/30/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91458	5.54	155659	6.75	130840	10.05
UPPER LIMIT	182916	6.038	311318	7.251	261680	10.549
LOWER LIMIT	45729	5.038	77829.5	6.251	65420	9.549
EPA SAMPLE NO.						
FILTER-CAKE	62556	5.54	124459	6.76	115572	10.05
VX0530WBL01	64533	5.55	129905	6.76	123011	10.05
VX0530WBS01	86619	5.55	153415	6.76	134836	10.05
VX0530WBSD01	81578	5.54	145807	6.76	127712	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.