

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135958

Review By	MOHAN	Review On	6/2/2025 10:14:34 AM
Supervise By	jaswal	Supervise On	6/2/2025 4:46:20 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85798,MP85799,MP85800,MP85801,MP85802,MP85803		
ICV Standard	MP85804		
CCV Standard	MP85806		
ICSA Standard			
CRI Standard	MP85808		
LCS Standard			
Chk Standard	MP85805,MP85807,MP85809,MP85811		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	05/30/25 11:45		MOHAN	OK
2	S0.2	S0.2	CAL2	05/30/25 11:48		MOHAN	OK
3	S2.5	S2.5	CAL3	05/30/25 11:50		MOHAN	OK
4	S5	S5	CAL4	05/30/25 11:58		MOHAN	OK
5	S7.5	S7.5	CAL5	05/30/25 12:03		MOHAN	OK
6	S10	S10	CAL6	05/30/25 12:10		MOHAN	OK
7	ICV22	ICV22	ICV	05/30/25 12:16		MOHAN	OK
8	ICB22	ICB22	ICB	05/30/25 12:18		MOHAN	OK
9	CCV79	CCV79	CCV	05/30/25 12:20		MOHAN	OK
10	CCB79	CCB79	CCB	05/30/25 12:23		MOHAN	OK
11	CRA	CRA	CRDL	05/30/25 12:25		MOHAN	OK
12	HighStd	HighStd	HIGH STD	05/30/25 12:27		MOHAN	OK
13	ChkStd	ChkStd	SAM	05/30/25 12:29		MOHAN	OK
14	PB168222BL	PB168222BL	MB	05/30/25 12:35		MOHAN	OK
15	PB168222BS	PB168222BS	LCS	05/30/25 12:42		MOHAN	OK
16	Q2136-01	OR-646-COMP-52	SAM	05/30/25 12:44		MOHAN	OK
17	Q2137-01	MOO-25-0151	SAM	05/30/25 12:46		MOHAN	OK
18	Q2139-01	BELL-25-0010	SAM	05/30/25 12:48		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135958

Review By	MOHAN	Review On	6/2/2025 10:14:34 AM
Supervise By	jaswal	Supervise On	6/2/2025 4:46:20 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85798,MP85799,MP85800,MP85801,MP85802,MP85803		
ICV Standard	MP85804		
CCV Standard	MP85806		
ICSA Standard			
CRI Standard	MP85808		
LCS Standard			
Chk Standard	MP85805,MP85807,MP85809,MP85811		

19	Q2141-01	VNJ-215	SAM	05/30/25 12:51		MOHAN	OK
20	Q2146-01	TP04-MHN-WC	SAM	05/30/25 12:53		MOHAN	OK
21	CCV80	CCV80	CCV	05/30/25 12:55		MOHAN	OK
22	CCB80	CCB80	CCB	05/30/25 12:58		MOHAN	OK
23	Q2150-01	TP-44	SAM	05/30/25 13:05		MOHAN	OK
24	Q2150-01DUP	TP-44DUP	DUP	05/30/25 13:07		MOHAN	OK
25	Q2150-01MS	TP-44MS	MS	05/30/25 13:12		MOHAN	OK
26	Q2150-01MSD	TP-44MSD	MSD	05/30/25 13:15		MOHAN	OK
27	Q2150-02	TP-42	SAM	05/30/25 13:17		MOHAN	OK
28	Q2150-03	TP-39	SAM	05/30/25 13:19		MOHAN	OK
29	Q2150-04	TP-48	SAM	05/30/25 13:22		MOHAN	OK
30	Q2150-05	TP-47	SAM	05/30/25 13:24		MOHAN	OK
31	Q2150-06	TP-50	SAM	05/30/25 13:26		MOHAN	OK
32	Q2150-07	TP-51	SAM	05/30/25 13:28		MOHAN	OK
33	CCV81	CCV81	CCV	05/30/25 13:31		MOHAN	OK
34	CCB81	CCB81	CCB	05/30/25 13:37		MOHAN	OK
35	Q2150-08	TP-52	SAM	05/30/25 13:40		MOHAN	OK
36	Q2150-09	TP-54	SAM	05/30/25 13:42		MOHAN	OK
37	Q2150-10	TP-53	SAM	05/30/25 13:44		MOHAN	OK
38	Q2151-01	WC-1	SAM	05/30/25 13:47		MOHAN	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB135958

Review By	MOHAN	Review On	6/2/2025 10:14:34 AM
Supervise By	jaswal	Supervise On	6/2/2025 4:46:20 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85798,MP85799,MP85800,MP85801,MP85802,MP85803		
ICV Standard	MP85804		
CCV Standard	MP85806		
ICSA Standard			
CRI Standard	MP85808		
LCS Standard			
Chk Standard	MP85805,MP85807,MP85809,MP85811		

39	Q2155-01	446	SAM	05/30/25 13:49		MOHAN	OK
40	Q2155-02	447-SOLID	SAM	05/30/25 13:51		MOHAN	OK
41	Q2157-01	291	SAM	05/30/25 13:53		MOHAN	OK
42	Q2150-01L	TP-44L	SD	05/30/25 13:56		MOHAN	OK
43	Q2150-01A	TP-44A	PS	05/30/25 13:58		MOHAN	OK
44	CCV82	CCV82	CCV	05/30/25 14:00		MOHAN	OK
45	CCB82	CCB82	CCB	05/30/25 14:02		MOHAN	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135981

Review By	jaswal	Review On	6/3/2025 4:30:35 PM
Supervise By	mohan	Supervise On	6/3/2025 4:36:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/02/25 12:12		Jaswal	OK
2	S1	S1	CAL2	06/02/25 12:16		Jaswal	OK
3	S2	S2	CAL3	06/02/25 12:21		Jaswal	OK
4	S3	S3	CAL4	06/02/25 12:25		Jaswal	OK
5	S4	S4	CAL5	06/02/25 12:29		Jaswal	OK
6	S5	S5	CAL6	06/02/25 12:33		Jaswal	OK
7	ICV01	ICV01	ICV	06/02/25 12:37		Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/02/25 12:59		Jaswal	OK
9	ICB01	ICB01	ICB	06/02/25 13:03		Jaswal	OK
10	CRI01	CRI01	CRDL	06/02/25 13:08		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/02/25 13:12		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/02/25 13:35		Jaswal	OK
13	ICSA01DL	ICSA01DL	ICSA	06/02/25 13:45		Jaswal	OK
14	ICSAB01DL	ICSAB01DL	ICSAB	06/02/25 13:56		Jaswal	OK
15	CCV01	CCV01	CCV	06/02/25 14:14		Jaswal	OK
16	CCB01	CCB01	CCB	06/02/25 14:28		Jaswal	OK
17	Q2136-05	OR-646-COMP-52	SAM	06/02/25 14:32		Jaswal	OK
18	Q2143-01	ELI-46-36-25-58-53	SAM	06/02/25 14:37		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135981

Review By	jaswal	Review On	6/3/2025 4:30:35 PM
Supervise By	mohan	Supervise On	6/3/2025 4:36:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

19	Q2143-02	ELI-57-43-35-26	SAM	06/02/25 14:41		Jaswal	OK
20	Q2146-04	TP04-MHN-WC	SAM	06/02/25 14:46		Jaswal	OK
21	Q2146-01	TP04-MHN-WC	SAM	06/02/25 14:50		Jaswal	OK
22	Q2146-01DUP	TP04-MHN-WCDUP	DUP	06/02/25 14:54		Jaswal	OK
23	Q2146-01L	TP04-MHN-WCL	SD	06/02/25 14:59		Jaswal	OK
24	Q2146-01MS	TP04-MHN-WCMS	MS	06/02/25 15:03		Jaswal	OK
25	Q2146-01MSD	TP04-MHN-WCMSD	MSD	06/02/25 15:07		Jaswal	OK
26	CCV02	CCV02	CCV	06/02/25 15:11		Jaswal	OK
27	CCB02	CCB02	CCB	06/02/25 15:16		Jaswal	OK
28	Q2146-01A	TP04-MHN-WCA	PS	06/02/25 15:20		Jaswal	OK
29	Q2102-05	LAW-25-0077	SAM	06/02/25 15:24		Jaswal	OK
30	Q2144-01	OILY-DEBRIS-COMP	SAM	06/02/25 15:28		Jaswal	OK
31	Q2144-02	OILY-DEBRIS-COMP	SAM	06/02/25 15:33		Jaswal	OK
32	Q2149-01	FILTER-CAKE	SAM	06/02/25 15:37		Jaswal	OK
33	Q2153-01	TR-04-0592025	SAM	06/02/25 15:42		Jaswal	OK
34	Q2152-01	OK-02-05292025	SAM	06/02/25 15:46		Jaswal	OK
35	Q2152-01DUP	OK-02-05292025DUP	DUP	06/02/25 15:50		Jaswal	OK
36	Q2152-01L	OK-02-05292025L	SD	06/02/25 15:54		Jaswal	OK
37	CCV03	CCV03	CCV	06/02/25 15:58		Jaswal	OK
38	CCB03	CCB03	CCB	06/02/25 16:02		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135981

Review By	jaswal	Review On	6/3/2025 4:30:35 PM
Supervise By	mohan	Supervise On	6/3/2025 4:36:15 PM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

39	Q2152-01MS	OK-02-05292025MS	MS	06/02/25 16:06		Jaswal	OK
40	Q2152-01MSD	OK-02-05292025MSD	MSD	06/02/25 16:10		Jaswal	OK
41	Q2152-01A	OK-02-05292025A	PS	06/02/25 16:14		Jaswal	OK
42	Q2151-01	WC-1	SAM	06/02/25 16:18		Jaswal	OK
43	Q2151-04	WC-1	SAM	06/02/25 16:22		Jaswal	OK
44	Q2151-04DUP	WC-1DUP	DUP	06/02/25 16:27		Jaswal	OK
45	Q2151-04L	WC-1L	SD	06/02/25 16:31		Jaswal	OK
46	Q2151-04MS	WC-1MS	MS	06/02/25 16:36		Jaswal	OK
47	Q2151-04MSD	WC-1MSD	MSD	06/02/25 16:40		Jaswal	OK
48	CCV04	CCV04	CCV	06/02/25 16:49		Jaswal	OK
49	CCB04	CCB04	CCB	06/02/25 16:55		Jaswal	OK
50	Q2151-04A	WC-1A	PS	06/02/25 17:00		Jaswal	OK
51	Q2137-03DL	MOO-25-0149DL	SAM	06/02/25 17:04	Straight 5x for all elements	Jaswal	OK
52	PB168190TB	PB168190TB	MB	06/02/25 17:08		Jaswal	OK
53	PB168214BL	PB168214BL	MB	06/02/25 17:13		Jaswal	OK
54	PB168214BS	PB168214BS	LCS	06/02/25 17:17		Jaswal	OK
55	PB168215BL	PB168215BL	MB	06/02/25 17:21		Jaswal	OK
56	PB168215BS	PB168215BS	LCS	06/02/25 17:25		Jaswal	OK
57	PB168200BL	PB168200BL	MB	06/02/25 17:29		Jaswal	OK
58	PB168200BS	PB168200BS	LCS	06/02/25 17:34		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135981

Review By	jaswal	Review On	6/3/2025 4:30:35 PM
Supervise By	mohan	Supervise On	6/3/2025 4:36:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

59	CCV05	CCV05	CCV	06/02/25 17:38		Jaswal	OK
60	CCB05	CCB05	CCB	06/02/25 17:42		Jaswal	OK
61	Q2150-01	TP-44	SAM	06/02/25 17:47		Jaswal	OK
62	Q2150-02	TP-42	SAM	06/02/25 17:51		Jaswal	OK
63	Q2150-03	TP-39	SAM	06/02/25 17:55		Jaswal	OK
64	Q2150-04	TP-48	SAM	06/02/25 17:59		Jaswal	OK
65	Q2150-05	TP-47	SAM	06/02/25 18:03		Jaswal	OK
66	Q2150-06	TP-50	SAM	06/02/25 18:07		Jaswal	OK
67	Q2150-07	TP-51	SAM	06/02/25 18:12		Jaswal	OK
68	Q2150-08	TP-52	SAM	06/02/25 18:16		Jaswal	OK
69	Q2150-09	TP-54	SAM	06/02/25 18:20		Jaswal	OK
70	CCV06	CCV06	CCV	06/02/25 18:24		Jaswal	OK
71	CCB06	CCB06	CCB	06/02/25 18:28		Jaswal	OK
72	Q2150-10	TP-53	SAM	06/02/25 18:33		Jaswal	OK
73	Q2133-01	MW1	SAM	06/02/25 18:37		Jaswal	OK
74	Q2133-02	MW5	SAM	06/02/25 18:41	Na oversaturated	Jaswal	Dilution
75	Q2133-03	SUPPLY	SAM	06/02/25 18:45	NOT USE	Jaswal	Not Ok
76	Q2133-03DUP	SUPPLYDUP	DUP	06/02/25 18:50	Bad injection	Jaswal	Not Ok
77	Q2133-03L	SUPPLYL	SD	06/02/25 18:54	NOT USE	Jaswal	Not Ok
78	Q2133-03MS	SUPPLYMS	MS	06/02/25 18:59	Bad injection	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135981

Review By	jaswal	Review On	6/3/2025 4:30:35 PM
Supervise By	mohan	Supervise On	6/3/2025 4:36:15 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546		
ICV Standard	MP85553		
CCV Standard	MP85556		
ICSA Standard	MP85554,MP85555		
CRI Standard	MP85552		
LCS Standard			
Chk Standard	MP85557,MP85558		

79	Q2133-03MSD	SUPPLYMSD	MSD	06/02/25 19:03	NOT USE	Jaswal	Not Ok
80	Q2133-03A	SUPPLYA	PS	06/02/25 19:07	NOT USE	Jaswal	Not Ok
81	CCV07	CCV07	CCV	06/02/25 19:11		Jaswal	OK
82	CCB07	CCB07	CCB	06/02/25 19:15		Jaswal	OK
83	Q2134-01	MW10	SAM	06/02/25 19:19	Mn high	Jaswal	Dilution
84	Q2101-01	TP-1-MHE	SAM	06/02/25 19:24	NOT USE	Jaswal	Not Ok
85	Q2101-01DUP	TP-1-MHEDUP	DUP	06/02/25 19:28	Need to Confirm	Jaswal	Not Ok
86	Q2101-01L	TP-1-MHEL	SD	06/02/25 19:32	NOT USE	Jaswal	Not Ok
87	Q2101-01MS	TP-1-MHEMS	MS	06/02/25 19:36	MS/MSD fail for all elements	Jaswal	Not Ok
88	Q2101-01MSD	TP-1-MHEMSD	MSD	06/02/25 19:40	MS/MSD fail for all elements	Jaswal	Not Ok
89	Q2101-01A	TP-1-MHEA	PS	06/02/25 19:45	NOT USE	Jaswal	Not Ok
90	Q2159-01	TP05-MHO-WC	SAM	06/02/25 19:49		Jaswal	OK
91	Q2160-01	TP04-MHG-WC	SAM	06/02/25 19:53		Jaswal	OK
92	CCV08	CCV08	CCV	06/02/25 19:57		Jaswal	OK
93	CCB08	CCB08	CCB	06/02/25 20:01		Jaswal	OK
94	Q2172-01	TP06-MHQ	SAM	06/02/25 20:06	NOT USE	Jaswal	Not Ok
95	Q2172-01DUP	TP06-MHQDUP	DUP	06/02/25 20:10	NOT USE	Jaswal	Not Ok
96	Q2172-01L	TP06-MHQL	SD	06/02/25 20:14	NOT USE	Jaswal	Not Ok
97	Q2172-01MS	TP06-MHQMS	MS	06/02/25 20:18	Bad injection	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135981

Review By	jaswal	Review On	6/3/2025 4:30:35 PM
Supervise By	mohan	Supervise On	6/3/2025 4:36:15 PM

STD. NAME	STD REF.#
ICAL Standard	MP85545,MP85552,MP85549,MP85548,MP85547,MP85546
ICV Standard	MP85553
CCV Standard	MP85556
ICSA Standard	MP85554,MP85555
CRI Standard	MP85552
LCS Standard	
Chk Standard	MP85557,MP85558

98	Q2172-01MSD	TP06-MHQMSD	MSD	06/02/25 20:23	NOT USE	Jaswal	Not Ok
99	Q2172-01A	TP06-MHQA	PS	06/02/25 20:27	NOT USE	Jaswal	Not Ok
100	Q2173-01	OR-400-CF-402B-CO	SAM	06/02/25 20:32		Jaswal	OK
101	Q2173-07	OR-400-CF-402B-CO	SAM	06/02/25 20:36	Bad injection	Jaswal	Not Ok
102	Q2173-13	OR-400-CF-402B-CO	SAM	06/02/25 20:41	Bad injection	Jaswal	Not Ok
103	CCV09	CCV09	CCV	06/02/25 20:45		Jaswal	OK
104	CCB09	CCB09	CCB	06/02/25 20:49		Jaswal	OK
105	PB168236BL	PB168236BL	MB	06/02/25 20:54	NOT USE	Jaswal	Not Ok
106	PB168236BS	PB168236BS	LCS	06/02/25 20:58	NOT USE	Jaswal	Not Ok
107	Q2175-02	52725	SAM	06/02/25 21:02	NOT USE	Jaswal	Not Ok
108	CCV10	CCV10	CCV	06/02/25 21:06	CCV fail for all aparemters	Jaswal	OK
109	CCB10	CCB10	CCB	06/02/25 21:11		Jaswal	OK