



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: CDM Smith SDG No.: Q2150
Contract: CAMP02 Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	95	70 - 130	CV	05/30/2025	12:25	LB135958
CRI01	Aluminum	102	100	102	65 - 135	P	06/02/2025	13:08	LB135981
	Antimony	48.0	50.0	96	65 - 135	P	06/02/2025	13:08	LB135981
	Arsenic	20.4	20.0	102	65 - 135	P	06/02/2025	13:08	LB135981
	Barium	92.5	100	92	65 - 135	P	06/02/2025	13:08	LB135981
	Beryllium	5.91	6.0	99	65 - 135	P	06/02/2025	13:08	LB135981
	Cadmium	5.90	6.0	98	65 - 135	P	06/02/2025	13:08	LB135981
	Calcium	1910	2000	95	65 - 135	P	06/02/2025	13:08	LB135981
	Chromium	9.95	10.0	100	65 - 135	P	06/02/2025	13:08	LB135981
	Cobalt	28.9	30.0	96	65 - 135	P	06/02/2025	13:08	LB135981
	Copper	21.2	20.0	106	65 - 135	P	06/02/2025	13:08	LB135981
	Iron	106	100	106	65 - 135	P	06/02/2025	13:08	LB135981
	Lead	13.1	12.0	109	65 - 135	P	06/02/2025	13:08	LB135981
	Magnesium	1930	2000	96	65 - 135	P	06/02/2025	13:08	LB135981
	Manganese	20.1	20.0	101	65 - 135	P	06/02/2025	13:08	LB135981
	Nickel	39.5	40.0	99	65 - 135	P	06/02/2025	13:08	LB135981
	Potassium	1740	2000	87	65 - 135	P	06/02/2025	13:08	LB135981
	Selenium	19.1	20.0	96	65 - 135	P	06/02/2025	13:08	LB135981
	Silver	9.30	10.0	93	65 - 135	P	06/02/2025	13:08	LB135981
	Sodium	1690	2000	85	65 - 135	P	06/02/2025	13:08	LB135981
	Thallium	35.9	40.0	90	65 - 135	P	06/02/2025	13:08	LB135981
	Vanadium	36.5	40.0	91	65 - 135	P	06/02/2025	13:08	LB135981
	Zinc	40.4	40.0	101	65 - 135	P	06/02/2025	13:08	LB135981