

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2150
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2150 **SAS No.:** Q2150

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168215BS							
Aluminum	mg/Kg	100	99.8		100	80 - 120	P
Antimony	mg/Kg	40.0	42.0		105	80 - 120	P
Arsenic	mg/Kg	40.0	41.0		102	80 - 120	P
Barium	mg/Kg	10.0	9.37		94	80 - 120	P
Beryllium	mg/Kg	10.0	10.1		101	80 - 120	P
Cadmium	mg/Kg	10.0	10.0		100	80 - 120	P
Calcium	mg/Kg	50.0	50.4	J	101	80 - 120	P
Chromium	mg/Kg	20.0	20.8		104	80 - 120	P
Cobalt	mg/Kg	10.0	10.0		100	80 - 120	P
Copper	mg/Kg	15.0	15.9		106	80 - 120	P
Iron	mg/Kg	150	152		101	80 - 120	P
Lead	mg/Kg	50.0	48.9		98	80 - 120	P
Magnesium	mg/Kg	100	97.4	J	97	80 - 120	P
Manganese	mg/Kg	10.0	9.85		98	80 - 120	P
Nickel	mg/Kg	25.0	25.4		102	80 - 120	P
Potassium	mg/Kg	500	484		97	80 - 120	P
Selenium	mg/Kg	100	104		104	80 - 120	P
Silver	mg/Kg	3.8	3.77		99	80 - 120	P
Sodium	mg/Kg	150	138		92	80 - 120	P
Thallium	mg/Kg	100	99.5		100	80 - 120	P
Vanadium	mg/Kg	15.0	14.7		98	80 - 120	P
Zinc	mg/Kg	10.0	10.3		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2150
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2150 **SAS No.:** Q2150

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168222BS Mercury	mg/Kg	0.25	0.21		86	80 - 120	CV