

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052125\
 Data File : FF015878.D
 Signal(s) : FID2B.ch
 Acq On : 21 May 2025 12:16
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 FID_F
 ClientSampleId :
 5 TRPH STD

Integration File: autoint1.e
 Quant Time: May 21 12:31:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052125.M
 Quant Title :
 QLast Update : Wed May 21 12:31:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.021	628555	5.146 ug/ml
Target Compounds			
1) N-OCTANE	2.033	571635	4.689 ug/ml
2) N-DECANE	4.571	505250	4.262 ug/ml
3) N-DODECANE	6.741	561760	4.465 ug/ml
4) N-TETRADECANE	8.583	531455	4.236 ug/ml
5) N-HEXADECANE	10.184	611741	4.664 ug/ml
6) N-OCTADECANE	11.620	676173	4.921 ug/ml
7) N-EICOSANE	12.927	721081	5.068 ug/ml
8) N-DOCOSANE	14.123	711955	5.125 ug/ml
10) N-TETRACOSANE	15.225	723436	5.190 ug/ml
11) N-HEXACOSANE	16.246	721950	5.205 ug/ml
12) N-OCTACOSANE	17.197	726562	5.285 ug/ml
13) N-TRIACONTANE	18.083	742337	5.332 ug/ml
14) N-DOTRIACONTANE	18.916	725714	5.375 ug/ml
15) N-TETRATRIACONTANE	19.700	649413	5.266 ug/ml
16) N-HEXATRIACONTANE	20.441	555212	5.112 ug/ml
17) N-OCTATRIACONTANE	21.177	456710	4.813 ug/ml
18) N-TETRACONTANE	22.092	392391	4.619 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF052125\
Data File : FF015878.D
Signal(s) : FID2B.ch
Acq On : 21 May 2025 12:16
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
FID_F
ClientSampleId :
5 TRPH STD

Integration File: autoint1.e
Quant Time: May 21 12:31:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF052125.M
Quant Title :
QLast Update : Wed May 21 12:31:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

