

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_S

Calibration Date(s): 06/02/2025 06/02/2025

Calibration Times: 09:13 10:49

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS030437.D</u>	CF 500 =	<u>PS030438.D</u>		
CF 750 =		<u>PS030439.D</u>	CF 1000 =	<u>PS030440.D</u>	CF 1500 =	<u>PS030441.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	26654600000	22537700000	21523600000	21039200000	19992000000	22349400000	12
2,4,5-TP(Silvex)	30373300000	24973400000	23533400000	22714000000	21490700000	24616900000	14
2,4-D	5421790000	4357790000	4143530000	4048250000	3895280000	4373330000	14
2,4-DB	3958090000	3446260000	3373160000	3385980000	3356080000	3503910000	7
2,4-DCAA	5145720000	4057020000	3830810000	3692630000	3535940000	4052420000	16
DICAMBA	20964400000	17146300000	16619900000	16172600000	15526100000	17285900000	12
DICHLORPROP	5206190000	4130820000	3894460000	3783800000	3645610000	4132180000	15
Dinoseb	23156800000	18828100000	17672200000	17195900000	16380800000	18646800000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_S Calibration Date(s): 06/02/2025 06/02/2025
Calibration Times: 09:13 10:49

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS030437.D</u>	CF 500 =	<u>PS030438.D</u>		
CF 750 =		<u>PS030439.D</u>	CF 1000 =	<u>PS030440.D</u>	CF 1500 =	<u>PS030441.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	16810500000	14903600000	14395300000	14103000000	13612500000	14765000000	8
2,4,5-TP(Silvex)	18734600000	16594300000	15960100000	15600700000	14942200000	16366400000	9
2,4-D	23900000000	20282500000	19630600000	19443300000	19164000000	20484100000	10
2,4-DB	15607200000	13423500000	13160500000	13285600000	13243200000	13744000000	8
2,4-DCAA	14740500000	12572600000	12209200000	12102100000	11998800000	12724600000	9
DICAMBA	86590200000	78255800000	77404500000	77103100000	76145700000	79099800000	5
DICHLORPROP	22047700000	18922400000	18012700000	17785400000	17504500000	18854500000	10
Dinoseb	13001900000	11375000000	10900700000	10837000000	10570200000	11337000000	9