

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/02/2025 06/02/2025

Client Sample No.: CCAL01 Date Analyzed: 06/02/2025

Lab Sample No.: HSTDCCC750 Data File : PS030456.D Time Analyzed: 18:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	8.884	8.785	8.985	711.700	712.500	-0.1
2,4,5-TP(Silvex)	8.608	8.509	8.709	699.100	712.500	-1.9
2,4-D	7.785	7.685	7.885	691.520	705.000	-1.9
2,4-DB	9.433	9.334	9.534	735.550	712.500	3.2
2,4-DCAA	6.732	6.633	6.833	721.540	750.000	-3.8
DICAMBA	6.904	6.803	7.003	690.790	705.000	-2.0
DICHLORPROP	7.570	7.470	7.670	681.540	705.000	-3.3
Dinoseb	10.573	10.474	10.674	677.420	705.000	-3.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/02/2025 06/02/2025

Client Sample No.: CCAL01 Date Analyzed: 06/02/2025

Lab Sample No.: HSTDCCC750 Data File : PS030456.D Time Analyzed: 18:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.743	9.645	9.845	712.480	712.500	0.0
2,4,5-TP(Silvex)	9.343	9.245	9.445	710.870	712.500	-0.2
2,4-D	8.475	8.376	8.576	694.340	705.000	-1.5
2,4-DB	10.296	10.197	10.397	698.450	712.500	-2.0
2,4-DCAA	7.301	7.202	7.402	741.240	750.000	-1.2
DICAMBA	7.482	7.383	7.583	708.370	705.000	0.5
DICHLORPROP	8.167	8.068	8.268	695.540	705.000	-1.3
Dinoseb	10.664	10.565	10.765	683.460	705.000	-3.1