

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL05 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030489.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.694	9.597	9.797	707.360	712.500	-0.7
2,4,5-TP(Silvex)	9.400	9.304	9.504	701.500	712.500	-1.5
2,4-D	8.508	8.412	8.612	688.650	705.000	-2.3
2,4-DB	10.275	10.179	10.379	711.140	712.500	-0.2
2,4-DCAA	7.366	7.269	7.469	719.780	750.000	-4.0
DICAMBA	7.558	7.461	7.661	694.210	705.000	-1.5
DICHLORPROP	8.275	8.179	8.379	678.650	705.000	-3.7
Dinoseb	11.498	11.402	11.602	681.220	705.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL05 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030489.D Time Analyzed: 16:50

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.367	10.268	10.468	697.940	712.500	-2.0
2,4,5-TP(Silvex)	9.940	9.842	10.042	690.320	712.500	-3.1
2,4-D	9.033	8.934	9.134	680.340	705.000	-3.5
2,4-DB	10.934	10.836	11.036	685.670	712.500	-3.8
2,4-DCAA	7.773	7.674	7.874	712.640	750.000	-5.0
DICAMBA	7.977	7.877	8.077	688.840	705.000	-2.3
DICHLORPROP	8.697	8.598	8.798	670.150	705.000	-4.9
Dinoseb	11.320	11.221	11.421	676.820	705.000	-4.0