

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL09 Date Analyzed: 06/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS030606.D Time Analyzed: 20:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.683	9.597	9.797	708.690	712.500	-0.5
2,4,5-TP(Silvex)	9.388	9.304	9.504	734.910	712.500	3.1
2,4-D	8.498	8.412	8.612	699.570	705.000	-0.8
2,4-DB	10.262	10.179	10.379	662.060	712.500	-7.1
2,4-DCAA	7.358	7.269	7.469	695.690	750.000	-7.2
DICAMBA	7.550	7.461	7.661	736.770	705.000	4.5
DICHLORPROP	8.265	8.179	8.379	709.770	705.000	0.7
Dinoseb	11.483	11.402	11.602	730.610	705.000	3.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL09 Date Analyzed: 06/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS030606.D Time Analyzed: 20:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.360	10.268	10.468	734.330	712.500	3.1
2,4,5-TP(Silvex)	9.934	9.842	10.042	739.790	712.500	3.8
2,4-D	9.027	8.934	9.134	724.670	705.000	2.8
2,4-DB	10.927	10.836	11.036	717.750	712.500	0.7
2,4-DCAA	7.769	7.674	7.874	736.310	750.000	-1.8
DICAMBA	7.972	7.877	8.077	732.020	705.000	3.8
DICHLORPROP	8.692	8.598	8.798	712.380	705.000	1.0
Dinoseb	11.313	11.221	11.421	743.840	705.000	5.5