

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2150</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS030577.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2130-01MS (Column 1)	TP-3MS											
	DICAMBA	184.4	0	82.9	ug/Kg	45				10	112	
	DICHLORPROP	184.4	0	93.1	ug/Kg	50				10	113	
	2,4-D	184.4	0	99.5	ug/Kg	54				10	144	
	2,4,5-TP(Silvex)	184.4	0	88.2	ug/Kg	48				10	114	
	2,4,5-T	184.4	0	79.9	ug/Kg	43				10	115	
	2,4-DB	184.4	0	46.7	ug/Kg	25				10	140	
Client Sample ID: Q2130-01MS (Column 2)	Dinoseb	184.4	0	16.3	ug/Kg	9	*			10	118	
	TP-3MS											
	DICAMBA	184.4	0	78.3	ug/Kg	42				10	112	
	DICHLORPROP	184.4	0	91.5	ug/Kg	50				10	113	
	2,4-D	184.4	0	84.7	ug/Kg	46				10	144	
	2,4,5-TP(Silvex)	184.4	0	92.9	ug/Kg	50				10	114	
	2,4,5-T	184.4	0	82.2	ug/Kg	45				10	115	
	2,4-DB	184.4	0	75.1	ug/Kg	41				10	140	
	Dinoseb	184.4	0	20.3	ug/Kg	11				10	118	