

Contract:	<u>CAMP02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2150</u>	SAS No.:	<u>Q2150</u>	SDG NO.:	<u>Q2150</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>06/02/2025</u>	<u>06/02/2025</u>		
		Calibration Times:		<u>09:13</u>	<u>10:49</u>		

LAB FILE ID:		RT 200 =	<u>PS030437.D</u>	RT 500 =	<u>PS030438.D</u>
RT 750 =	PS030439.D	RT 1000 =	PS030440.D	RT 1500 =	PS030441.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Instrument ID: ECD_S Calibration Date(s): 06/02/2025 06/02/2025

Calibration Times: 09:13 10:49

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS030437.D</u>	RT 500 = <u>PS030438.D</u>
	RT 750 = <u>PS030439.D</u>	RT 1000 = <u>PS030440.D</u>
		RT 1500 = <u>PS030441.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW	
							FROM	TO
2,4,5-T	9.74	9.74	9.75	9.74	9.74	9.74	9.64	9.84
2,4,5-TP(Silvex)	9.34	9.34	9.35	9.34	9.34	9.34	9.24	9.44
2,4-D	8.48	8.48	8.48	8.48	8.48	8.48	8.38	8.58
2,4-DB	10.30	10.30	10.30	10.30	10.30	10.30	10.20	10.40
2,4-DCAA	7.30	7.30	7.30	7.30	7.30	7.30	7.20	7.40
DICAMBA	7.48	7.48	7.48	7.48	7.48	7.48	7.38	7.58
DICHLORPROP	8.17	8.17	8.17	8.17	8.17	8.17	8.07	8.27
Dinoseb	10.66	10.66	10.67	10.67	10.66	10.66	10.56	10.76