

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 06/02/2025 Initial Calibration Date(s): 06/02/2025 06/02/2025

Continuing Calib Time: 18:10 Initial Calibration Time(s): 09:13 10:49

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 6.90       | 6.90      | 6.80      | 7.00  | 0.00       |
| 2,4-DCAA         | 6.73       | 6.73      | 6.63      | 6.83  | 0.00       |
| DICHLORPROP      | 7.57       | 7.57      | 7.47      | 7.67  | 0.00       |
| 2,4-D            | 7.79       | 7.79      | 7.69      | 7.89  | 0.01       |
| 2,4,5-TP(Silvex) | 8.61       | 8.61      | 8.51      | 8.71  | 0.00       |
| 2,4,5-T          | 8.88       | 8.89      | 8.79      | 8.99  | 0.01       |
| 2,4-DB           | 9.43       | 9.43      | 9.33      | 9.53  | 0.00       |
| Dinoseb          | 10.57      | 10.57     | 10.47     | 10.67 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 06/02/2025 Initial Calibration Date(s): 06/02/2025 06/02/2025

Continuing Calib Time: 18:10 Initial Calibration Time(s): 09:13 10:49

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.48       | 7.48      | 7.38      | 7.58  | 0.00       |
| 2,4-DCAA         | 7.30       | 7.30      | 7.20      | 7.40  | 0.00       |
| DICHLORPROP      | 8.17       | 8.17      | 8.07      | 8.27  | 0.00       |
| 2,4-D            | 8.48       | 8.48      | 8.38      | 8.58  | 0.00       |
| 2,4,5-TP(Silvex) | 9.34       | 9.35      | 9.25      | 9.45  | 0.01       |
| 2,4,5-T          | 9.74       | 9.75      | 9.65      | 9.85  | 0.01       |
| 2,4-DB           | 10.30      | 10.30     | 10.20     | 10.40 | 0.00       |
| Dinoseb          | 10.66      | 10.67     | 10.57     | 10.77 | 0.01       |