

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 06/03/2025 Initial Calibration Date(s): 06/02/2025 06/02/2025

Continuing Calib Time: 00:59 Initial Calibration Time(s): 09:13 10:49

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	6.90	6.90	6.80	7.00	0.00
2,4-DCAA	6.73	6.73	6.63	6.83	0.00
DICHLORPROP	7.57	7.57	7.47	7.67	0.00
2,4-D	7.78	7.79	7.69	7.89	0.01
2,4,5-TP(Silvex)	8.61	8.61	8.51	8.71	0.00
2,4,5-T	8.88	8.89	8.79	8.99	0.01
2,4-DB	9.43	9.43	9.33	9.53	0.00
Dinoseb	10.57	10.57	10.47	10.67	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM **Case No.:** Q2150 **SAS No.:** Q2150 **SDG NO.:** Q2150

Continuing Calib Date: 06/03/2025 **Initial Calibration Date(s):** 06/02/2025 06/02/2025

Continuing Calib Time: 00:59 **Initial Calibration Time(s):** 09:13 10:49

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.48	7.48	7.38	7.58	0.00
2,4-DCAA	7.30	7.30	7.20	7.40	0.00
DICHLORPROP	8.17	8.17	8.07	8.27	0.00
2,4-D	8.48	8.48	8.38	8.58	0.00
2,4,5-TP(Silvex)	9.34	9.35	9.25	9.45	0.01
2,4,5-T	9.74	9.75	9.65	9.85	0.01
2,4-DB	10.30	10.30	10.20	10.40	0.00
Dinoseb	10.66	10.67	10.57	10.77	0.01