

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 06/04/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 14:23 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|------------------|------------|-----------|------------------------|-------|------------|
| DICAMBA          | 7.56       | 7.56      | 7.46                   | 7.66  | 0.00       |
| 2,4-DCAA         | 7.37       | 7.37      | 7.27                   | 7.47  | 0.00       |
| DICHLORPROP      | 8.28       | 8.28      | 8.18                   | 8.38  | 0.00       |
| 2,4-D            | 8.51       | 8.51      | 8.41                   | 8.61  | 0.00       |
| 2,4,5-TP(Silvex) | 9.40       | 9.40      | 9.30                   | 9.50  | 0.00       |
| 2,4,5-T          | 9.70       | 9.70      | 9.60                   | 9.80  | 0.00       |
| 2,4-DB           | 10.28      | 10.28     | 10.18                  | 10.38 | 0.00       |
| Dinoseb          | 11.50      | 11.50     | 11.40                  | 11.60 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

Continuing Calib Date: 06/04/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 14:23 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP2 ID: 0.32 (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|------------------|------------|-----------|-----------|-------|------------|
|                  |            |           | FROM      | TO    |            |
| DICAMBA          | 7.98       | 7.98      | 7.88      | 8.08  | 0.00       |
| 2,4-DCAA         | 7.77       | 7.77      | 7.67      | 7.87  | 0.00       |
| DICHLORPROP      | 8.70       | 8.70      | 8.60      | 8.80  | 0.00       |
| 2,4-D            | 9.03       | 9.03      | 8.93      | 9.13  | 0.00       |
| 2,4,5-TP(Silvex) | 9.94       | 9.94      | 9.84      | 10.04 | 0.00       |
| 2,4,5-T          | 10.37      | 10.37     | 10.27     | 10.47 | 0.00       |
| 2,4-DB           | 10.94      | 10.94     | 10.84     | 11.04 | 0.00       |
| Dinoseb          | 11.32      | 11.32     | 11.22     | 11.42 | 0.00       |