

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/02/2025 06/02/2025

Client Sample No.: CCAL02 Date Analyzed: 06/02/2025

Lab Sample No.: HSTDCCC750 Data File : PS030467.D Time Analyzed: 22:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	8.883	8.785	8.985	714.810	712.500	0.3
2,4,5-TP(Silvex)	8.608	8.509	8.709	700.540	712.500	-1.7
2,4-D	7.785	7.685	7.885	693.760	705.000	-1.6
2,4-DB	9.432	9.334	9.534	737.030	712.500	3.4
2,4-DCAA	6.732	6.633	6.833	723.490	750.000	-3.5
DICAMBA	6.903	6.803	7.003	692.130	705.000	-1.8
DICHLORPROP	7.569	7.470	7.670	682.100	705.000	-3.2
Dinoseb	10.571	10.474	10.674	682.660	705.000	-3.2

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/02/2025 06/02/2025

Client Sample No.: CCAL02 Date Analyzed: 06/02/2025

Lab Sample No.: HSTDCCC750 Data File : PS030467.D Time Analyzed: 22:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.743	9.645	9.845	711.940	712.500	-0.1
2,4,5-TP(Silvex)	9.343	9.245	9.445	709.200	712.500	-0.5
2,4-D	8.475	8.376	8.576	695.710	705.000	-1.3
2,4-DB	10.295	10.197	10.397	701.530	712.500	-1.5
2,4-DCAA	7.301	7.202	7.402	738.810	750.000	-1.5
DICAMBA	7.482	7.383	7.583	706.810	705.000	0.3
DICHLORPROP	8.167	8.068	8.268	702.350	705.000	-0.4
Dinoseb	10.663	10.565	10.765	685.830	705.000	-2.7