

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/02/2025 06/02/2025

Client Sample No.: CCAL03 Date Analyzed: 06/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS030473.D Time Analyzed: 00:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	8.884	8.785	8.985	719.320	712.500	1.0
2,4,5-TP(Silvex)	8.607	8.509	8.709	703.280	712.500	-1.3
2,4-D	7.784	7.685	7.885	699.740	705.000	-0.7
2,4-DB	9.432	9.334	9.534	745.590	712.500	4.6
2,4-DCAA	6.733	6.633	6.833	724.260	750.000	-3.4
DICAMBA	6.903	6.803	7.003	693.600	705.000	-1.6
DICHLORPROP	7.569	7.470	7.670	684.280	705.000	-2.9
Dinoseb	10.572	10.474	10.674	687.260	705.000	-2.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/02/2025 06/02/2025

Client Sample No.: CCAL03 Date Analyzed: 06/03/2025

Lab Sample No.: HSTDCCC750 Data File : PS030473.D Time Analyzed: 00:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.743	9.645	9.845	724.500	712.500	1.7
2,4,5-TP(Silvex)	9.343	9.245	9.445	718.510	712.500	0.8
2,4-D	8.475	8.376	8.576	705.970	705.000	0.1
2,4-DB	10.295	10.197	10.397	712.370	712.500	0.0
2,4-DCAA	7.301	7.202	7.402	746.250	750.000	-0.5
DICAMBA	7.481	7.383	7.583	714.410	705.000	1.3
DICHLORPROP	8.166	8.068	8.268	700.070	705.000	-0.7
Dinoseb	10.663	10.565	10.765	700.530	705.000	-0.6