

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030483.D Time Analyzed: 14:23

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.697	9.597	9.797	708.550	712.500	-0.6
2,4,5-TP(Silvex)	9.401	9.304	9.504	703.720	712.500	-1.2
2,4-D	8.509	8.412	8.612	686.470	705.000	-2.6
2,4-DB	10.276	10.179	10.379	703.340	712.500	-1.3
2,4-DCAA	7.367	7.269	7.469	725.390	750.000	-3.3
DICAMBA	7.560	7.461	7.661	694.190	705.000	-1.5
DICHLORPROP	8.276	8.179	8.379	677.610	705.000	-3.9
Dinoseb	11.500	11.402	11.602	693.920	705.000	-1.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/04/2025

Lab Sample No.: HSTDCCC750 Data File : PS030483.D Time Analyzed: 14:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.367	10.268	10.468	690.040	712.500	-3.2
2,4,5-TP(Silvex)	9.941	9.842	10.042	686.650	712.500	-3.6
2,4-D	9.033	8.934	9.134	674.550	705.000	-4.3
2,4-DB	10.935	10.836	11.036	678.160	712.500	-4.8
2,4-DCAA	7.773	7.674	7.874	700.400	750.000	-6.6
DICAMBA	7.977	7.877	8.077	680.500	705.000	-3.5
DICHLORPROP	8.697	8.598	8.798	672.350	705.000	-4.6
Dinoseb	11.321	11.221	11.421	684.430	705.000	-2.9