

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL07 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030580.D Time Analyzed: 19:53

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.684	9.597	9.797	698.110	712.500	-2.0
2,4,5-TP(Silvex)	9.390	9.304	9.504	724.120	712.500	1.6
2,4-D	8.499	8.412	8.612	689.840	705.000	-2.2
2,4-DB	10.264	10.179	10.379	649.610	712.500	-8.8
2,4-DCAA	7.360	7.269	7.469	732.930	750.000	-2.3
DICAMBA	7.552	7.461	7.661	729.450	705.000	3.5
DICHLORPROP	8.268	8.179	8.379	700.780	705.000	-0.6
Dinoseb	11.486	11.402	11.602	716.830	705.000	1.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL07 Date Analyzed: 06/09/2025

Lab Sample No.: HSTDCCC750 Data File : PS030580.D Time Analyzed: 19:53

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.361	10.268	10.468	710.000	712.500	-0.4
2,4,5-TP(Silvex)	9.935	9.842	10.042	715.970	712.500	0.5
2,4-D	9.028	8.934	9.134	701.870	705.000	-0.4
2,4-DB	10.928	10.836	11.036	686.990	712.500	-3.6
2,4-DCAA	7.769	7.674	7.874	723.920	750.000	-3.5
DICAMBA	7.973	7.877	8.077	709.970	705.000	0.7
DICHLORPROP	8.693	8.598	8.798	701.790	705.000	-0.5
Dinoseb	11.314	11.221	11.421	712.940	705.000	1.1