

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL08 Date Analyzed: 06/11/2025

Lab Sample No.: HSTDCCC750 Data File : PS030621.D Time Analyzed: 09:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.680	9.597	9.797	700.660	712.500	-1.7
2,4,5-TP(Silvex)	9.386	9.304	9.504	736.560	712.500	3.4
2,4-D	8.496	8.412	8.612	691.670	705.000	-1.9
2,4-DB	10.259	10.179	10.379	643.460	712.500	-9.7
2,4-DCAA	7.356	7.269	7.469	737.450	750.000	-1.7
DICAMBA	7.549	7.461	7.661	744.120	705.000	5.5
DICHLORPROP	8.264	8.179	8.379	713.190	705.000	1.2
Dinoseb	11.481	11.402	11.602	718.070	705.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL08 Date Analyzed: 06/11/2025

Lab Sample No.: HSTDCCC750 Data File : PS030621.D Time Analyzed: 09:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.359	10.268	10.468	750.020	712.500	5.3
2,4,5-TP(Silvex)	9.932	9.842	10.042	754.780	712.500	5.9
2,4-D	9.026	8.934	9.134	740.500	705.000	5.0
2,4-DB	10.925	10.836	11.036	732.100	712.500	2.8
2,4-DCAA	7.768	7.674	7.874	751.350	750.000	0.2
DICAMBA	7.971	7.877	8.077	750.620	705.000	6.5
DICHLORPROP	8.691	8.598	8.798	728.610	705.000	3.3
Dinoseb	11.311	11.221	11.421	752.930	705.000	6.8