

Surrogate Summary

SDG No.: **Q2150**

Client: **CDM Smith**

Analytical Method: **8151A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS030436.D	PIBLK-PS030436.D	2,4-DCAA	1	500	465	93		61	136
		2,4-DCAA	2	500	470	94		61	136
I.BLK-PS030455.D	PIBLK-PS030455.D	2,4-DCAA	1	500	467	93		61	136
		2,4-DCAA	2	500	483	97		61	136
Q2150-01	TP-44	2,4-DCAA	1	500	173	35		10	141
		2,4-DCAA	2	500	167	33		10	141
Q2150-02	TP-42	2,4-DCAA	1	500	275	55		10	141
		2,4-DCAA	2	500	280	56		10	141
Q2150-03	TP-39	2,4-DCAA	1	500	267	53		10	141
		2,4-DCAA	2	500	255	51		10	141
Q2150-04	TP-48	2,4-DCAA	1	500	239	48		10	141
		2,4-DCAA	2	500	221	44		10	141
Q2150-05	TP-47	2,4-DCAA	1	500	254	51		10	141
		2,4-DCAA	2	500	237	47		10	141
I.BLK-PS030466.D	PIBLK-PS030466.D	2,4-DCAA	1	500	477	95		61	136
		2,4-DCAA	2	500	485	97		61	136
Q2150-07	TP-51	2,4-DCAA	1	500	247	49		10	141
		2,4-DCAA	2	500	244	49		10	141
Q2150-08	TP-52	2,4-DCAA	1	500	173	35		10	141
		2,4-DCAA	2	500	165	33		10	141
Q2150-09	TP-54	2,4-DCAA	1	500	331	66		10	141
		2,4-DCAA	2	500	320	64		10	141
Q2150-10	TP-53	2,4-DCAA	1	500	292	58		10	141
		2,4-DCAA	2	500	285	57		10	141
I.BLK-PS030472.D	PIBLK-PS030472.D	2,4-DCAA	1	500	476	95		61	136
		2,4-DCAA	2	500	490	98		61	136
I.BLK-PS030475.D	PIBLK-PS030475.D	2,4-DCAA	1	500	414	83		61	136
		2,4-DCAA	2	500	445	89		61	136
I.BLK-PS030482.D	PIBLK-PS030482.D	2,4-DCAA	1	500	444	89		61	136
		2,4-DCAA	2	500	449	90		61	136
PB168207BL	PB168207BL	2,4-DCAA	1	500	436	87		10	141
		2,4-DCAA	2	500	430	86		10	141
PB168207BS	PB168207BS	2,4-DCAA	1	500	527	105		10	141
		2,4-DCAA	2	500	493	99		10	141
I.BLK-PS030488.D	PIBLK-PS030488.D	2,4-DCAA	1	500	436	87		61	136
		2,4-DCAA	2	500	451	90		61	136
I.BLK-PS030574.D	PIBLK-PS030574.D	2,4-DCAA	1	500	437	87		61	136
		2,4-DCAA	2	500	454	91		61	136
Q2130-01MS	TP-3MS	2,4-DCAA	1	500	252	50		10	141
		2,4-DCAA	2	500	210	42		10	141
Q2130-01MSD	TP-3MSD	2,4-DCAA	1	500	250	50		10	141

Surrogate Summary

SDG No.: Q2150

Client: CDM Smith

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2130-01MSD	TP-3MSD	2,4-DCAA	2	500	206	41		10	141
I.BLK-PS030579.D	PIBLK-PS030579.D	2,4-DCAA	1	500	450	90		61	136
		2,4-DCAA	2	500	466	93		61	136
I.BLK-PS030620.D	PIBLK-PS030620.D	2,4-DCAA	1	500	449	90		61	136
		2,4-DCAA	2	500	487	97		61	136
Q2150-06	TP-50	2,4-DCAA	1	500	540	108		10	141
		2,4-DCAA	2	500	597	119		10	141
I.BLK-PS030629.D	PIBLK-PS030629.D	2,4-DCAA	1	500	452	90		61	136
		2,4-DCAA	2	500	492	98		61	136