

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2150</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS030578.D</u>

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q2130-01MSD (Column 1)	TP-3MSD											
	DICAMBA	184.1	0	82.5	ug/Kg	45		0		10	112	20
	DICHLORPROP	184.1	0	92.4	ug/Kg	50		0		10	113	20
	2,4-D	184.1	0	99.3	ug/Kg	54		0		10	144	20
	2,4,5-TP(Silvex)	184.1	0	87.6	ug/Kg	48		0		10	114	20
	2,4,5-T	184.1	0	80.3	ug/Kg	44		2		10	115	20
	2,4-DB	184.1	0	44.9	ug/Kg	24		4		10	140	20
Client Sample ID: Q2130-01MSD (Column 2)	Dinoseb	184.1	0	16.3	ug/Kg	9	*	0		10	118	20
	TP-3MSD											
	DICAMBA	184.1	0	78.1	ug/Kg	42		0		10	112	20
	DICHLORPROP	184.1	0	91.1	ug/Kg	49		2		10	113	20
	2,4-D	184.1	0	94.6	ug/Kg	51		10		10	144	20
	2,4,5-TP(Silvex)	184.1	0	92.0	ug/Kg	50		0		10	114	20
	2,4,5-T	184.1	0	81.6	ug/Kg	44		2		10	115	20
	2,4-DB	184.1	0	74.6	ug/Kg	41		0		10	140	20
	Dinoseb	184.1	0	20.1	ug/Kg	11		0		10	118	20