

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS060925

Review By	Abdul	Review On	6/10/2025 10:46:12 AM
Supervise By	mohammad	Supervise On	6/11/2025 2:16:12 AM
SubDirectory	PS060925	HP Acquire Method	HP Processing Method ps060425 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030561.D	09 Jun 2025 11:29		AR\AJ	Ok
2	I.BLK	I.BLK	PS030562.D	09 Jun 2025 11:53		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030563.D	09 Jun 2025 12:17	Picloram low in 1st column	AR\AJ	Ok,M
4	Q2235-03	WC-A2-08-C	PS030564.D	09 Jun 2025 13:21		AR\AJ	Ok,M
5	Q2236-03	WC-A4-05A-C	PS030565.D	09 Jun 2025 13:45		AR\AJ	Ok
6	Q2236-07	WC-A2-04-C	PS030566.D	09 Jun 2025 14:09		AR\AJ	Ok
7	Q2236-11	WC-A2-05-C	PS030567.D	09 Jun 2025 14:33		AR\AJ	Ok
8	Q2236-15	WC-A2-06-C	PS030568.D	09 Jun 2025 14:57		AR\AJ	Ok
9	Q2236-19	WC-A2-07-C	PS030569.D	09 Jun 2025 15:22		AR\AJ	Ok
10	Q2198-02	B-202-SB02	PS030570.D	09 Jun 2025 15:46		AR\AJ	Ok,M
11	Q2198-04	B-207-SB02	PS030571.D	09 Jun 2025 16:10		AR\AJ	Ok
12	Q2198-04MS	B-207-SB02MS	PS030572.D	09 Jun 2025 16:34	some compound recovery fail	AR\AJ	Ok,M
13	Q2198-04MSD	B-207-SB02MSD	PS030573.D	09 Jun 2025 16:58	some compound recovery fail	AR\AJ	Ok,M
14	I.BLK	I.BLK	PS030574.D	09 Jun 2025 17:22		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS030575.D	09 Jun 2025 17:46		AR\AJ	Ok,M
16	Q2130-01	TP-3	PS030576.D	09 Jun 2025 18:17		AR\AJ	Ok,M
17	Q2130-01MS	TP-3MS	PS030577.D	09 Jun 2025 18:41		AR\AJ	Ok,M
18	Q2130-01MSD	TP-3MSD	PS030578.D	09 Jun 2025 19:05		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS060925

Review By	Abdul	Review On	6/10/2025 10:46:12 AM
Supervise By	mohammad	Supervise On	6/11/2025 2:16:12 AM
SubDirectory	PS060925	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

19	I.BLK	I.BLK	PS030579.D	09 Jun 2025 19:29		AR\AJ	Ok
20	HSTDCCC750	HSTDCCC750	PS030580.D	09 Jun 2025 19:53		AR\AJ	Ok,M
21	Q2206-01	TP-1	PS030581.D	09 Jun 2025 20:17	Surrogate low in 2nd column	AR\AJ	ReRun
22	Q2226-01	TP06-MHI-WC	PS030582.D	09 Jun 2025 20:41		AR\AJ	Ok
23	Q2240-01	TP-3	PS030583.D	09 Jun 2025 21:05		AR\AJ	Ok
24	Q2240-05	TP-2	PS030584.D	09 Jun 2025 21:29	Surrogate low in both column	AR\AJ	Not Ok
25	Q2240-09	TP-1	PS030585.D	09 Jun 2025 21:53	Surrogate low in both column	AR\AJ	Not Ok
26	Q2241-01	TP-N	PS030586.D	09 Jun 2025 22:17		AR\AJ	Ok
27	Q2241-05	TP-S	PS030587.D	09 Jun 2025 22:42		AR\AJ	Ok
28	Q2242-01	TP09-MHJ	PS030588.D	09 Jun 2025 23:06		AR\AJ	Ok
29	Q2244-01	TP03-MHC	PS030589.D	09 Jun 2025 23:30		AR\AJ	Ok
30	I.BLK	I.BLK	PS030590.D	09 Jun 2025 23:54		AR\AJ	Ok
31	HSTDCCC750	HSTDCCC750	PS030591.D	10 Jun 2025 00:18		AR\AJ	Ok,M

M : Manual Integration