

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072461.D Time Analyzed: 10:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.655	5.552	5.752	422.630	500.000	-15.5
Aroclor-1016-2	5.676	5.574	5.774	470.170	500.000	-6.0
Aroclor-1016-3	5.738	5.636	5.836	468.320	500.000	-6.3
Aroclor-1016-4	5.836	5.734	5.934	488.430	500.000	-2.3
Aroclor-1016-5	6.128	6.027	6.227	478.290	500.000	-4.3
Aroclor-1260-1	7.247	7.147	7.347	455.460	500.000	-8.9
Aroclor-1260-2	7.501	7.401	7.601	471.800	500.000	-5.6
Aroclor-1260-3	7.859	7.760	7.960	493.710	500.000	-1.3
Aroclor-1260-4	8.083	7.985	8.185	484.210	500.000	-3.2
Aroclor-1260-5	8.402	8.303	8.503	506.790	500.000	1.4
Decachlorobiphenyl	10.210	10.113	10.313	52.720	50.000	5.4
Tetrachloro-m-xylene	4.502	4.396	4.596	47.950	50.000	-4.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072461.D Time Analyzed: 10:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	498.390	500.000	-0.3
Aroclor-1016-2	4.896	4.795	4.995	499.930	500.000	0.0
Aroclor-1016-3	5.073	4.973	5.173	502.840	500.000	0.6
Aroclor-1016-4	5.114	5.015	5.215	506.060	500.000	1.2
Aroclor-1016-5	5.329	5.229	5.429	496.780	500.000	-0.6
Aroclor-1260-1	6.361	6.264	6.464	523.800	500.000	4.8
Aroclor-1260-2	6.550	6.453	6.653	499.500	500.000	-0.1
Aroclor-1260-3	6.702	6.606	6.806	530.710	500.000	6.1
Aroclor-1260-4	7.173	7.077	7.277	518.770	500.000	3.8
Aroclor-1260-5	7.415	7.319	7.519	517.690	500.000	3.5
Decachlorobiphenyl	8.816	8.723	8.923	55.760	50.000	11.5
Tetrachloro-m-xylene	3.796	3.691	3.891	49.680	50.000	-0.6