

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072476.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.655	5.552	5.752	416.340	500.000	-16.7
Aroclor-1016-2	5.677	5.574	5.774	449.250	500.000	-10.2
Aroclor-1016-3	5.739	5.636	5.836	436.950	500.000	-12.6
Aroclor-1016-4	5.837	5.734	5.934	456.260	500.000	-8.7
Aroclor-1016-5	6.129	6.027	6.227	446.620	500.000	-10.7
Aroclor-1260-1	7.248	7.147	7.347	438.650	500.000	-12.3
Aroclor-1260-2	7.501	7.401	7.601	447.180	500.000	-10.6
Aroclor-1260-3	7.860	7.760	7.960	462.100	500.000	-7.6
Aroclor-1260-4	8.084	7.985	8.185	456.590	500.000	-8.7
Aroclor-1260-5	8.403	8.303	8.503	467.390	500.000	-6.5
Decachlorobiphenyl	10.211	10.113	10.313	48.580	50.000	-2.8
Tetrachloro-m-xylene	4.503	4.396	4.596	46.790	50.000	-6.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072476.D Time Analyzed: 16:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.878	4.777	4.977	482.170	500.000	-3.6
Aroclor-1016-2	4.896	4.795	4.995	495.180	500.000	-1.0
Aroclor-1016-3	5.074	4.973	5.173	495.170	500.000	-1.0
Aroclor-1016-4	5.115	5.015	5.215	501.040	500.000	0.2
Aroclor-1016-5	5.330	5.229	5.429	493.810	500.000	-1.2
Aroclor-1260-1	6.362	6.264	6.464	516.350	500.000	3.3
Aroclor-1260-2	6.551	6.453	6.653	491.560	500.000	-1.7
Aroclor-1260-3	6.704	6.606	6.806	527.110	500.000	5.4
Aroclor-1260-4	7.175	7.077	7.277	515.270	500.000	3.1
Aroclor-1260-5	7.416	7.319	7.519	509.650	500.000	1.9
Decachlorobiphenyl	8.819	8.723	8.923	51.310	50.000	2.6
Tetrachloro-m-xylene	3.796	3.691	3.891	50.540	50.000	1.1