

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072491.D Time Analyzed: 21:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.656	5.552	5.752	414.050	500.000	-17.2
Aroclor-1016-2	5.677	5.574	5.774	443.170	500.000	-11.4
Aroclor-1016-3	5.739	5.636	5.836	435.990	500.000	-12.8
Aroclor-1016-4	5.837	5.734	5.934	447.390	500.000	-10.5
Aroclor-1016-5	6.129	6.027	6.227	438.220	500.000	-12.4
Aroclor-1260-1	7.248	7.147	7.347	426.950	500.000	-14.6
Aroclor-1260-2	7.501	7.401	7.601	419.840	500.000	-16.0
Aroclor-1260-3	7.860	7.760	7.960	435.350	500.000	-12.9
Aroclor-1260-4	8.084	7.985	8.185	424.930	500.000	-15.0
Aroclor-1260-5	8.403	8.303	8.503	444.970	500.000	-11.0
Decachlorobiphenyl	10.211	10.113	10.313	46.890	50.000	-6.2
Tetrachloro-m-xylene	4.503	4.396	4.596	46.710	50.000	-6.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2150 SAS No.: Q2150 SDG NO.: Q2150

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/30/2025

Lab Sample No.: AR1660CCC500 Data File : PP072491.D Time Analyzed: 21:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.880	4.777	4.977	500.510	500.000	0.1
Aroclor-1016-2	4.897	4.795	4.995	505.890	500.000	1.2
Aroclor-1016-3	5.074	4.973	5.173	512.590	500.000	2.5
Aroclor-1016-4	5.116	5.015	5.215	515.110	500.000	3.0
Aroclor-1016-5	5.330	5.229	5.429	506.240	500.000	1.2
Aroclor-1260-1	6.363	6.264	6.464	499.960	500.000	0.0
Aroclor-1260-2	6.552	6.453	6.653	476.940	500.000	-4.6
Aroclor-1260-3	6.705	6.606	6.806	508.980	500.000	1.8
Aroclor-1260-4	7.175	7.077	7.277	507.230	500.000	1.4
Aroclor-1260-5	7.417	7.319	7.519	506.090	500.000	1.2
Decachlorobiphenyl	8.819	8.723	8.923	52.090	50.000	4.2
Tetrachloro-m-xylene	3.797	3.691	3.891	50.530	50.000	1.1